

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MACGILLIVARY, SEAN
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 07/03/2019 00:51:40

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.081	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.190	00:47
Breath Volume	1.785	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.188	00:50
Breath Volume	2.265	00:50
Air Blank	0.000	00:50
Control Test	0.081	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: TAMBURRO

DATE/TIME OF TEST: 07/05/2019 00:33:31

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.081	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.080	00:29
Breath Volume	1.742	00:29
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.079	00:32
Breath Volume	2.054	00:32
Air Blank	0.000	00:32
Control Test	0.081	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PONDER,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 07/05/2019 03:55:15

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.081	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.127	03:50
Breath Volume	1.945	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.128	03:53
Breath Volume	1.480	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HOLLAND,TROY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GLOVER

DATE/TIME OF TEST: 07/05/2019 04:09:31

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.079	04:04
Air Blank		0.000	04:04
Subject Sample	# 1	0.225	04:05
Breath Volume		2.660	04:05
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.230	04:08
Breath Volume		2.367	04:08
Air Blank		0.000	04:08
Control Test		0.080	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: BEDFORD

DATE/TIME OF TEST: 07/05/2019 04:30:04

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.080	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.000	04:25
Breath Volume	2.187	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.000	04:28
Breath Volume	2.492	04:28
Air Blank	0.000	04:29
Control Test	0.081	04:29
Air Blank	0.000	04:29
Diagnostics Check	OK	04:30

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ROSS,JACOB
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: AGULIAR MIRANDA

DATE/TIME OF TEST: 07/07/2019 00:15:21

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.080	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.188	00:10
Breath Volume	1.816	00:10
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.197	00:13
Breath Volume	1.343	00:13
Air Blank	0.000	00:14
Control Test	0.081	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DELCASTILLO,BRET
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: BERNARDO VASQUEZ

DATE/TIME OF TEST: 07/07/2019 04:57:22

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.081	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.166	04:52
Breath Volume	3.023	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.170	04:55
Breath Volume	1.742	04:55
Air Blank	0.000	04:56
Control Test	0.082	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SACCENTE,LYNN
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: EVANDER

DATE/TIME OF TEST: 07/07/2019 05:04:48

Diagnostics Check	OK	04:58
Air Blank	0.000	04:59
Control Test	0.080	04:59
Air Blank	0.000	05:00
Subject Sample # 1	0.163	05:00
Breath Volume	3.492	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:02
Subject Sample # 2	0.162	05:03
Breath Volume	3.449	05:03
Air Blank	0.000	05:03
Control Test	0.081	05:04
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANCION,CALIX,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ORTIZ,PAUL
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 06:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:14

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 07/07/2019 07:41:40

Diagnostics Check	OK	07:35
Air Blank	0.000	07:35
Control Test	0.080	07:36
Air Blank	0.000	07:36
Subject Sample # 1	0.092	07:37
Breath Volume	1.253	07:37
Air Blank	0.000	07:37
Air Blank	0.000	07:39
Subject Sample # 2	0.092	07:40
Breath Volume	1.496	07:40
Air Blank	0.000	07:40
Control Test	0.081	07:41
Air Blank	0.000	07:41
Diagnostics Check	OK	07:41

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KNIGHT, RYAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: BELANSKY

DATE/TIME OF TEST: 07/09/2019 03:39:58

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.081	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.196	03:35
Breath Volume	1.753	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.196	03:38
Breath Volume	1.539	03:38
Air Blank	0.000	03:39
Control Test	0.081	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DONNELLY, SHANE
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: SETAL

DATE/TIME OF TEST: 07/09/2019 20:21:34

Diagnostics Check	OK	20:15
Air Blank	0.000	20:15
Control Test	0.081	20:15
Air Blank	0.000	20:16
Subject Sample # 1	0.163	20:16
Breath Volume	2.582	20:16
Air Blank	0.000	20:17
Air Blank	0.000	20:19
Subject Sample # 2	0.163	20:20
Breath Volume	2.226	20:20
Air Blank	0.000	20:20
Control Test	0.081	20:20
Air Blank	0.000	20:21
Diagnostics Check	OK	20:21

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PONDER,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: ABDOLLAHI

DATE/TIME OF TEST: 07/11/2019 04:27:37

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.080	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.106	04:23
Breath Volume	1.886	04:23
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.103	04:26
Breath Volume	1.921	04:26
Air Blank	0.000	04:26
Control Test	0.080	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIDDLE,JAMES,A
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: MARION,RICHARD
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 06:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:19

SUBJECT'S LAST NAME: DEES

DATE/TIME OF TEST: 07/11/2019 07:57:00

Diagnostics Check	OK	07:45	
Air Blank	0.000	07:45	
Control Test	0.080	07:46	
Air Blank	0.000	07:46	
Subject Sample	# 1 0.195	07:49	Volume Not Met
Breath Volume	0.355	07:49	
Air Blank	0.000	07:50	
Air Blank	0.000	07:52	
Subject Sample	# 2 0.180	07:55	Volume Not Met
Breath Volume	0.828	07:55	
Air Blank	0.000	07:56	
Control Test	0.081	07:56	
Air Blank	0.000	07:56	
Diagnostics Check	OK	07:57	

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,TROY,M
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HOLLAND,TROY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 07/12/2019 20:50:35

Diagnostics Check	OK	20:43
Air Blank	0.000	20:44
Control Test	0.080	20:44
Air Blank	0.000	20:45
Subject Sample # 1	0.205	20:45
Breath Volume	1.359	20:45
Air Blank	0.000	20:46
Air Blank	0.000	20:48
Subject Sample # 2	0.199	20:49
Breath Volume	1.140	20:49
Air Blank	0.000	20:49
Control Test	0.081	20:50
Air Blank	0.000	20:50
Diagnostics Check	OK	20:50

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SALGADO,JOSE
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: FAULKNER

DATE/TIME OF TEST: 07/13/2019 01:40:54

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:30	
Control Test	0.080	01:30	
Air Blank	0.000	01:31	
Subject Sample	# 1 0.137	01:34	Volume Not Met
Breath Volume	0.316	01:34	
Air Blank	0.000	01:35	
Air Blank	0.000	01:36	
Subject Sample	# 2	01:39	Subject Test Refused
Breath Volume	0.035	01:39	
Air Blank	0.000	01:40	
Control Test	0.080	01:40	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: LOHRI, BRENDAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: VERBAL

DATE/TIME OF TEST: 07/14/2019 01:36:40

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.081	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.102	01:32
Breath Volume	3.136	01:32
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.097	01:35
Breath Volume	2.457	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HOLLAND, TROY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ROLLE

DATE/TIME OF TEST: 07/14/2019 02:21:26

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.081	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.000	02:16
Breath Volume	2.050	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.000	02:20
Breath Volume	1.394	02:20
Air Blank	0.000	02:20
Control Test	0.081	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TORRES, JOSUE
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: RAMIREZ WOOD

DATE/TIME OF TEST: 07/14/2019 05:23:54

Diagnostics Check	OK	05:17
Air Blank	0.000	05:18
Control Test	0.080	05:18
Air Blank	0.000	05:18
Subject Sample # 1	0.151	05:19
Breath Volume	1.312	05:19
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.143	05:22
Breath Volume	1.460	05:22
Air Blank	0.000	05:23
Control Test	0.081	05:23
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: JOHNSON, ADAM
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 07/17/2019 03:20:03

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.000	03:15
Breath Volume	1.484	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.000	03:18
Breath Volume	1.898	03:18
Air Blank	0.000	03:19
Control Test	0.082	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:20

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HILL,SEAN
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: COLON CORTADA

DATE/TIME OF TEST: 07/21/2019 01:08:17

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.080	01:02
Air Blank	0.000	01:03
Subject Sample # 1	0.094	01:03
Breath Volume	1.976	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.088	01:06
Breath Volume	1.789	01:06
Air Blank	0.000	01:07
Control Test	0.081	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANCION,CALIX,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BURNS,C
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 11:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:55

SUBJECT'S LAST NAME: LYTTLE

DATE/TIME OF TEST: 07/21/2019 12:34:27

Diagnostics Check	OK	12:25
Air Blank	0.000	12:25
Control Test	0.078	12:26
Air Blank	0.000	12:26
Subject Sample # 1	0.227	12:27
Breath Volume	1.332	12:27
Air Blank	0.000	12:28
Air Blank	0.000	12:30
Subject Sample # 2	0.227	12:32
Breath Volume	1.304	12:32
Air Blank	0.000	12:33
Control Test	0.080	12:33
Air Blank	0.000	12:34
Diagnostics Check	OK	12:34

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: LOHRI, BRENDAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 07/22/2019 03:54:13

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.098	03:49
Breath Volume	1.218	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.095	03:52
Breath Volume	1.203	03:52
Air Blank	0.000	03:53
Control Test	0.081	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,TROY,M
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HOLLAND,TROY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: RESTREPO DIAZ

DATE/TIME OF TEST: 07/23/2019 05:09:50

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.081	05:03
Air Blank		0.000	05:04
Subject Sample	# 1	0.165	05:05
Breath Volume		2.601	05:05
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.163	05:08
Breath Volume		1.695	05:08
Air Blank		0.000	05:08
Control Test		0.082	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CRAWFORD,L,P

BREATH TEST OPERATOR AGENCY: Lake Mary PD

ARRESTING OFFICER: NAPIER,J

ARRESTING OFFICER AGENCY: Casselberry PD

LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 915310

DRY GAS EXPIRY DATE: 03/07/2021

VIOLATION CODE: DUI

ARREST TIME: 16:33

GENDER: Female

OBSERVATION PERIOD BEGAN: 17:32

SUBJECT'S LAST NAME: BELANGER

DATE/TIME OF TEST: 07/23/2019 17:57:04

Diagnostics Check	OK	17:56	
Air Blank	0.000	17:56	
		17:56	RFI Detect
Air Blank	0.000	17:57	

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,L,P
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: NAPIER,J
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:33

SUBJECT'S LAST NAME: BELANGER

DATE/TIME OF TEST: 07/23/2019 18:10:21

Diagnostics Check	OK	18:04
Air Blank	0.000	18:04
Control Test	0.081	18:04
Air Blank	0.000	18:05
Subject Sample # 1	0.114	18:05
Breath Volume	1.175	18:05
Air Blank	0.000	18:06
Air Blank	0.000	18:08
Subject Sample # 2	0.108	18:08
Breath Volume	1.828	18:08
Air Blank	0.000	18:09
Control Test	0.081	18:09
Air Blank	0.000	18:10
Diagnostics Check	OK	18:10

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SMITH,BLAKE
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 07/25/2019 04:08:42

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:03
Subject Sample # 1	0.076	04:04
Breath Volume	2.085	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.072	04:07
Breath Volume	1.882	04:07
Air Blank	0.000	04:07
Control Test	0.082	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 07/27/2019 01:20:03

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.081	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.177	01:15
Breath Volume	1.332	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.167	01:18
Breath Volume	2.273	01:18
Air Blank	0.000	01:19
Control Test	0.082	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:20

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SANTIAGO,EDUARDO
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: ENGLISH

DATE/TIME OF TEST: 07/27/2019 01:43:31

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.080	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.249	01:39
Breath Volume	2.789	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.248	01:42
Breath Volume	2.996	01:42
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HOLLAND,TROY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/29/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: ROSARIO MANDES

DATE/TIME OF TEST: 07/27/2019 02:47:47

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.217	02:43
Breath Volume	1.304	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:45
Subject Sample # 2	0.213	02:46
Breath Volume	1.312	02:46
Air Blank	0.000	02:46
Control Test	0.080	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 08/03/2019 04:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 07/01/2019 02:39:32

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:37	
Control Test	0.081	02:37	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:38	Subject Test Refused
Breath Volume	0.000	02:38	
Air Blank	0.000	02:38	
Control Test	0.081	02:38	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:39	

Data Download Date/Time: 07/10/2019 20:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: FRASZ

DATE/TIME OF TEST: 07/02/2019 20:56:40

Diagnostics Check		OK	20:48
Air Blank		0.000	20:49
Control Test		0.081	20:49
Air Blank		0.000	20:50
Subject Sample	# 1	0.216	20:51
Breath Volume		1.460	20:51
Air Blank		0.000	20:51
Air Blank		0.000	20:53
Subject Sample	# 2	0.211	20:55
Breath Volume		1.121	20:55
Air Blank		0.000	20:55
Control Test		0.081	20:56
Air Blank		0.000	20:56
Diagnostics Check		OK	20:56

Data Download Date/Time: 07/10/2019 20:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: MC ELROY

DATE/TIME OF TEST: 07/03/2019 20:37:28

Diagnostics Check	OK	20:31
Air Blank	0.000	20:31
Control Test	0.082	20:31
Air Blank	0.000	20:32
Subject Sample # 1	0.301	20:32
Breath Volume	2.574	20:32
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.315	20:35
Breath Volume	2.839	20:35
Air Blank	0.000	20:36
Control Test	0.082	20:36
Air Blank	0.000	20:37
Diagnostics Check	OK	20:37

Data Download Date/Time: 07/10/2019 20:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: DOUTY

DATE/TIME OF TEST: 07/04/2019 01:28:14

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:25	
Control Test	0.081	01:26	
Air Blank	0.000	01:26	
Subject Sample	# 1	01:26	Subject Test Refused
Breath Volume	0.000	01:26	
Air Blank	0.000	01:27	
Control Test	0.081	01:27	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 07/10/2019 20:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/04/2019 04:49:13

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:46	
Control Test	0.082	04:47	
Air Blank	0.000	04:47	
Subject Sample	# 1	04:47	Subject Test Refused
Breath Volume	0.000	04:47	
Air Blank	0.000	04:48	
Control Test	0.082	04:48	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

Data Download Date/Time: 07/10/2019 20:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: REIDY,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:49

SUBJECT'S LAST NAME: NECKER

DATE/TIME OF TEST: 07/04/2019 07:16:05

Diagnostics Check	OK	07:09
Air Blank	0.000	07:09
Control Test	0.082	07:10
Air Blank	0.000	07:10
Subject Sample # 1	0.152	07:11
Breath Volume	2.058	07:11
Air Blank	0.000	07:12
Air Blank	0.000	07:14
Subject Sample # 2	0.146	07:14
Breath Volume	2.070	07:14
Air Blank	0.000	07:15
Control Test	0.082	07:15
Air Blank	0.000	07:15
Diagnostics Check	OK	07:16

Data Download Date/Time: 07/10/2019 20:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: DEFRANCO

DATE/TIME OF TEST: 07/05/2019 04:00:51

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.081	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.074	03:56
Breath Volume	2.816	03:56
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.072	03:59
Breath Volume	2.003	03:59
Air Blank	0.000	03:59
Control Test	0.081	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 07/10/2019 20:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/06/2019 04:04:01

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:01	
Control Test	0.082	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:02	Subject Test Refused
Breath Volume	0.000	04:02	
Air Blank	0.000	04:03	
Control Test	0.081	04:03	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:04	

Data Download Date/Time: 07/10/2019 20:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:14

SUBJECT'S LAST NAME: COTTER

DATE/TIME OF TEST: 07/06/2019 16:53:09

Diagnostics Check	OK	16:50	
Air Blank	0.000	16:50	
Control Test	0.081	16:51	
Air Blank	0.000	16:51	
Subject Sample	# 1	16:51	Subject Test Refused
Breath Volume	0.000	16:51	
Air Blank	0.000	16:52	
Control Test	0.081	16:52	
Air Blank	0.000	16:53	
Diagnostics Check	OK	16:53	

Data Download Date/Time: 07/10/2019 20:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BUSCH,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: MCCALL

DATE/TIME OF TEST: 07/08/2019 08:27:55

Diagnostics Check	OK	08:21
Air Blank	0.000	08:21
Control Test	0.082	08:21
Air Blank	0.000	08:22
Subject Sample # 1	0.267	08:23
Breath Volume	2.589	08:23
Air Blank	0.000	08:23
Air Blank	0.000	08:25
Subject Sample # 2	0.264	08:26
Breath Volume	2.203	08:26
Air Blank	0.000	08:27
Control Test	0.082	08:27
Air Blank	0.000	08:27
Diagnostics Check	OK	08:27

Data Download Date/Time: 07/10/2019 20:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DORMAN,ZAKARIAH
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: DULLEA

DATE/TIME OF TEST: 07/09/2019 03:52:33

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:49	
Control Test	0.082	03:49	
Air Blank	0.000	03:50	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:51	
Control Test	0.082	03:51	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:52	

Data Download Date/Time: 07/10/2019 20:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VICKERS,JAMES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BELCASTRO

DATE/TIME OF TEST: 07/09/2019 04:05:12

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.082	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.253	04:00
Breath Volume	3.011	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:02
Subject Sample # 2	0.246	04:03
Breath Volume	3.421	04:03
Air Blank	0.000	04:04
Control Test	0.082	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 07/10/2019 20:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBSEN,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 07/11/2019 00:11:36

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:09	
Control Test	0.081	00:09	
Air Blank	0.000	00:09	
Subject Sample	# 1	00:10	Subject Test Refused
Breath Volume	0.000	00:10	
Air Blank	0.000	00:10	
Control Test	0.081	00:11	
Air Blank	0.000	00:11	
Diagnostics Check	OK	00:11	

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: SPILLMAN

DATE/TIME OF TEST: 07/11/2019 23:35:00

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.082	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.236	23:29
Breath Volume	1.656	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.232	23:33
Breath Volume	1.277	23:33
Air Blank	0.000	23:34
Control Test	0.082	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:35

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: KIRTS

DATE/TIME OF TEST: 07/12/2019 01:27:55

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:17	
Control Test	0.082	01:17	
Air Blank	0.000	01:18	
Subject Sample	# 1	01:21	No Sample Provided
Breath Volume	0.000	01:21	
Air Blank	0.000	01:21	
Air Blank	0.000	01:23	
Subject Sample	# 2	01:26	Subject Test Refused
Breath Volume	0.000	01:26	
Air Blank	0.000	01:27	
Control Test	0.082	01:27	
Air Blank	0.000	01:27	
Diagnostics Check	OK	01:27	

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 07/13/2019 03:49:41

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.085	03:45
Breath Volume	2.382	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.076	03:48
Breath Volume	1.167	03:48
Air Blank	0.000	03:48
Control Test	0.079	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: MARTINO

DATE/TIME OF TEST: 07/13/2019 23:52:32

Diagnostics Check	OK	23:44
Air Blank	0.000	23:45
Control Test	0.081	23:45
Air Blank	0.000	23:46
Subject Sample # 1	0.207	23:47
Breath Volume	1.484	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:49
Subject Sample # 2	0.210	23:51
Breath Volume	1.250	23:51
Air Blank	0.000	23:51
Control Test	0.081	23:51
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: ZAKI

DATE/TIME OF TEST: 07/14/2019 00:19:34

Diagnostics Check	OK	00:12
Air Blank	0.000	00:12
Control Test	0.081	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.032	00:15
Breath Volume	2.398	00:15
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.029	00:18
Breath Volume	1.886	00:18
Air Blank	0.000	00:18
Control Test	0.080	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: PISCITELLI

DATE/TIME OF TEST: 07/14/2019 01:02:04

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.081	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.156	00:57
Breath Volume		1.972	00:57
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.158	01:00
Breath Volume		2.007	01:00
Air Blank		0.000	01:01
Control Test		0.081	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:02

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: WARNER

DATE/TIME OF TEST: 07/14/2019 04:41:33

Diagnostics Check		OK	04:33
Air Blank		0.000	04:33
Control Test		0.081	04:33
Air Blank		0.000	04:34
Subject Sample	# 1	0.158	04:36
Breath Volume		2.570	04:36
Air Blank		0.000	04:37
Air Blank		0.000	04:39
Subject Sample	# 2	0.153	04:40
Breath Volume		2.050	04:40
Air Blank		0.000	04:40
Control Test		0.082	04:40
Air Blank		0.000	04:41
Diagnostics Check		OK	04:41

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVIS
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUID,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 07/14/2019 06:27:26

Diagnostics Check	OK	06:21
Air Blank	0.000	06:21
Control Test	0.080	06:21
Air Blank	0.000	06:22
Subject Sample # 1	0.150	06:22
Breath Volume	2.410	06:22
Air Blank	0.000	06:23
Air Blank	0.000	06:25
Subject Sample # 2	0.148	06:25
Breath Volume	2.480	06:25
Air Blank	0.000	06:26
Control Test	0.081	06:26
Air Blank	0.000	06:27
Diagnostics Check	OK	06:27

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS,M,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COOK,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CRUZ-PENA

DATE/TIME OF TEST: 07/15/2019 03:31:56

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:29	
Control Test	0.082	03:29	
Air Blank	0.000	03:29	
Subject Sample	# 1	03:30	RFI Detect
Breath Volume	0.671	03:30	
Air Blank	0.000	03:31	
Control Test	0.081	03:31	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS,M,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COOK,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CRUZ-PENA

DATE/TIME OF TEST: 07/15/2019 03:37:21

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.082	03:35	
Air Blank	0.000	03:36	
Subject Sample	# 1	03:36	RFI Detect
Breath Volume	0.000	03:36	
Air Blank		03:36	RFI Detect
Air Blank	0.000	03:37	

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 07/18/2019 03:45:27

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.081	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.041	03:41
Breath Volume	1.667	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.040	03:43
Breath Volume	1.355	03:43
Air Blank	0.000	03:44
Control Test	0.081	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: COLMAN

DATE/TIME OF TEST: 07/18/2019 05:04:26

Diagnostics Check	OK	04:57
Air Blank	0.000	04:58
Control Test	0.081	04:58
Air Blank	0.000	04:59
Subject Sample # 1	0.153	04:59
Breath Volume	2.488	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.149	05:02
Breath Volume	2.460	05:02
Air Blank	0.000	05:03
Control Test	0.081	05:03
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: GUNTUKU

DATE/TIME OF TEST: 07/18/2019 23:44:07

Diagnostics Check	OK	23:39	
Air Blank	0.000	23:39	
Control Test	0.081	23:40	
Air Blank	0.000	23:40	
Subject Sample	# 1	23:42	Subject Test Refused
Breath Volume	0.000	23:42	
Air Blank	0.000	23:43	
Control Test	0.081	23:43	
Air Blank	0.000	23:44	
Diagnostics Check	OK	23:44	

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: RAILTON

DATE/TIME OF TEST: 07/19/2019 00:36:06

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:28	
Control Test	0.081	00:28	
Air Blank	0.000	00:28	
Subject Sample	# 1	00:32	No Sample Provided
Breath Volume	0.000	00:32	
Air Blank	0.000	00:32	
Air Blank	0.000	00:34	
Subject Sample	# 2	00:34	Subject Test Refused
Breath Volume	0.000	00:34	
Air Blank	0.000	00:35	
Control Test	0.082	00:35	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: FERWORN

DATE/TIME OF TEST: 07/20/2019 02:30:33

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:27	
Control Test	0.082	02:28	
Air Blank	0.000	02:28	
Subject Sample	# 1	02:29	Subject Test Refused
Breath Volume	0.000	02:29	
Air Blank	0.000	02:29	
Control Test	0.081	02:29	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: TRANCHITA

DATE/TIME OF TEST: 07/20/2019 04:59:35

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:51	
Control Test	0.081	04:51	
Air Blank	0.000	04:52	
Subject Sample	# 1	04:55	No Sample Provided
Breath Volume	0.000	04:55	
Air Blank	0.000	04:55	
Air Blank	0.000	04:57	
Subject Sample	# 2	04:58	Subject Test Refused
Breath Volume	0.000	04:58	
Air Blank	0.000	04:58	
Control Test	0.082	04:59	
Air Blank	0.000	04:59	
Diagnostics Check	OK	04:59	

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:19

SUBJECT'S LAST NAME: WESLEY

DATE/TIME OF TEST: 07/20/2019 06:57:37

Diagnostics Check		OK	06:47
Air Blank		0.000	06:48
Control Test		0.081	06:48
Air Blank		0.000	06:49
Subject Sample	# 1	0.164	06:49
Breath Volume		2.789	06:49
Air Blank		0.000	06:50
Air Blank		0.000	06:52
Subject Sample	# 2	0.140	06:52
Breath Volume		1.156	06:52
Air Blank		0.000	06:53
Air Blank		0.000	06:55
Subject Sample	# 3	0.141	06:56
Breath Volume		1.101	06:56
Air Blank		0.000	06:56
Control Test		0.081	06:57
Air Blank		0.000	06:57
Diagnostics Check		OK	06:57

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: SANCHEZ-FLORES

DATE/TIME OF TEST: 07/21/2019 01:54:02

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.081	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.136	01:49
Breath Volume	1.765	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.135	01:52
Breath Volume	2.246	01:52
Air Blank	0.000	01:53
Control Test	0.081	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:54

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: FLORES ORTIS

DATE/TIME OF TEST: 07/21/2019 04:15:03

Diagnostics Check	OK	04:06
Air Blank	0.000	04:07
Control Test	0.081	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.187	04:09
Breath Volume	2.386	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.187	04:13
Breath Volume	2.589	04:13
Air Blank	0.000	04:14
Control Test	0.081	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:15

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: BOUNDS

DATE/TIME OF TEST: 07/21/2019 20:31:29

Diagnostics Check	OK	20:24	
Air Blank	0.000	20:25	
Control Test	0.082	20:25	
Air Blank	0.000	20:25	
Subject Sample	# 1 0.108	20:26	
Breath Volume	1.550	20:26	
Air Blank	0.000	20:27	
Air Blank	0.000	20:28	
Subject Sample	# 2	20:30	Subject Test Refused
Breath Volume	0.000	20:30	
Air Blank	0.000	20:30	
Control Test	0.081	20:30	
Air Blank	0.000	20:31	
Diagnostics Check	OK	20:31	

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: BOUNDS

DATE/TIME OF TEST: 07/21/2019 20:52:19

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:47	
Control Test	0.082	20:47	
Air Blank	0.000	20:48	
Subject Sample	# 1	20:50	Subject Test Refused
Breath Volume	0.000	20:50	
Air Blank	0.000	20:51	
Control Test	0.081	20:51	
Air Blank	0.000	20:52	
Diagnostics Check	OK	20:52	

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: HAIGH KING

DATE/TIME OF TEST: 07/24/2019 19:12:56

Diagnostics Check	OK	19:06
Air Blank	0.000	19:07
Control Test	0.082	19:07
Air Blank	0.000	19:07
Subject Sample # 1	0.200	19:08
Breath Volume	1.886	19:08
Air Blank	0.000	19:09
Air Blank	0.000	19:10
Subject Sample # 2	0.219	19:11
Breath Volume	1.917	19:11
Air Blank	0.000	19:12
Control Test	0.081	19:12
Air Blank	0.000	19:12
Diagnostics Check	OK	19:12

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: SHULL

DATE/TIME OF TEST: 07/27/2019 03:54:55

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.081	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.105	03:50
Breath Volume	2.207	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.108	03:53
Breath Volume	2.023	03:53
Air Blank	0.000	03:54
Control Test	0.082	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: PURDY

DATE/TIME OF TEST: 07/28/2019 00:27:11

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.081	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.045	00:22
Breath Volume	1.718	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.045	00:25
Breath Volume	1.625	00:25
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,JARED
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PROSTOSERDOV

DATE/TIME OF TEST: 07/28/2019 02:32:03

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.082	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.045	02:27
Breath Volume	2.128	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.048	02:30
Breath Volume	2.894	02:30
Air Blank	0.000	02:31
Control Test	0.082	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:32

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: REIDY,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: BAUGHMAN

DATE/TIME OF TEST: 07/28/2019 03:11:09

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:08	
Control Test	0.081	03:08	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:10	
Control Test	0.081	03:10	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

Data Download Date/Time: 08/03/2019 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: VALENZUELA PILEGGI

DATE/TIME OF TEST: 07/28/2019 05:47:29

Diagnostics Check	OK	05:40
Air Blank	0.000	05:41
Control Test	0.081	05:41
Air Blank	0.000	05:42
Subject Sample # 1	0.164	05:42
Breath Volume	1.492	05:42
Air Blank	0.000	05:43
Air Blank	0.000	05:45
Subject Sample # 2	0.173	05:45
Breath Volume	2.257	05:45
Air Blank	0.000	05:46
Control Test	0.082	05:46
Air Blank	0.000	05:47
Diagnostics Check	OK	05:47

Data Download Date/Time: 08/03/2019 20:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LAMOUREUX,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: BAIRD

DATE/TIME OF TEST: 07/28/2019 17:54:11

Diagnostics Check	OK	17:47
Air Blank	0.000	17:48
Control Test	0.081	17:48
Air Blank	0.000	17:49
Subject Sample # 1	0.157	17:49
Breath Volume	1.648	17:49
Air Blank	0.000	17:50
Air Blank	0.000	17:52
Subject Sample # 2	0.164	17:52
Breath Volume	2.500	17:52
Air Blank	0.000	17:53
Control Test	0.080	17:53
Air Blank	0.000	17:54
Diagnostics Check	OK	17:54

Data Download Date/Time: 08/03/2019 20:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,STEVEN,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: JAMES,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: GASQUEZ

DATE/TIME OF TEST: 07/28/2019 21:59:29

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.081	21:52
Air Blank	0.000	21:52
Subject Sample # 1	0.258	21:54
Breath Volume	3.984	21:54
Air Blank	0.000	21:55
Air Blank	0.000	21:56
Subject Sample # 2	0.249	21:57
Breath Volume	3.457	21:57
Air Blank	0.000	21:58
Control Test	0.079	21:58
Air Blank	0.000	21:59
Diagnostics Check	OK	21:59

Data Download Date/Time: 08/03/2019 20:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MORRIS,BRIDGETT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: MACCAGLIA

DATE/TIME OF TEST: 07/29/2019 00:59:16

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.081	00:53
Air Blank	0.000	00:53
Subject Sample # 1	0.082	00:54
Breath Volume	2.214	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:56
Subject Sample # 2	0.078	00:57
Breath Volume	1.347	00:57
Air Blank	0.000	00:58
Control Test	0.082	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 08/03/2019 20:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SKALKO,RYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/29/2019 02:11:10

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.082	02:04
Air Blank	0.000	02:04
Subject Sample # 1	0.076	02:06
Breath Volume	1.398	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.072	02:09
Breath Volume	1.253	02:09
Air Blank	0.000	02:10
Control Test	0.081	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 08/03/2019 20:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,SCOTT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BUSTOS,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:16

SUBJECT'S LAST NAME: TORMEY

DATE/TIME OF TEST: 07/30/2019 12:46:53

Diagnostics Check	OK	12:40
Air Blank	0.000	12:40
Control Test	0.082	12:41
Air Blank	0.000	12:41
Subject Sample # 1	0.018	12:42
Breath Volume	2.187	12:42
Air Blank	0.000	12:42
Air Blank	0.000	12:44
Subject Sample # 2	0.016	12:45
Breath Volume	1.218	12:45
Air Blank	0.000	12:45
Control Test	0.081	12:46
Air Blank	0.000	12:46
Diagnostics Check	OK	12:46

Data Download Date/Time: 08/03/2019 20:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 07/30/2019 23:29:58

Diagnostics Check	OK	23:26	
Air Blank	0.000	23:27	
Control Test	0.081	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:28	RFI Detect
Breath Volume	0.000	23:28	
Air Blank	0.000	23:29	
Control Test	0.082	23:29	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

Data Download Date/Time: 08/03/2019 20:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 07/30/2019 23:34:35

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:32	
Control Test	0.081	23:32	
Air Blank	0.000	23:33	
Subject Sample	# 1	23:33	Subject Test Refused
Breath Volume	0.000	23:33	
Air Blank	0.000	23:33	
Control Test	0.081	23:33	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 08/03/2019 20:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANSEN,DUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: KOLLAR

DATE/TIME OF TEST: 07/31/2019 14:37:56

Diagnostics Check	OK	14:34	
Air Blank	0.000	14:35	
Control Test	0.082	14:35	
Air Blank	0.000	14:36	
Subject Sample	# 1	14:36	Subject Test Refused
Breath Volume	0.000	14:36	
Air Blank	0.000	14:37	
Control Test	0.082	14:37	
Air Blank	0.000	14:37	
Diagnostics Check	OK	14:37	

Data Download Date/Time: 08/03/2019 20:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCKENZIE,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:56

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 07/01/2019 07:19:57

Diagnostics Check	OK	07:16	
Air Blank	0.000	07:16	
Control Test	0.079	07:17	
Air Blank	0.000	07:17	
Subject Sample # 1	0.184	07:18	
Breath Volume	3.722	07:18	
Air Blank		07:18	RFI Detect
Air Blank	0.000	07:19	
Control Test	0.080	07:19	
Air Blank	0.000	07:19	
Diagnostics Check	OK	07:19	

Data Download Date/Time: 07/10/2019 23:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCKENZIE,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:56

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 07/01/2019 07:27:44

Diagnostics Check	OK	07:21
Air Blank	0.000	07:21
Control Test	0.078	07:22
Air Blank	0.000	07:22
Subject Sample # 1	0.184	07:23
Breath Volume	2.910	07:23
Air Blank	0.000	07:23
Air Blank	0.000	07:25
Subject Sample # 2	0.175	07:26
Breath Volume	2.414	07:26
Air Blank	0.000	07:26
Control Test	0.079	07:27
Air Blank	0.000	07:27
Diagnostics Check	OK	07:27

Data Download Date/Time: 07/10/2019 23:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CARTER,TIMOTHY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/05/2019 22:47:41

Diagnostics Check	OK	22:44	
Air Blank	0.000	22:45	
Control Test	0.078	22:45	
Air Blank	0.000	22:46	
Subject Sample	# 1	22:46	Subject Test Refused
Breath Volume	0.000	22:46	
Air Blank	0.000	22:46	
Control Test	0.077	22:47	
Air Blank	0.000	22:47	
Diagnostics Check	OK	22:47	

Data Download Date/Time: 07/10/2019 23:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GASSON,STEVE
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: LANGAME

DATE/TIME OF TEST: 07/05/2019 23:24:08

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:21	
Control Test	0.079	23:22	
Air Blank	0.000	23:22	
Subject Sample	# 1	23:22	Subject Test Refused
Breath Volume	0.000	23:22	
Air Blank	0.000	23:23	
Control Test	0.079	23:23	
Air Blank	0.000	23:24	
Diagnostics Check	OK	23:24	

Data Download Date/Time: 07/10/2019 23:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VANWORP,JUDSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 07/05/2019 23:56:09

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:53	
Control Test	0.079	23:54	
Air Blank	0.000	23:54	
Subject Sample	# 1	23:54	Subject Test Refused
Breath Volume	0.000	23:54	
Air Blank	0.000	23:55	
Control Test	0.079	23:55	
Air Blank	0.000	23:56	
Diagnostics Check	OK	23:56	

Data Download Date/Time: 07/10/2019 23:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: DAVID

DATE/TIME OF TEST: 07/06/2019 01:03:38

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.079	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.137	00:59
Breath Volume	3.007	00:59
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.131	01:01
Breath Volume	2.839	01:01
Air Blank	0.000	01:02
Control Test	0.078	01:02
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 07/10/2019 23:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 07/06/2019 02:43:23

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:40	
Control Test	0.079	02:41	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:41	Subject Test Refused
Breath Volume	0.000	02:41	
Air Blank	0.000	02:42	
Control Test	0.078	02:42	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 07/10/2019 23:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ARBANAS,SHANE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: DILLON

DATE/TIME OF TEST: 07/06/2019 03:08:47

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:06	
Control Test	0.079	03:06	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:07	
Control Test	0.079	03:08	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 07/10/2019 23:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FISHER,CLIFF
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/06/2019 03:53:46

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:51	
Control Test	0.079	03:51	
Air Blank	0.000	03:52	
Subject Sample	# 1	03:52	Subject Test Refused
Breath Volume	0.000	03:52	
Air Blank	0.000	03:52	
Control Test	0.078	03:53	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 07/10/2019 23:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CARTER,TIMOTHY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 07/06/2019 04:32:04

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.079	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.081	04:25
Breath Volume	1.449	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.076	04:30
Breath Volume	1.187	04:30
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:32

Data Download Date/Time: 07/10/2019 23:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS,M,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COOK,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CRUZ-PENA

DATE/TIME OF TEST: 07/15/2019 03:21:38

Diagnostics Check	OK	03:21	
Air Blank		03:21	RFI Detect
Air Blank		03:21	RFI Detect

Data Download Date/Time: 08/03/2019 23:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS,M,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COOK,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CRUZ-PENA

DATE/TIME OF TEST: 07/15/2019 03:24:18

Diagnostics Check	OK	03:24	
Air Blank		03:24	RFI Detect
Air Blank		03:24	RFI Detect

Data Download Date/Time: 08/03/2019 23:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCKENZIE,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 07/16/2019 18:57:31

Diagnostics Check	OK	18:49
Air Blank	0.000	18:50
Control Test	0.078	18:50
Air Blank	0.000	18:50
Subject Sample # 1	0.242	18:52
Breath Volume	1.605	18:52
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample # 2	0.226	18:55
Breath Volume	1.152	18:55
Air Blank	0.000	18:56
Control Test	0.079	18:56
Air Blank	0.000	18:57
Diagnostics Check	OK	18:57

Data Download Date/Time: 08/03/2019 23:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,JAMES,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILHELM,JAMES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/17/2019 01:34:01

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:31	
Control Test	0.080	01:31	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:32	Subject Test Refused
Breath Volume	0.000	01:32	
Air Blank	0.000	01:33	
Control Test	0.080	01:33	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 08/03/2019 23:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: ZONJA

DATE/TIME OF TEST: 07/18/2019 03:37:32

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:34	
Control Test	0.079	03:35	
Air Blank	0.000	03:35	
Subject Sample	# 1	03:36	Subject Test Refused
Breath Volume	0.000	03:36	
Air Blank	0.000	03:36	
Control Test	0.079	03:36	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

Data Download Date/Time: 08/03/2019 23:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCKENZIE,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:05

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 07/30/2019 09:28:39

Diagnostics Check	OK	09:25	
Air Blank	0.000	09:25	
Control Test	0.076	09:26	
Air Blank	0.000	09:26	
Subject Sample	# 1	09:27	Subject Test Refused
Breath Volume	0.000	09:27	
Air Blank	0.000	09:27	
Control Test	0.079	09:28	
Air Blank	0.000	09:28	
Diagnostics Check	OK	09:28	

Data Download Date/Time: 08/03/2019 23:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBSON,ZACHARY,A
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: MARTIN,ZACHARY
ARRESTING OFFICER AGENCY: Melbourne Beach PD
LAST AGENCY INSPECTION: 06/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 07/03/2019 22:45:35

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.079	22:39
Air Blank	0.000	22:39
Subject Sample # 1	0.000	22:41
Breath Volume	1.816	22:41
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.000	22:44
Breath Volume	1.328	22:44
Air Blank	0.000	22:44
Control Test	0.078	22:45
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 08/14/2019 04:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KALLMEYER,DANIEL,K
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 07/09/2019 09:37:43

Diagnostics Check	OK	09:34	
Air Blank	0.000	09:34	
Control Test	0.080	09:35	
Air Blank	0.000	09:35	
Subject Sample	# 1	09:36	Interferent Detect
Breath Volume	1.789	09:36	
Air Blank	0.000	09:36	
Control Test	0.084	09:37	
Air Blank	0.000	09:37	
Diagnostics Check	OK	09:37	

Data Download Date/Time: 08/22/2019 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KALLMEYER,DANIEL,K
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: HITLEY

DATE/TIME OF TEST: 07/09/2019 09:48:04

Diagnostics Check	OK	09:42
Air Blank	0.000	09:42
Control Test	0.082	09:42
Air Blank	0.000	09:43
Subject Sample # 1	0.000	09:43
Breath Volume	2.230	09:43
Air Blank	0.000	09:44
Air Blank	0.000	09:46
Subject Sample # 2	0.000	09:46
Breath Volume	2.218	09:46
Air Blank	0.000	09:47
Control Test	0.085	09:47
Air Blank	0.000	09:47
Diagnostics Check	OK	09:48

Data Download Date/Time: 08/22/2019 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KALLMEYER,DANIEL,K
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: LAUX

DATE/TIME OF TEST: 07/09/2019 10:00:24

Diagnostics Check	OK	09:54
Air Blank	0.000	09:54
Control Test	0.082	09:55
Air Blank	0.000	09:55
Subject Sample # 1	0.000	09:56
Breath Volume	2.449	09:56
Air Blank	0.000	09:56
Air Blank	0.000	09:58
Subject Sample # 2	0.000	09:59
Breath Volume	2.433	09:59
Air Blank	0.000	09:59
Control Test	0.080	09:59
Air Blank	0.000	10:00
Diagnostics Check	OK	10:00

Data Download Date/Time: 08/22/2019 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: JACKSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: COURT ORD
ARREST TIME: 13:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: DICKS

DATE/TIME OF TEST: 07/09/2019 14:25:37

Diagnostics Check	OK	14:19
Air Blank	0.000	14:20
Control Test	0.082	14:20
Air Blank	0.000	14:21
Subject Sample # 1	0.007	14:21
Breath Volume	2.039	14:21
Air Blank	0.000	14:21
Air Blank	0.000	14:23
Subject Sample # 2	0.000	14:24
Breath Volume	2.773	14:24
Air Blank	0.000	14:24
Control Test	0.080	14:25
Air Blank	0.000	14:25
Diagnostics Check	OK	14:25

Data Download Date/Time: 08/22/2019 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: JACKSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: COURT ORD
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: DEEMER

DATE/TIME OF TEST: 07/10/2019 14:42:01

Diagnostics Check	OK	14:33	
Air Blank	0.000	14:34	
Control Test	0.082	14:34	
Air Blank	0.000	14:35	
Subject Sample # 1	0.000	14:35	
Breath Volume	1.558	14:35	
Air Blank	0.000	14:36	
Air Blank	0.000	14:38	
Subject Sample # 2		14:40	Subject Test Refused
Breath Volume	0.457	14:40	
Air Blank	0.000	14:41	
Control Test	0.080	14:41	
Air Blank	0.000	14:41	
Diagnostics Check	OK	14:42	

Data Download Date/Time: 08/22/2019 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: JACKSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: COURT ORD
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: DEEMER

DATE/TIME OF TEST: 07/10/2019 14:49:49

Diagnostics Check	OK	14:44
Air Blank	0.000	14:44
Control Test	0.080	14:44
Air Blank	0.000	14:45
Subject Sample # 1	0.000	14:45
Breath Volume	1.789	14:45
Air Blank	0.000	14:46
Air Blank	0.000	14:48
Subject Sample # 2	0.000	14:48
Breath Volume	1.335	14:48
Air Blank	0.000	14:48
Control Test	0.080	14:49
Air Blank	0.000	14:49
Diagnostics Check	OK	14:49

Data Download Date/Time: 08/22/2019 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: AASERUD

DATE/TIME OF TEST: 07/20/2019 21:07:32

Diagnostics Check	OK	21:01
Air Blank	0.000	21:01
Control Test	0.082	21:01
Air Blank	0.000	21:02
Subject Sample # 1	0.097	21:03
Breath Volume	2.007	21:03
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.099	21:06
Breath Volume	2.246	21:06
Air Blank	0.000	21:06
Control Test	0.082	21:06
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

Data Download Date/Time: 08/22/2019 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAMILTON,J,W

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HAMILTON,J

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 02:08

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ROMILUS

DATE/TIME OF TEST: 07/02/2019 02:41:16

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.156	02:36
Breath Volume	1.640	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.158	02:39
Breath Volume	2.031	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 07/29/2019 02:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN, JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: TRIMBOLI, MELANIE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: SIDERS

DATE/TIME OF TEST: 07/05/2019 00:57:50

Diagnostics Check		OK	00:51
Air Blank		0.000	00:52
Control Test		0.079	00:52
Air Blank		0.000	00:52
Subject Sample	# 1	0.205	00:53
Breath Volume		1.617	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:55
Subject Sample	# 2	0.212	00:56
Breath Volume		1.156	00:56
Air Blank		0.000	00:56
Control Test		0.079	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 07/29/2019 02:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SPAULDING,CODY,R

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: SPAULDING,CODY

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 12:44

GENDER: Female

OBSERVATION PERIOD BEGAN: 13:07

SUBJECT'S LAST NAME: TOWNS

DATE/TIME OF TEST: 07/05/2019 13:35:48

Diagnostics Check	OK	13:31	
Air Blank	0.000	13:32	
Control Test	0.079	13:32	
Air Blank	0.000	13:33	
Subject Sample	# 1 0.366	13:34	
Breath Volume	1.453	13:34	
Air Blank		13:35	Purge Fail
Air Blank	0.000	13:35	

Data Download Date/Time: 07/29/2019 02:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SPAULDING,CODY,R

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: SPAULDING,CODY

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 12:44

GENDER: Female

OBSERVATION PERIOD BEGAN: 13:07

SUBJECT'S LAST NAME: TOWNS

DATE/TIME OF TEST: 07/05/2019 13:44:43

Diagnostics Check	OK	13:39	
Air Blank	0.000	13:39	
Control Test	0.075	13:39	
Air Blank	0.000	13:40	
Subject Sample	# 1 0.329	13:43	Volume Not Met
Breath Volume	0.070	13:43	
Air Blank		13:44	Purge Fail
Air Blank	0.000	13:44	

Data Download Date/Time: 07/29/2019 02:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPAULDING,CODY,R
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: SPAULDING,CODY
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:07

SUBJECT'S LAST NAME: TOWNS

DATE/TIME OF TEST: 07/05/2019 14:05:34

Diagnostics Check		OK	13:53	
Air Blank		0.000	13:53	
Control Test		0.076	13:54	
Air Blank		0.000	13:54	
Subject Sample	# 1	0.307	13:57	Volume Not Met
Breath Volume		1.500	13:57	
Air Blank		0.000	13:58	
Air Blank		0.000	14:00	
Subject Sample	# 2	0.332	14:01	
Breath Volume		1.828	14:01	
Air Blank		0.000	14:01	
Air Blank		0.000	14:03	
Subject Sample	# 3	0.349	14:03	
Breath Volume		1.343	14:03	
Air Blank		0.000	14:04	
Control Test		0.076	14:04	
Air Blank		0.000	14:05	
Diagnostics Check		OK	14:05	

Data Download Date/Time: 07/29/2019 02:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAMILTON,J,W

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HAMILTON,J

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24818080A2

DRY GAS EXPIRY DATE: 10/05/2020

VIOLATION CODE: DUI

ARREST TIME: 23:53

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 07/07/2019 00:32:24

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:28	
Control Test	0.079	00:28	
Air Blank	0.000	00:29	
Subject Sample	# 1	00:30	Subject Test Refused
Breath Volume	0.050	00:30	
Air Blank	0.000	00:31	
Control Test	0.079	00:31	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 07/29/2019 02:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: BUCKLEY

DATE/TIME OF TEST: 07/09/2019 22:43:40

Diagnostics Check	OK	22:33
Air Blank	0.000	22:34
Control Test	0.079	22:34
Air Blank	0.000	22:35
Subject Sample # 1	0.176	22:37
Breath Volume	1.363	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.177	22:42
Breath Volume	1.359	22:42
Air Blank	0.000	22:42
Control Test	0.079	22:43
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 07/29/2019 02:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN, JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MILIAN, S
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: GROSKE

DATE/TIME OF TEST: 07/12/2019 01:17:21

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:08	
Control Test	0.079	01:08	
Air Blank	0.000	01:09	
Subject Sample	# 1	01:12	No Sample Provided
Breath Volume	0.000	01:12	
Air Blank	0.000	01:12	
Air Blank	0.000	01:14	
Subject Sample	# 2	01:15	Subject Test Refused
Breath Volume	0.000	01:15	
Air Blank	0.000	01:16	
Control Test	0.079	01:16	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 07/29/2019 02:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN, JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MILIAN, S
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: GROSKE

DATE/TIME OF TEST: 07/12/2019 01:37:56

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.090	01:33
Breath Volume	1.593	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.085	01:36
Breath Volume	1.410	01:36
Air Blank	0.000	01:37
Control Test	0.079	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 07/29/2019 02:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/14/2019 18:42:27

Diagnostics Check	OK	18:36
Air Blank	0.000	18:36
Control Test	0.079	18:37
Air Blank	0.000	18:37
Subject Sample # 1	0.082	18:37
Breath Volume	2.914	18:37
Air Blank	0.000	18:38
Air Blank	0.000	18:40
Subject Sample # 2	0.077	18:40
Breath Volume	3.316	18:40
Air Blank	0.000	18:41
Control Test	0.079	18:41
Air Blank	0.000	18:42
Diagnostics Check	OK	18:42

Data Download Date/Time: 07/29/2019 02:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MCDONALD,C
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: GROSSMAN

DATE/TIME OF TEST: 07/22/2019 01:17:42

Diagnostics Check		OK	01:10
Air Blank		0.000	01:11
Control Test		0.080	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.126	01:12
Breath Volume		2.183	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.130	01:16
Breath Volume		2.562	01:16
Air Blank		0.000	01:16
Control Test		0.080	01:17
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 07/29/2019 02:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LOGAN,JONATHAN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 07/22/2019 23:08:24

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.079	23:03
Air Blank		0.000	23:03
Subject Sample	# 1	0.108	23:03
Breath Volume		2.019	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.102	23:06
Breath Volume		1.351	23:06
Air Blank		0.000	23:07
Control Test		0.079	23:07
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 07/29/2019 02:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN, JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: BERRY, VICTORIA
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: HUDDLESTON

DATE/TIME OF TEST: 07/23/2019 21:06:57

Diagnostics Check		OK	21:00
Air Blank		0.000	21:01
Control Test		0.079	21:01
Air Blank		0.000	21:02
Subject Sample	# 1	0.054	21:02
Breath Volume		2.890	21:02
Air Blank		0.000	21:03
Air Blank		0.000	21:04
Subject Sample	# 2	0.052	21:05
Breath Volume		2.515	21:05
Air Blank		0.000	21:06
Control Test		0.079	21:06
Air Blank		0.000	21:06
Diagnostics Check		OK	21:06

Data Download Date/Time: 07/29/2019 02:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: SMITH,MICHAEL
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 07/24/2019 01:30:10

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.079	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.170	01:25
Breath Volume	2.867	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.159	01:28
Breath Volume	3.636	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 07/29/2019 02:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: COZZOLINO

DATE/TIME OF TEST: 07/02/2019 20:16:08

Diagnostics Check	OK	20:13	
Air Blank	0.000	20:13	
Control Test	0.079	20:13	
Air Blank	0.000	20:14	
Subject Sample	# 1	20:14	Subject Test Refused
Breath Volume	0.000	20:14	
Air Blank	0.000	20:15	
Control Test	0.078	20:15	
Air Blank	0.000	20:16	
Diagnostics Check	OK	20:16	

Data Download Date/Time: 07/05/2019 16:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: MARES

DATE/TIME OF TEST: 07/05/2019 00:51:10

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.078	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.101	00:46
Breath Volume	1.101	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:48
Subject Sample # 2	0.103	00:49
Breath Volume	1.437	00:49
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 07/05/2019 16:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: THAYER

DATE/TIME OF TEST: 07/05/2019 02:21:08

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:18	
Control Test	0.079	02:19	
Air Blank	0.000	02:19	
Subject Sample	# 1	02:19	Subject Test Refused
Breath Volume	0.000	02:19	
Air Blank	0.000	02:20	
Control Test	0.078	02:20	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time: 07/05/2019 16:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: CAVALLO

DATE/TIME OF TEST: 07/13/2019 01:41:10

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:34
Subject Sample # 1	0.167	01:35
Breath Volume	1.277	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.157	01:39
Breath Volume	1.242	01:39
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 08/22/2019 17:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: CHEVALIER

DATE/TIME OF TEST: 07/14/2019 04:09:00

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.079	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.140	04:04
Breath Volume	1.644	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.139	04:07
Breath Volume	1.558	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:09

Data Download Date/Time: 08/22/2019 17:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKINNER,SEAN,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SKINNER,SEAN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: LEONE

DATE/TIME OF TEST: 07/16/2019 19:05:13

Diagnostics Check	OK	18:59
Air Blank	0.000	18:59
Control Test	0.078	18:59
Air Blank	0.000	19:00
Subject Sample # 1	0.000	19:00
Breath Volume	1.152	19:00
Air Blank	0.000	19:01
Air Blank	0.000	19:03
Subject Sample # 2	0.000	19:03
Breath Volume	1.808	19:03
Air Blank	0.000	19:04
Control Test	0.081	19:04
Air Blank	0.000	19:05
Diagnostics Check	OK	19:05

Data Download Date/Time: 08/22/2019 17:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/19/2019 23:29:03

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.079	23:23
Air Blank	0.000	23:24
Subject Sample # 1	0.000	23:24
Breath Volume	2.269	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.000	23:27
Breath Volume	2.378	23:27
Air Blank	0.000	23:28
Control Test	0.080	23:28
Air Blank	0.000	23:28
Diagnostics Check	OK	23:29

Data Download Date/Time: 08/22/2019 17:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: CLOTHIER

DATE/TIME OF TEST: 07/21/2019 01:14:14

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.079	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.157	01:07
Breath Volume	1.351	01:07
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.162	01:12
Breath Volume	1.519	01:12
Air Blank	0.000	01:13
Control Test	0.079	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 08/22/2019 17:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 07/21/2019 02:35:09

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.121	02:30
Breath Volume	2.976	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.119	02:33
Breath Volume	2.042	02:33
Air Blank	0.000	02:34
Control Test	0.079	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 08/22/2019 17:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUFF,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HUFF,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/21/2019 06:27:17

Diagnostics Check	OK	06:15	
Air Blank	0.000	06:15	
Control Test	0.078	06:16	
Air Blank	0.000	06:16	
Subject Sample	# 1	06:19	No Sample Provided
Breath Volume	0.000	06:19	
Air Blank	0.000	06:20	
Air Blank	0.000	06:22	
Subject Sample	# 2	0.151	06:22
Breath Volume	1.843	06:22	
Air Blank	0.000	06:23	
Air Blank	0.000	06:25	
Subject Sample	# 3	0.149	06:25
Breath Volume	2.667	06:25	
Air Blank	0.000	06:26	
Control Test	0.080	06:26	
Air Blank	0.000	06:27	
Diagnostics Check	OK	06:27	

Data Download Date/Time: 08/22/2019 17:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: WEST OHARA

DATE/TIME OF TEST: 07/25/2019 22:36:28

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.079	22:30
Air Blank	0.000	22:31
Subject Sample # 1	0.159	22:32
Breath Volume	1.906	22:32
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.167	22:35
Breath Volume	1.773	22:35
Air Blank	0.000	22:35
Control Test	0.079	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 08/22/2019 17:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GUY, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/05/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: ECKROTE

DATE/TIME OF TEST: 07/26/2019 23:57:18

Diagnostics Check	OK	23:54	
Air Blank	0.000	23:54	
Control Test	0.079	23:55	
Air Blank	0.000	23:55	
Subject Sample	# 1	23:55	Subject Test Refused
Breath Volume	0.000	23:55	
Air Blank	0.000	23:56	
Control Test	0.079	23:56	
Air Blank	0.000	23:57	
Diagnostics Check	OK	23:57	

Data Download Date/Time: 08/22/2019 17:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GARRISON,JOHN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PRICKETT

DATE/TIME OF TEST: 07/05/2019 03:23:40

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:21	
Control Test	0.079	03:21	
Air Blank	0.000	03:22	
Subject Sample	# 1	03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:22	
Control Test	0.079	03:23	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 08/21/2019 02:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,DAVID,J
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: MORALES,DAVID
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ELLER

DATE/TIME OF TEST: 07/06/2019 03:36:20

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.077	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.232	03:31
Breath Volume		1.777	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.236	03:34
Breath Volume		1.941	03:34
Air Blank		0.000	03:35
Control Test		0.078	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 07/29/2019 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,DAVID,J
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: MORALES,DAVID
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: VINCENT

DATE/TIME OF TEST: 07/07/2019 03:19:03

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:16	
Control Test	0.077	03:17	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:18	
Control Test	0.078	03:18	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 07/29/2019 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: TWENSEY

DATE/TIME OF TEST: 07/13/2019 23:00:25

Diagnostics Check	OK	22:57	
Air Blank	0.000	22:58	
Control Test	0.077	22:58	
Air Blank	0.000	22:58	
Subject Sample	# 1	22:59	Subject Test Refused
Breath Volume	0.000	22:59	
Air Blank	0.000	22:59	
Control Test	0.078	22:59	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:00	

Data Download Date/Time: 07/29/2019 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: SAN

DATE/TIME OF TEST: 07/14/2019 23:24:18

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.079	23:18
Air Blank	0.000	23:19
Subject Sample # 1	0.190	23:19
Breath Volume	1.523	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.190	23:22
Breath Volume	1.593	23:22
Air Blank	0.000	23:23
Control Test	0.077	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 07/29/2019 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,GAVIN,O
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/15/2019 02:42:59

Diagnostics Check		OK	02:33
Air Blank		0.000	02:33
Control Test		0.078	02:33
Air Blank		0.000	02:34
Subject Sample	# 1	0.304	02:34
Breath Volume		2.136	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.277	02:38
Breath Volume		1.625	02:38
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 3	0.271	02:41
Breath Volume		1.335	02:41
Air Blank		0.000	02:42
Control Test		0.076	02:42
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 07/29/2019 16:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEISTHAMMEL,SEAN,P
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: PICOTTE,CHASE
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: OLANO

DATE/TIME OF TEST: 07/26/2019 06:09:47

Diagnostics Check	OK	06:02
Air Blank	0.000	06:03
Control Test	0.078	06:03
Air Blank	0.000	06:04
Subject Sample # 1	0.192	06:04
Breath Volume	4.460	06:04
Air Blank	0.000	06:05
Air Blank	0.000	06:07
Subject Sample # 2	0.196	06:08
Breath Volume	4.375	06:08
Air Blank	0.000	06:08
Control Test	0.078	06:09
Air Blank	0.000	06:09
Diagnostics Check	OK	06:09

Data Download Date/Time: 07/29/2019 16:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 07/01/2019 01:50:07

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.040	01:45
Breath Volume	3.417	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.040	01:48
Breath Volume	3.121	01:48
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELDER,E,V
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TURPIN,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:33

SUBJECT'S LAST NAME: DAMATO

DATE/TIME OF TEST: 07/01/2019 16:59:48

Diagnostics Check	OK	16:53
Air Blank	0.000	16:54
Control Test	0.080	16:54
Air Blank	0.000	16:54
Subject Sample # 1	0.170	16:55
Breath Volume	1.542	16:55
Air Blank	0.000	16:55
Air Blank	0.000	16:57
Subject Sample # 2	0.163	16:58
Breath Volume	1.843	16:58
Air Blank	0.000	16:58
Control Test	0.080	16:59
Air Blank	0.000	16:59
Diagnostics Check	OK	16:59

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: BONETA

DATE/TIME OF TEST: 07/01/2019 22:14:32

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.079	22:08
Air Blank	0.000	22:09
Subject Sample # 1	0.160	22:09
Breath Volume	3.183	22:09
Air Blank	0.000	22:10
Air Blank	0.000	22:12
Subject Sample # 2	0.160	22:12
Breath Volume	3.140	22:12
Air Blank	0.000	22:13
Control Test	0.079	22:13
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MCBROOM,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:35

SUBJECT'S LAST NAME: REDISH

DATE/TIME OF TEST: 07/02/2019 09:59:29

Diagnostics Check	OK	09:56	
Air Blank	0.000	09:56	
Control Test	0.080	09:56	
Air Blank	0.000	09:57	
Subject Sample	# 1	09:57	Interferent Detect
Breath Volume	1.335	09:57	
Air Blank	0.000	09:58	
Control Test	0.079	09:58	
Air Blank	0.000	09:59	
Diagnostics Check	OK	09:59	

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MCBROOM,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:35

SUBJECT'S LAST NAME: REDISH

DATE/TIME OF TEST: 07/02/2019 10:07:27

Diagnostics Check	OK	10:04	
Air Blank	0.000	10:04	
Control Test	0.080	10:04	
Air Blank	0.000	10:05	
Subject Sample	# 1	10:05	Interferent Detect
Breath Volume	2.320	10:05	
Air Blank	0.000	10:06	
Control Test	0.080	10:06	
Air Blank	0.000	10:07	
Diagnostics Check	OK	10:07	

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MCBROOM,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:48

SUBJECT'S LAST NAME: REDISH

DATE/TIME OF TEST: 07/02/2019 11:14:58

Diagnostics Check	OK	11:08
Air Blank	0.000	11:09
Control Test	0.080	11:09
Air Blank	0.000	11:09
Subject Sample # 1	0.072	11:10
Breath Volume	2.613	11:10
Air Blank	0.000	11:11
Air Blank	0.000	11:12
Subject Sample # 2	0.066	11:13
Breath Volume	2.468	11:13
Air Blank	0.000	11:14
Control Test	0.079	11:14
Air Blank	0.000	11:14
Diagnostics Check	OK	11:14

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: KUNITSKIY

DATE/TIME OF TEST: 07/04/2019 20:18:40

Diagnostics Check	OK	20:15	
Air Blank	0.000	20:16	
Control Test	0.080	20:16	
Air Blank	0.000	20:17	
Subject Sample	# 1	20:17	Subject Test Refused
Breath Volume	0.000	20:17	
Air Blank	0.000	20:17	
Control Test	0.079	20:18	
Air Blank	0.000	20:18	
Diagnostics Check	OK	20:18	

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 07/05/2019 00:46:03

Diagnostics Check	OK	00:38
Air Blank	0.000	00:39
Control Test	0.079	00:39
Air Blank	0.000	00:40
Subject Sample # 1	0.203	00:41
Breath Volume	2.234	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:43
Subject Sample # 2	0.197	00:44
Breath Volume	1.582	00:44
Air Blank	0.000	00:45
Control Test	0.079	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:46

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ, M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: GRANELA

DATE/TIME OF TEST: 07/05/2019 01:24:22

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:18
Subject Sample # 1	0.102	01:19
Breath Volume	2.433	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.106	01:22
Breath Volume	3.152	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DIAZ, C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: SALADINO

DATE/TIME OF TEST: 07/05/2019 03:24:46

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.000	03:20
Breath Volume	3.554	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.000	03:23
Breath Volume	3.113	03:23
Air Blank	0.000	03:23
Control Test	0.080	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ, M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 07/05/2019 04:19:15

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.080	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.174	04:14
Breath Volume	2.324	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.172	04:17
Breath Volume	2.457	04:17
Air Blank	0.000	04:18
Control Test	0.079	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: BARR

DATE/TIME OF TEST: 07/06/2019 19:53:14

Diagnostics Check	OK	19:46
Air Blank	0.000	19:47
Control Test	0.079	19:47
Air Blank	0.000	19:48
Subject Sample # 1	0.162	19:48
Breath Volume	2.339	19:48
Air Blank	0.000	19:49
Air Blank	0.000	19:51
Subject Sample # 2	0.159	19:51
Breath Volume	1.531	19:51
Air Blank	0.000	19:52
Control Test	0.079	19:52
Air Blank	0.000	19:53
Diagnostics Check	OK	19:53

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER, JOSHUA, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: DYESS

DATE/TIME OF TEST: 07/06/2019 21:16:49

Diagnostics Check	OK	21:13	
Air Blank	0.000	21:14	
Control Test	0.079	21:14	
Air Blank	0.000	21:15	
Subject Sample	# 1	21:15	Subject Test Refused
Breath Volume	0.000	21:15	
Air Blank	0.000	21:15	
Control Test	0.079	21:16	
Air Blank	0.000	21:16	
Diagnostics Check	OK	21:16	

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: AKERS

DATE/TIME OF TEST: 07/07/2019 02:24:27

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.150	02:19
Breath Volume	2.367	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.154	02:22
Breath Volume	1.914	02:22
Air Blank	0.000	02:23
Control Test	0.078	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HYSMITH,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 07/07/2019 03:21:24

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.079	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.180	03:16
Breath Volume	2.988	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.187	03:19
Breath Volume	3.472	03:19
Air Blank	0.000	03:20
Control Test	0.079	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: ALBRIGHT

DATE/TIME OF TEST: 07/07/2019 04:24:31

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:23	
Control Test	0.080	04:23	
Air Blank		04:24	RFI Detect
Air Blank	0.000	04:24	

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: ALBRIGHT

DATE/TIME OF TEST: 07/07/2019 04:28:30

Diagnostics Check	OK	04:25	
Air Blank	0.000	04:26	
Control Test	0.079	04:26	
Air Blank	0.000	04:26	
Subject Sample	# 1	04:27	Subject Test Refused
Breath Volume	0.000	04:27	
Air Blank	0.000	04:27	
Control Test	0.079	04:27	
Air Blank	0.000	04:28	
Diagnostics Check	OK	04:28	

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLDEN,E,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:29

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 07/07/2019 12:52:55

Diagnostics Check	OK	12:50	
Air Blank	0.000	12:50	
Control Test	0.079	12:50	
Air Blank	0.000	12:51	
Subject Sample	# 1	12:51	Subject Test Refused
Breath Volume	0.000	12:51	
Air Blank	0.000	12:52	
Control Test	0.080	12:52	
Air Blank	0.000	12:52	
Diagnostics Check	OK	12:52	

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: SANG

DATE/TIME OF TEST: 07/07/2019 22:55:57

Diagnostics Check	OK	22:49
Air Blank	0.000	22:50
Control Test	0.080	22:50
Air Blank	0.000	22:51
Subject Sample # 1	0.154	22:51
Breath Volume	2.808	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:53
Subject Sample # 2	0.163	22:54
Breath Volume	2.800	22:54
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 07/08/2019 01:15:12

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:12	
Control Test	0.080	01:13	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:13	Subject Test Refused
Breath Volume	0.000	01:13	
Air Blank	0.000	01:14	
Control Test	0.081	01:14	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:15	

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BUSATTA,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: IER

DATE/TIME OF TEST: 07/08/2019 05:40:50

Diagnostics Check	OK	05:37	
Air Blank	0.000	05:38	
Control Test	0.081	05:38	
Air Blank	0.000	05:39	
Subject Sample	# 1	05:39	Subject Test Refused
Breath Volume	0.000	05:39	
Air Blank	0.000	05:39	
Control Test	0.081	05:40	
Air Blank	0.000	05:40	
Diagnostics Check	OK	05:40	

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: NOBORE,O
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/09/2019 03:33:50

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.081	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.157	03:29
Breath Volume	3.187	03:29
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.164	03:32
Breath Volume	3.554	03:32
Air Blank	0.000	03:32
Control Test	0.079	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PATTERSON,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 07/13/2019 04:59:06

Diagnostics Check		OK	04:52
Air Blank		0.000	04:53
Control Test		0.080	04:53
Air Blank		0.000	04:54
Subject Sample	# 1	0.100	04:54
Breath Volume		1.222	04:54
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.100	04:57
Breath Volume		1.687	04:57
Air Blank		0.000	04:58
Control Test		0.080	04:58
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELDER,E,V
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: OSTEEN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: ALDACO

DATE/TIME OF TEST: 07/16/2019 14:37:48

Diagnostics Check	OK	14:31
Air Blank	0.000	14:32
Control Test	0.080	14:32
Air Blank	0.000	14:32
Subject Sample # 1	0.045	14:33
Breath Volume	3.292	14:33
Air Blank	0.000	14:34
Air Blank	0.000	14:35
Subject Sample # 2	0.042	14:36
Breath Volume	1.664	14:36
Air Blank	0.000	14:36
Control Test	0.080	14:37
Air Blank	0.000	14:37
Diagnostics Check	OK	14:37

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: WARNOCK

DATE/TIME OF TEST: 07/16/2019 21:40:00

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.079	21:34
Air Blank	0.000	21:34
Subject Sample # 1	0.206	21:35
Breath Volume	1.570	21:35
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.203	21:38
Breath Volume	1.453	21:38
Air Blank	0.000	21:39
Control Test	0.080	21:39
Air Blank	0.000	21:39
Diagnostics Check	OK	21:40

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: GOLLIHUE

DATE/TIME OF TEST: 07/17/2019 05:19:36

Diagnostics Check		OK	05:13
Air Blank		0.000	05:13
Control Test		0.080	05:14
Air Blank		0.000	05:14
Subject Sample	# 1	0.204	05:15
Breath Volume		3.699	05:15
Air Blank		0.000	05:15
Air Blank		0.000	05:17
Subject Sample	# 2	0.208	05:17
Breath Volume		3.394	05:17
Air Blank		0.000	05:18
Control Test		0.079	05:18
Air Blank		0.000	05:19
Diagnostics Check		OK	05:19

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: PINKNEY

DATE/TIME OF TEST: 07/17/2019 23:33:33

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.080	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.012	23:29
Breath Volume	2.007	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.011	23:32
Breath Volume	2.335	23:32
Air Blank	0.000	23:32
Control Test	0.080	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 07/18/2019 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: TURPIN

DATE/TIME OF TEST: 07/18/2019 23:26:41

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.087	23:22
Breath Volume	2.335	23:22
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.084	23:25
Breath Volume	2.472	23:25
Air Blank	0.000	23:25
Control Test	0.080	23:26
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 07/30/2019 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: BOSTON

DATE/TIME OF TEST: 07/19/2019 03:22:54

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.078	03:18
Breath Volume	3.703	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.078	03:21
Breath Volume	4.269	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 07/30/2019 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELDER,E,V
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FLORES,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: DUKES

DATE/TIME OF TEST: 07/19/2019 07:52:19

Diagnostics Check	OK	07:45
Air Blank	0.000	07:46
Control Test	0.080	07:46
Air Blank	0.000	07:47
Subject Sample # 1	0.234	07:47
Breath Volume	3.195	07:47
Air Blank	0.000	07:48
Air Blank	0.000	07:50
Subject Sample # 2	0.228	07:50
Breath Volume	3.718	07:50
Air Blank	0.000	07:51
Control Test	0.079	07:51
Air Blank	0.000	07:52
Diagnostics Check	OK	07:52

Data Download Date/Time: 07/30/2019 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,A,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FRANZ,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:36

SUBJECT'S LAST NAME: NIXON

DATE/TIME OF TEST: 07/20/2019 08:03:32

Diagnostics Check	OK	07:56
Air Blank	0.000	07:57
Control Test	0.080	07:57
Air Blank	0.000	07:58
Subject Sample # 1	0.152	07:58
Breath Volume	1.128	07:58
Air Blank	0.000	07:59
Air Blank	0.000	08:01
Subject Sample # 2	0.150	08:01
Breath Volume	1.148	08:01
Air Blank	0.000	08:02
Control Test	0.079	08:02
Air Blank	0.000	08:03
Diagnostics Check	OK	08:03

Data Download Date/Time: 07/30/2019 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,A,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ACKERMAN,O
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: CARIVEAUX

DATE/TIME OF TEST: 07/20/2019 16:07:26

Diagnostics Check	OK	16:01
Air Blank	0.000	16:01
Control Test	0.080	16:01
Air Blank	0.000	16:02
Subject Sample # 1	0.000	16:02
Breath Volume	2.328	16:02
Air Blank	0.000	16:03
Air Blank	0.000	16:05
Subject Sample # 2	0.000	16:06
Breath Volume	2.156	16:06
Air Blank	0.000	16:06
Control Test	0.080	16:06
Air Blank	0.000	16:07
Diagnostics Check	OK	16:07

Data Download Date/Time: 07/30/2019 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,A,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:24

SUBJECT'S LAST NAME: WILCOX

DATE/TIME OF TEST: 07/24/2019 12:51:51

Diagnostics Check	OK	12:45
Air Blank	0.000	12:45
Control Test	0.081	12:46
Air Blank	0.000	12:46
Subject Sample # 1	0.181	12:47
Breath Volume	2.785	12:47
Air Blank	0.000	12:47
Air Blank	0.000	12:49
Subject Sample # 2	0.177	12:50
Breath Volume	2.656	12:50
Air Blank	0.000	12:50
Control Test	0.080	12:51
Air Blank	0.000	12:51
Diagnostics Check	OK	12:51

Data Download Date/Time: 07/30/2019 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,A,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MACK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:49

SUBJECT'S LAST NAME: SEABENECK

DATE/TIME OF TEST: 07/24/2019 15:13:35

Diagnostics Check	OK	15:10	
Air Blank	0.000	15:11	
Control Test	0.081	15:11	
Air Blank	0.000	15:11	
Subject Sample	# 1	15:12	Subject Test Refused
Breath Volume	0.000	15:12	
Air Blank	0.000	15:12	
Control Test	0.080	15:12	
Air Blank	0.000	15:13	
Diagnostics Check	OK	15:13	

Data Download Date/Time: 07/30/2019 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLDEN,E,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOYLE,B
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: YAKEL

DATE/TIME OF TEST: 07/25/2019 17:38:58

Diagnostics Check	OK	17:32
Air Blank	0.000	17:32
Control Test	0.080	17:33
Air Blank	0.000	17:33
Subject Sample # 1	0.201	17:34
Breath Volume	2.652	17:34
Air Blank	0.000	17:34
Air Blank	0.000	17:36
Subject Sample # 2	0.192	17:37
Breath Volume	1.964	17:37
Air Blank	0.000	17:38
Control Test	0.079	17:38
Air Blank	0.000	17:38
Diagnostics Check	OK	17:38

Data Download Date/Time: 07/30/2019 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HYSMITH,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: DESCHENES

DATE/TIME OF TEST: 07/26/2019 01:03:06

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:00	
Control Test	0.081	01:00	
Air Blank	0.000	01:01	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:02	
Control Test	0.080	01:02	
Air Blank	0.000	01:03	
Diagnostics Check	OK	01:03	

Data Download Date/Time: 07/30/2019 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: GIBBS

DATE/TIME OF TEST: 07/26/2019 22:40:15

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:37	
Control Test	0.080	22:38	
Air Blank	0.000	22:38	
Subject Sample	# 1	22:38	Subject Test Refused
Breath Volume	0.000	22:38	
Air Blank	0.000	22:39	
Control Test	0.081	22:39	
Air Blank	0.000	22:40	
Diagnostics Check	OK	22:40	

Data Download Date/Time: 07/30/2019 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: DELANCY

DATE/TIME OF TEST: 07/27/2019 03:01:58

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.081	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.108	02:57
Breath Volume	1.453	02:57
Air Blank	0.000	02:58
Air Blank	0.000	02:59
Subject Sample # 2	0.110	03:00
Breath Volume	1.933	03:00
Air Blank	0.000	03:01
Control Test	0.081	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 07/30/2019 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMAS,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 07/27/2019 23:11:32

Diagnostics Check	OK	23:04
Air Blank	0.000	23:05
Control Test	0.080	23:05
Air Blank	0.000	23:06
Subject Sample # 1	0.183	23:06
Breath Volume	3.863	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.183	23:09
Breath Volume	3.253	23:09
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 07/30/2019 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 07/29/2019 02:26:27

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.174	02:21
Breath Volume	2.046	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.168	02:24
Breath Volume	2.292	02:24
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 07/30/2019 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GRACKL,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: CREMEANS

DATE/TIME OF TEST: 07/27/2019 22:05:34

Diagnostics Check	OK	21:59
Air Blank	0.000	21:59
Control Test	0.081	22:00
Air Blank	0.000	22:00
Subject Sample # 1	0.000	22:01
Breath Volume	2.773	22:01
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.000	22:04
Breath Volume	2.292	22:04
Air Blank	0.000	22:04
Control Test	0.081	22:04
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 08/28/2019 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JENNINGS,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: SABOURIN

DATE/TIME OF TEST: 07/31/2019 10:34:39

Diagnostics Check	OK	10:20
Air Blank	0.000	10:21
Control Test	0.080	10:21
Air Blank	0.000	10:22
Subject Sample # 1	0.000	10:23
Breath Volume	1.546	10:23
Air Blank	0.000	10:23
Air Blank	0.000	10:32
Subject Sample # 2	0.009	10:33
Breath Volume	1.351	10:33
Air Blank	0.000	10:33
Control Test	0.085	10:34
Air Blank	0.000	10:34
Diagnostics Check	OK	10:34

Data Download Date/Time: 08/28/2019 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: MOYE

DATE/TIME OF TEST: 07/31/2019 23:14:38

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.080	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.170	23:09
Breath Volume	1.636	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:11
Subject Sample # 2	0.155	23:13
Breath Volume	1.140	23:13
Air Blank	0.000	23:13
Control Test	0.079	23:14
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 08/28/2019 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001048

INSTRUMENT REGISTERED TO: LAKE ALFRED PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GETTLE,ANTHONY,W
BREATH TEST OPERATOR AGENCY: Lake Alfred PD
ARRESTING OFFICER: GETTLE,ANTHONY
ARRESTING OFFICER AGENCY: Lake Alfred PD
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: PROB/PAROL
ARREST TIME: 20:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/18/2019 21:22:09

Diagnostics Check	OK	21:16
Air Blank	0.000	21:16
Control Test	0.081	21:17
Air Blank	0.000	21:17
Subject Sample # 1	0.000	21:18
Breath Volume	2.285	21:18
Air Blank	0.000	21:18
Air Blank	0.000	21:20
Subject Sample # 2	0.000	21:20
Breath Volume	3.109	21:20
Air Blank	0.000	21:21
Control Test	0.079	21:21
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 07/26/2019 18:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/18/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/01/2019 23:31:51

Diagnostics Check	OK	23:28	
Air Blank	0.000	23:28	
Control Test	0.080	23:28	
Air Blank	0.000	23:29	
Subject Sample	# 1	23:30	Subject Test Refused
Breath Volume	0.000	23:30	
Air Blank	0.000	23:30	
Control Test	0.080	23:31	
Air Blank	0.000	23:31	
Diagnostics Check	OK	23:31	

Data Download Date/Time: 07/13/2019 18:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONNELL,JACOB,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NARINE,NANDKISHORE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: PADEN

DATE/TIME OF TEST: 07/06/2019 00:04:46

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:56	
Control Test	0.080	23:57	
Air Blank	0.000	23:57	
Subject Sample	# 1	00:00	No Sample Provided
Breath Volume	0.000	00:00	
Air Blank	0.000	00:01	
Air Blank	0.000	00:03	
Subject Sample	# 2	00:03	Subject Test Refused
Breath Volume	0.000	00:03	
Air Blank	0.000	00:03	
Control Test	0.080	00:04	
Air Blank	0.000	00:04	
Diagnostics Check	OK	00:04	

Data Download Date/Time: 07/13/2019 18:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,SUMMER,N
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,SUMMER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: DAUENHAUER

DATE/TIME OF TEST: 07/14/2019 22:47:40

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.080	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.163	22:42
Breath Volume	2.433	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.157	22:46
Breath Volume	2.402	22:46
Air Blank	0.000	22:46
Control Test	0.080	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 09/26/2019 17:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONNELL,JACOB,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CONNELL,JACOB
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: LE

DATE/TIME OF TEST: 07/16/2019 02:56:30

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:48	
Control Test	0.080	02:48	
Air Blank	0.000	02:49	
Subject Sample	# 1	02:52	No Sample Provided
Breath Volume	0.000	02:52	
Air Blank	0.000	02:52	
Air Blank	0.000	02:54	
Subject Sample	# 2	02:55	Subject Test Refused
Breath Volume	0.000	02:55	
Air Blank	0.000	02:55	
Control Test	0.080	02:55	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

Data Download Date/Time: 09/26/2019 17:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONNELL,JACOB,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CLOUTIER,AUSTIN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 07/16/2019 04:24:21

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:16	
Control Test	0.081	04:16	
Air Blank	0.000	04:17	
Subject Sample	# 1	04:20	No Sample Provided
Breath Volume	0.000	04:20	
Air Blank	0.000	04:20	
Air Blank	0.000	04:22	
Subject Sample	# 2	04:23	Subject Test Refused
Breath Volume	0.000	04:23	
Air Blank	0.000	04:23	
Control Test	0.081	04:23	
Air Blank	0.000	04:24	
Diagnostics Check	OK	04:24	

Data Download Date/Time: 09/26/2019 17:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HAMMELMAN,TYLER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: VALENTI

DATE/TIME OF TEST: 07/18/2019 03:34:07

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.000	03:29
Breath Volume	4.710	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.000	03:32
Breath Volume	4.062	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 09/26/2019 17:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONNELL,JACOB,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CONNELL,JACOB
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 07/22/2019 01:58:52

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.079	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.000	01:54
Breath Volume	2.625	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.000	01:57
Breath Volume	2.109	01:57
Air Blank	0.000	01:57
Control Test	0.079	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 09/26/2019 17:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KABAN,SEAN,W
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: KABAN,SEAN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: REQUEJO

DATE/TIME OF TEST: 07/07/2019 23:01:08

Diagnostics Check	OK	22:54
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:55
Subject Sample # 1	0.000	22:56
Breath Volume	4.394	22:56
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.000	22:59
Breath Volume	4.589	22:59
Air Blank	0.000	23:00
Control Test	0.080	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 07/31/2019 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAYE, MAKINA, A
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: THOMAS, ALEX
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/09/2019 04:13:18

Diagnostics Check	OK	04:10	
Air Blank	0.000	04:10	
Control Test	0.079	04:10	
Air Blank	0.000	04:11	
Subject Sample	# 1	04:11	Subject Test Refused
Breath Volume	0.000	04:11	
Air Blank	0.000	04:12	
Control Test	0.079	04:12	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 07/31/2019 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,LUIS,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: FAIRWEATHER

DATE/TIME OF TEST: 07/02/2019 23:22:57

Diagnostics Check	OK	23:17	
Air Blank	0.000	23:18	
Control Test	0.080	23:18	
Air Blank	0.000	23:19	
Subject Sample	# 1	23:21	Subject Test Refused
Breath Volume	0.000	23:21	
Air Blank	0.000	23:22	
Control Test	0.081	23:22	
Air Blank	0.000	23:22	
Diagnostics Check	OK	23:22	

Data Download Date/Time: 07/11/2019 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LESPERANCE,DANIEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/06/2019 03:50:29

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.080	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.187	03:45
Breath Volume	2.550	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.177	03:49
Breath Volume	1.878	03:49
Air Blank	0.000	03:49
Control Test	0.077	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 07/11/2019 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: SABROSO

DATE/TIME OF TEST: 07/07/2019 04:58:04

Diagnostics Check	OK	04:51
Air Blank	0.000	04:51
Control Test	0.080	04:52
Air Blank	0.000	04:52
Subject Sample # 1	0.188	04:53
Breath Volume	1.843	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 2	0.178	04:56
Breath Volume	1.781	04:56
Air Blank	0.000	04:57
Control Test	0.080	04:57
Air Blank	0.000	04:57
Diagnostics Check	OK	04:58

Data Download Date/Time: 07/11/2019 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,CALVIN,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: CALLAGHAN

DATE/TIME OF TEST: 07/08/2019 21:22:55

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:18	
Control Test	0.080	21:18	
Air Blank	0.000	21:19	
Subject Sample	# 1	21:22	Sequence Aborted
Breath Volume	0.000	21:22	
Air Blank	0.000	21:22	

Data Download Date/Time: 07/11/2019 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,CALVIN,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: CALLAGHAN

DATE/TIME OF TEST: 07/08/2019 21:40:17

Diagnostics Check	OK	21:28		
Air Blank	0.000	21:29		
Control Test	0.080	21:29		
Air Blank	0.000	21:30		
Subject Sample	# 1	21:33	No Sample Provided	
Breath Volume	0.000	21:33		
Air Blank	0.000	21:33		
Air Blank	0.000	21:35		
Subject Sample	# 2	0.190	21:38	Volume Not Met
Breath Volume	0.394	21:38		
Air Blank	0.000	21:39		
Control Test	0.078	21:39		
Air Blank	0.000	21:40		
Diagnostics Check	OK	21:40		

Data Download Date/Time: 07/11/2019 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAIA,MICHAEL,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SAIA,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: MCABEE

DATE/TIME OF TEST: 07/09/2019 07:03:48

Diagnostics Check	OK	06:56	
Air Blank	0.000	06:56	
Control Test	0.079	06:57	
Air Blank	0.000	06:57	
Subject Sample	# 1 0.137	07:00	Volume Not Met
Breath Volume	0.593	07:00	
Air Blank	0.000	07:01	
Air Blank		07:03	Ambient Fail
Air Blank	0.000	07:03	

Data Download Date/Time: 07/11/2019 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LESPERANCE,DANIEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/11/2019 19:44:18

Diagnostics Check	OK	19:40	
Air Blank	0.000	19:41	
Control Test	0.081	19:41	
Air Blank	0.000	19:42	
Subject Sample	# 1	19:42	Interferent Detect
Breath Volume	1.667	19:42	
Air Blank	0.000	19:43	
Control Test	0.038	19:43	Interferent Detect
Air Blank	0.000	19:44	

Data Download Date/Time: 08/05/2019 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LESPERANCE,DANIEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/11/2019 19:48:21

Diagnostics Check	OK	19:46	
Air Blank	0.000	19:46	
Control Test	0.080	19:47	
Air Blank	0.000	19:47	
Subject Sample	# 1	19:47	Sequence Aborted
Breath Volume	0.000	19:47	
Air Blank	0.000	19:48	

Data Download Date/Time: 08/05/2019 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LESPERANCE,DANIEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/11/2019 20:12:40

Diagnostics Check	OK	20:06
Air Blank	0.000	20:06
Control Test	0.080	20:07
Air Blank	0.000	20:07
Subject Sample # 1	0.149	20:08
Breath Volume	2.371	20:08
Air Blank	0.000	20:08
Air Blank	0.000	20:10
Subject Sample # 2	0.144	20:11
Breath Volume	1.941	20:11
Air Blank	0.000	20:11
Control Test	0.080	20:12
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

Data Download Date/Time: 08/05/2019 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: WEBER

DATE/TIME OF TEST: 07/12/2019 01:22:53

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:11	
Control Test	0.080	01:12	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:15	No Sample Provided
Breath Volume	0.000	01:15	
Air Blank	0.000	01:16	
Air Blank	0.000	01:18	
Subject Sample	# 2	01:21	No Sample Provided
Breath Volume	0.000	01:21	
Air Blank	0.000	01:21	
Control Test	0.081	01:22	
Air Blank	0.000	01:22	
Diagnostics Check	OK	01:22	

Data Download Date/Time: 08/05/2019 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE, RONALD,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS, KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: WEITKEMPER

DATE/TIME OF TEST: 07/12/2019 18:33:02

Diagnostics Check	OK	18:26
Air Blank	0.000	18:26
Control Test	0.081	18:27
Air Blank	0.000	18:27
Subject Sample # 1	0.305	18:28
Breath Volume	2.339	18:28
Air Blank	0.000	18:29
Air Blank	0.000	18:30
Subject Sample # 2	0.300	18:31
Breath Volume	2.605	18:31
Air Blank	0.000	18:32
Control Test	0.077	18:32
Air Blank	0.000	18:32
Diagnostics Check	OK	18:33

Data Download Date/Time: 08/05/2019 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: BLOSSER

DATE/TIME OF TEST: 07/12/2019 20:52:20

Diagnostics Check	OK	20:41
Air Blank	0.000	20:42
Control Test	0.080	20:42
Air Blank	0.000	20:43
Subject Sample # 1	0.215	20:43
Breath Volume	2.039	20:43
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.192	20:47
Breath Volume	1.597	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 3	0.208	20:50
Breath Volume	2.148	20:50
Air Blank	0.000	20:51
Control Test	0.076	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 08/05/2019 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/13/2019 02:44:41

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:40	
Control Test	0.080	02:41	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:43	Subject Test Refused
Breath Volume	0.000	02:43	
Air Blank	0.000	02:43	
Control Test	0.080	02:44	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 08/05/2019 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAIA,MICHAEL,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HOBART,CLAYTON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 07/16/2019 00:39:08

Diagnostics Check	OK	00:32
Air Blank	0.000	00:33
Control Test	0.081	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.252	00:34
Breath Volume	1.351	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:36
Subject Sample # 2	0.247	00:37
Breath Volume	2.804	00:37
Air Blank	0.000	00:38
Control Test	0.078	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 08/05/2019 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,Z
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BRISACH

DATE/TIME OF TEST: 07/18/2019 01:14:05

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:11	
Control Test	0.080	01:11	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:13	
Control Test	0.080	01:13	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

Data Download Date/Time: 08/05/2019 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER, MATTHEW, R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER, MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: BARGMAN

DATE/TIME OF TEST: 07/18/2019 17:58:51

Diagnostics Check	OK	17:52
Air Blank	0.000	17:52
Control Test	0.079	17:52
Air Blank	0.000	17:53
Subject Sample # 1	0.076	17:54
Breath Volume	1.292	17:54
Air Blank	0.000	17:54
Air Blank	0.000	17:56
Subject Sample # 2	0.080	17:57
Breath Volume	1.265	17:57
Air Blank	0.000	17:57
Control Test	0.080	17:58
Air Blank	0.000	17:58
Diagnostics Check	OK	17:58

Data Download Date/Time: 08/05/2019 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: PAINTER

DATE/TIME OF TEST: 07/19/2019 23:41:36

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.080	23:33
Air Blank	0.000	23:34
Subject Sample # 1	0.186	23:37
Breath Volume	1.289	23:37
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.198	23:40
Breath Volume	1.394	23:40
Air Blank	0.000	23:40
Control Test	0.079	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 08/05/2019 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESPERANCE,DANIEL,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 07/20/2019 00:53:07

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.079	00:45
Air Blank		0.000	00:46
Subject Sample	# 1	0.251	00:48
Breath Volume		1.464	00:48
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.253	00:51
Breath Volume		2.238	00:51
Air Blank		0.000	00:52
Control Test		0.078	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 08/05/2019 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/25/2019 00:24:38

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.080	00:19
Air Blank	0.000	00:19
Subject Sample # 1	0.000	00:20
Breath Volume	2.183	00:20
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.000	00:23
Breath Volume	2.640	00:23
Air Blank	0.000	00:23
Control Test	0.080	00:24
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 08/05/2019 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: PRUITT

DATE/TIME OF TEST: 07/25/2019 21:34:56

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:26	
Control Test	0.080	21:26	
Air Blank	0.000	21:27	
Subject Sample	# 1	21:30	No Sample Provided
Breath Volume	0.000	21:30	
Air Blank	0.000	21:30	
Air Blank	0.000	21:32	
Subject Sample	# 2	21:33	Subject Test Refused
Breath Volume	0.000	21:33	
Air Blank	0.000	21:34	
Control Test	0.080	21:34	
Air Blank	0.000	21:34	
Diagnostics Check	OK	21:34	

Data Download Date/Time: 08/05/2019 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: PECK

DATE/TIME OF TEST: 07/25/2019 22:20:22

Diagnostics Check	OK	22:14	
Air Blank	0.000	22:14	
Control Test	0.080	22:14	
Air Blank	0.000	22:15	
Subject Sample	# 1 0.158	22:18	Volume Not Met
Breath Volume	0.371	22:18	
Air Blank	0.000	22:19	
Air Blank	0.000	22:20	

Data Download Date/Time: 08/05/2019 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 07/29/2019 02:35:48

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:24	
Control Test	0.079	02:24	
Air Blank	0.000	02:25	
Subject Sample	# 1	02:28	No Sample Provided
Breath Volume	0.000	02:28	
Air Blank	0.000	02:29	
Air Blank	0.000	02:31	
Subject Sample	# 2	02:34	Volume Not Met
Breath Volume	0.183	02:34	
Air Blank	0.000	02:34	
Control Test	0.080	02:35	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 08/05/2019 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,B,K
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GOETHE,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: YAKUNOVICH

DATE/TIME OF TEST: 07/01/2019 04:06:25

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:02	
Control Test	0.081	04:02	
Air Blank	0.000	04:03	
Subject Sample	# 1	04:05	Subject Test Refused
Breath Volume	0.000	04:05	
Air Blank	0.000	04:05	
Control Test	0.081	04:05	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 07/21/2019 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,E,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: KLINE

DATE/TIME OF TEST: 07/02/2019 00:05:57

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:03	
Control Test	0.080	00:03	
Air Blank	0.000	00:04	
Subject Sample	# 1	00:04	Subject Test Refused
Breath Volume	0.000	00:04	
Air Blank	0.000	00:05	
Control Test	0.080	00:05	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:05	

Data Download Date/Time: 07/21/2019 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,CATHERINE,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WIGGINS,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: GAULT

DATE/TIME OF TEST: 07/04/2019 04:57:47

Diagnostics Check	OK	04:51
Air Blank	0.000	04:51
Control Test	0.081	04:52
Air Blank	0.000	04:52
Subject Sample # 1	0.000	04:53
Breath Volume	2.203	04:53
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.000	04:56
Breath Volume	1.386	04:56
Air Blank	0.000	04:56
Control Test	0.081	04:57
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 07/21/2019 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,E,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: KUYPERS

DATE/TIME OF TEST: 07/06/2019 02:56:38

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:54	
Control Test	0.081	02:54	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:55	Subject Test Refused
Breath Volume	0.000	02:55	
Air Blank	0.000	02:55	
Control Test	0.081	02:56	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

Data Download Date/Time: 07/21/2019 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HOLLAND,B
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: DUKE

DATE/TIME OF TEST: 07/06/2019 03:46:17

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.081	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.000	03:41
Breath Volume	1.109	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.000	03:44
Breath Volume	1.812	03:44
Air Blank	0.000	03:45
Control Test	0.081	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 07/21/2019 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HOLLAND,B
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: DYER

DATE/TIME OF TEST: 07/07/2019 03:52:14

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.092	03:47
Breath Volume	1.941	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.085	03:50
Breath Volume	1.789	03:50
Air Blank	0.000	03:51
Control Test	0.081	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 07/21/2019 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,A,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SANDIFER,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 07/13/2019 05:33:44

Diagnostics Check	OK	05:21	
Air Blank	0.000	05:21	
Control Test	0.080	05:22	
Air Blank	0.000	05:22	
Subject Sample # 1	0.000	05:23	
Breath Volume	2.949	05:23	
Air Blank	0.000	05:24	
Air Blank	0.000	05:26	
Subject Sample # 2		05:29	No Sample Provided
Breath Volume	0.000	05:29	
Air Blank	0.000	05:29	
Air Blank	0.000	05:31	
Subject Sample # 3	0.000	05:32	
Breath Volume	2.457	05:32	
Air Blank	0.000	05:32	
Control Test	0.081	05:33	
Air Blank	0.000	05:33	
Diagnostics Check	OK	05:33	

Data Download Date/Time: 07/21/2019 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIGGINBOTHAM,B,K
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GORDON,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: ALBERTIE

DATE/TIME OF TEST: 07/14/2019 01:39:49

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:37	
Control Test	0.081	01:37	
Air Blank	0.000	01:38	
Subject Sample	# 1	01:38	Subject Test Refused
Breath Volume	0.000	01:38	
Air Blank	0.000	01:38	
Control Test	0.082	01:39	
Air Blank	0.000	01:39	
Diagnostics Check	OK	01:39	

Data Download Date/Time: 07/21/2019 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,E,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WIGGINS,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 07/15/2019 20:58:44

Diagnostics Check	OK	20:50
Air Blank	0.000	20:50
Control Test	0.080	20:51
Air Blank	0.000	20:51
Subject Sample # 1	0.276	20:53
Breath Volume	1.714	20:53
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample # 2	0.292	20:57
Breath Volume	1.226	20:57
Air Blank	0.000	20:57
Control Test	0.078	20:58
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 07/21/2019 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIS,A,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BALLARD,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: MCCORKEL

DATE/TIME OF TEST: 07/18/2019 22:26:27

Diagnostics Check	OK	22:23	
Air Blank	0.000	22:24	
Control Test	0.082	22:24	
Air Blank	0.000	22:24	
Subject Sample	# 1	22:25	Subject Test Refused
Breath Volume	0.000	22:25	
Air Blank	0.000	22:25	
Control Test	0.081	22:25	
Air Blank	0.000	22:26	
Diagnostics Check	OK	22:26	

Data Download Date/Time: 07/21/2019 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: WHORTON

DATE/TIME OF TEST: 07/20/2019 03:22:02

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:19	
Control Test	0.081	03:19	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:21	
Control Test	0.080	03:21	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 07/21/2019 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,E,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: COPHER,C
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 07/21/2019 02:37:21

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:34	
Control Test	0.081	02:35	
Air Blank	0.000	02:35	
Subject Sample	# 1	02:36	Subject Test Refused
Breath Volume	0.000	02:36	
Air Blank	0.000	02:36	
Control Test	0.081	02:36	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 07/21/2019 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELACRUZ,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/21/2019 04:17:36

Diagnostics Check		OK	04:11
Air Blank		0.000	04:11
Control Test		0.080	04:12
Air Blank		0.000	04:12
Subject Sample	# 1	0.146	04:13
Breath Volume		2.312	04:13
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.146	04:16
Breath Volume		2.714	04:16
Air Blank		0.000	04:16
Control Test		0.081	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 07/21/2019 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON',DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WHITTINGTON,B
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/02/2019 00:41:53

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.079	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.075	00:37
Breath Volume	2.410	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.073	00:40
Breath Volume	1.937	00:40
Air Blank	0.000	00:40
Control Test	0.080	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 07/04/2019 18:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WELLS,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: MAKEPEACE

DATE/TIME OF TEST: 07/02/2019 02:15:00

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.079	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.159	02:10
Breath Volume	2.355	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.145	02:13
Breath Volume	1.101	02:13
Air Blank	0.000	02:14
Control Test	0.079	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:15

Data Download Date/Time: 07/04/2019 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,K,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: HOLMES,K
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: COURT ORD
ARREST TIME: 13:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:12

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 07/03/2019 13:43:31

Diagnostics Check	OK	13:37
Air Blank	0.000	13:37
Control Test	0.080	13:38
Air Blank	0.000	13:38
Subject Sample # 1	0.000	13:39
Breath Volume	2.378	13:39
Air Blank	0.000	13:39
Air Blank	0.000	13:41
Subject Sample # 2	0.000	13:42
Breath Volume	2.144	13:42
Air Blank	0.000	13:42
Control Test	0.081	13:42
Air Blank	0.000	13:43
Diagnostics Check	OK	13:43

Data Download Date/Time: 07/04/2019 18:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,K,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: HOLMES,K
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: COURT ORD
ARREST TIME: 13:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:48

SUBJECT'S LAST NAME: HOOKS

DATE/TIME OF TEST: 07/03/2019 14:19:20

Diagnostics Check	OK	14:13
Air Blank	0.000	14:13
Control Test	0.080	14:13
Air Blank	0.000	14:14
Subject Sample # 1	0.000	14:14
Breath Volume	1.789	14:14
Air Blank	0.000	14:15
Air Blank	0.000	14:17
Subject Sample # 2	0.000	14:17
Breath Volume	1.570	14:17
Air Blank	0.000	14:18
Control Test	0.080	14:18
Air Blank	0.000	14:19
Diagnostics Check	OK	14:19

Data Download Date/Time: 07/04/2019 18:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,K,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: HOLMES,K
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: COURT ORD
ARREST TIME: 14:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:01

SUBJECT'S LAST NAME: EDLOW

DATE/TIME OF TEST: 07/03/2019 14:31:27

Diagnostics Check	OK	14:25
Air Blank	0.000	14:25
Control Test	0.080	14:25
Air Blank	0.000	14:26
Subject Sample # 1	0.000	14:26
Breath Volume	1.878	14:26
Air Blank	0.000	14:27
Air Blank	0.000	14:29
Subject Sample # 2	0.000	14:30
Breath Volume	2.835	14:30
Air Blank	0.000	14:30
Control Test	0.080	14:30
Air Blank	0.000	14:31
Diagnostics Check	OK	14:31

Data Download Date/Time: 07/04/2019 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,K,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: HOLMES,K
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: COURT ORD
ARREST TIME: 15:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: ALBRITTON

DATE/TIME OF TEST: 07/03/2019 15:30:37

Diagnostics Check	OK	15:24
Air Blank	0.000	15:24
Control Test	0.080	15:25
Air Blank	0.000	15:25
Subject Sample # 1	0.000	15:26
Breath Volume	2.121	15:26
Air Blank	0.000	15:26
Air Blank	0.000	15:28
Subject Sample # 2	0.000	15:29
Breath Volume	2.398	15:29
Air Blank	0.000	15:29
Control Test	0.080	15:30
Air Blank	0.000	15:30
Diagnostics Check	OK	15:30

Data Download Date/Time: 07/04/2019 18:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: HINISH,BARRY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 06/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: COURT ORD
ARREST TIME: 19:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: JOHN

DATE/TIME OF TEST: 07/03/2019 20:08:45

Diagnostics Check	OK	20:02
Air Blank	0.000	20:02
Control Test	0.080	20:03
Air Blank	0.000	20:03
Subject Sample # 1	0.000	20:04
Breath Volume	1.691	20:04
Air Blank	0.000	20:04
Air Blank	0.000	20:06
Subject Sample # 2	0.000	20:07
Breath Volume	2.828	20:07
Air Blank	0.000	20:07
Control Test	0.080	20:08
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 07/04/2019 18:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: HODGE,KYLER
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 06/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: FLAHERTY

DATE/TIME OF TEST: 07/04/2019 00:30:40

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:27	
Control Test	0.080	00:27	
Air Blank	0.000	00:28	
Subject Sample	# 1	00:29	Subject Test Refused
Breath Volume	0.000	00:29	
Air Blank	0.000	00:29	
Control Test	0.078	00:30	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:30	

Data Download Date/Time: 07/04/2019 18:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BEUFORD,SEAN
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 06/15/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: WISDOM

DATE/TIME OF TEST: 07/04/2019 02:10:34

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.080	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.102	02:06
Breath Volume	1.644	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.098	02:09
Breath Volume	2.613	02:09
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 07/04/2019 18:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: STEINER

DATE/TIME OF TEST: 07/07/2019 00:16:57

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:13	
Control Test	0.079	00:14	
Air Blank	0.000	00:14	
Subject Sample	# 1	00:15	Subject Test Refused
Breath Volume	0.000	00:15	
Air Blank	0.000	00:16	
Control Test	0.079	00:16	
Air Blank	0.000	00:16	
Diagnostics Check	OK	00:16	

Data Download Date/Time: 08/21/2019 01:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,J,V
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: ENNIS

DATE/TIME OF TEST: 07/07/2019 23:51:06

Diagnostics Check	OK	23:48	
Air Blank	0.000	23:48	
Control Test	0.080	23:48	
Air Blank	0.000	23:49	
Subject Sample	# 1	23:49	Subject Test Refused
Breath Volume	0.000	23:49	
Air Blank	0.000	23:50	
Control Test	0.078	23:50	
Air Blank	0.000	23:51	
Diagnostics Check	OK	23:51	

Data Download Date/Time: 08/21/2019 01:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRACKETT,MARK,O
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BRACKETT,MARK
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: ZERO TOL
ARREST TIME: 11:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: PETILL

DATE/TIME OF TEST: 07/08/2019 12:17:43

Diagnostics Check	OK	12:11
Air Blank	0.000	12:11
Control Test	0.080	12:11
Air Blank	0.000	12:12
Subject Sample # 1	0.000	12:13
Breath Volume	2.136	12:13
Air Blank	0.000	12:13
Air Blank	0.000	12:15
Subject Sample # 2	0.000	12:16
Breath Volume	2.554	12:16
Air Blank	0.000	12:16
Control Test	0.080	12:17
Air Blank	0.000	12:17
Diagnostics Check	OK	12:17

Data Download Date/Time: 08/21/2019 01:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN,SHON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ABELN,SHON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: COURT ORD
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 07/11/2019 14:34:06

Diagnostics Check	OK	14:28
Air Blank	0.000	14:28
Control Test	0.079	14:29
Air Blank	0.000	14:29
Subject Sample # 1	0.000	14:29
Breath Volume	3.039	14:29
Air Blank	0.000	14:30
Air Blank	0.000	14:32
Subject Sample # 2	0.000	14:32
Breath Volume	3.046	14:32
Air Blank	0.000	14:33
Control Test	0.080	14:33
Air Blank	0.000	14:34
Diagnostics Check	OK	14:34

Data Download Date/Time: 08/21/2019 01:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PANTOJA,YOSMANY
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 07/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: HAEHLE

DATE/TIME OF TEST: 07/16/2019 19:38:40

Diagnostics Check	OK	19:35	
Air Blank	0.000	19:36	
Control Test	0.080	19:36	
Air Blank	0.000	19:37	
Subject Sample	# 1	19:37	Subject Test Refused
Breath Volume	0.000	19:37	
Air Blank	0.000	19:37	
Control Test	0.079	19:38	
Air Blank	0.000	19:38	
Diagnostics Check	OK	19:38	

Data Download Date/Time: 08/21/2019 01:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ABELN,SETH
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 07/17/2019 03:48:27

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:45	
Control Test	0.080	03:46	
Air Blank	0.000	03:46	
Subject Sample	# 1	03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:47	
Control Test	0.079	03:47	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 08/21/2019 01:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BUEFORD,S
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 07/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: GENDRON

DATE/TIME OF TEST: 07/19/2019 06:13:18

Diagnostics Check	OK	06:10	
Air Blank	0.000	06:10	
Control Test	0.080	06:11	
Air Blank	0.000	06:11	
Subject Sample	# 1	06:11	Subject Test Refused
Breath Volume	0.000	06:11	
Air Blank	0.000	06:12	
Control Test	0.079	06:12	
Air Blank	0.000	06:13	
Diagnostics Check	OK	06:13	

Data Download Date/Time: 08/21/2019 01:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BROCK,ALEXANDRA
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: NOTTINGHAM

DATE/TIME OF TEST: 07/21/2019 02:17:12

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:14	
Control Test	0.079	02:14	
Air Blank	0.000	02:15	
Subject Sample	# 1	02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:16	
Control Test	0.080	02:16	
Air Blank	0.000	02:17	
Diagnostics Check	OK	02:17	

Data Download Date/Time: 08/21/2019 01:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRACKETT,MARK,O
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BRACKETT,MARK
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: ZERO TOL
ARREST TIME: 09:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/22/2019 10:17:57

Diagnostics Check	OK	10:11
Air Blank	0.000	10:11
Control Test	0.079	10:11
Air Blank	0.000	10:12
Subject Sample # 1	0.000	10:13
Breath Volume	2.523	10:13
Air Blank	0.000	10:14
Air Blank	0.000	10:16
Subject Sample # 2	0.000	10:16
Breath Volume	2.890	10:16
Air Blank	0.000	10:17
Control Test	0.079	10:17
Air Blank	0.000	10:17
Diagnostics Check	OK	10:17

Data Download Date/Time: 08/21/2019 01:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGOODIAL,TRISTON,T
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: CARRANZA,RUBEN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: CROTEAU

DATE/TIME OF TEST: 07/25/2019 14:31:34

Diagnostics Check	OK	14:24
Air Blank	0.000	14:25
Control Test	0.079	14:25
Air Blank	0.000	14:25
Subject Sample # 1	0.056	14:26
Breath Volume	1.285	14:26
Air Blank	0.000	14:27
Air Blank	0.000	14:29
Subject Sample # 2	0.059	14:30
Breath Volume	1.710	14:30
Air Blank	0.000	14:30
Control Test	0.079	14:30
Air Blank	0.000	14:31
Diagnostics Check	OK	14:31

Data Download Date/Time: 08/21/2019 01:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,J,V
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: NONE,NONE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: OTHER
ARREST TIME: 18:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: JOHN

DATE/TIME OF TEST: 07/25/2019 19:00:51

Diagnostics Check	OK	18:54
Air Blank	0.000	18:55
Control Test	0.079	18:55
Air Blank	0.000	18:56
Subject Sample # 1	0.000	18:56
Breath Volume	2.253	18:56
Air Blank	0.000	18:57
Air Blank	0.000	18:59
Subject Sample # 2	0.000	18:59
Breath Volume	2.144	18:59
Air Blank	0.000	18:59
Control Test	0.080	19:00
Air Blank	0.000	19:00
Diagnostics Check	OK	19:00

Data Download Date/Time: 08/21/2019 01:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNELL,C,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SNELL,C
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: COURT ORD
ARREST TIME: 11:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 07/26/2019 11:36:47

Diagnostics Check	OK	11:30
Air Blank	0.000	11:31
Control Test	0.080	11:31
Air Blank	0.000	11:31
Subject Sample # 1	0.000	11:32
Breath Volume	1.738	11:32
Air Blank	0.000	11:32
Air Blank	0.000	11:34
Subject Sample # 2	0.000	11:35
Breath Volume	1.839	11:35
Air Blank	0.000	11:35
Control Test	0.080	11:36
Air Blank	0.000	11:36
Diagnostics Check	OK	11:36

Data Download Date/Time: 08/21/2019 01:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BOGUSKY,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 07/28/2019 01:52:03

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.000	01:47
Breath Volume	2.593	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.000	01:50
Breath Volume	3.324	01:50
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:52

Data Download Date/Time: 08/21/2019 01:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GARRISON,JOHN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 07/28/2019 03:07:30

Diagnostics Check	OK	03:04	
Air Blank	0.000	03:05	
Control Test	0.080	03:05	
Air Blank	0.000	03:05	
Subject Sample	# 1	03:06	Subject Test Refused
Breath Volume	0.000	03:06	
Air Blank	0.000	03:06	
Control Test	0.080	03:06	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

Data Download Date/Time: 08/21/2019 01:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRACKETT,MARK,O
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BRACKETT,MARK
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/04/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: ZERO TOL
ARREST TIME: 13:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: SAMPLES

DATE/TIME OF TEST: 07/31/2019 14:20:39

Diagnostics Check	OK	14:14
Air Blank	0.000	14:14
Control Test	0.080	14:15
Air Blank	0.000	14:15
Subject Sample # 1	0.000	14:16
Breath Volume	1.515	14:16
Air Blank	0.000	14:16
Air Blank	0.000	14:18
Subject Sample # 2	0.000	14:19
Breath Volume	2.132	14:19
Air Blank	0.000	14:19
Control Test	0.080	14:20
Air Blank	0.000	14:20
Diagnostics Check	OK	14:20

Data Download Date/Time: 08/21/2019 01:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA, MATTHEW, C
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: CORREA, MATTHEW
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 07/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MASKE

DATE/TIME OF TEST: 07/16/2019 23:49:45

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.080	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.172	23:45
Breath Volume	2.195	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.173	23:48
Breath Volume	2.859	23:48
Air Blank	0.000	23:48
Control Test	0.082	23:49
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 08/08/2019 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAINES,DAVID,K
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: YOUNG,JOSHUA
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 07/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: RANDALL

DATE/TIME OF TEST: 07/20/2019 01:35:17

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.082	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.180	01:30
Breath Volume	3.328	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.182	01:33
Breath Volume	3.414	01:33
Air Blank	0.000	01:34
Control Test	0.081	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 08/08/2019 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAHAM,HUGH,E
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: CARRINGTON,LEANNA
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 07/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 07/22/2019 05:17:51

Diagnostics Check	OK	05:11
Air Blank	0.000	05:12
Control Test	0.081	05:12
Air Blank	0.000	05:13
Subject Sample # 1	0.108	05:13
Breath Volume	2.132	05:13
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.113	05:16
Breath Volume	1.773	05:16
Air Blank	0.000	05:16
Control Test	0.081	05:17
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 08/08/2019 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAHAM,HUGH,E
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: GRAHAM,HUGH
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 07/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: MARTEL

DATE/TIME OF TEST: 07/27/2019 04:04:07

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.081	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.181	03:59
Breath Volume	1.601	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.172	04:02
Breath Volume	2.375	04:02
Air Blank	0.000	04:03
Control Test	0.078	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 08/08/2019 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAHAM,HUGH,E
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: CORREA,MATTHEW
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 07/09/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG827002
DRY GAS EXPIRY DATE: 09/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 07/29/2019 02:23:52

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:12	
Control Test	0.080	02:13	
Air Blank	0.000	02:13	
Subject Sample	# 1	02:16	No Sample Provided
Breath Volume	0.000	02:16	
Air Blank	0.000	02:17	
Air Blank	0.000	02:19	
Subject Sample	# 2	02:22	No Sample Provided
Breath Volume	0.000	02:22	
Air Blank	0.000	02:22	
Control Test	0.080	02:23	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

Data Download Date/Time: 08/08/2019 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAROTTA,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: CLAGUE,ERIC
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 06/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 07/06/2019 00:56:18

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:46	
Control Test	0.079	00:46	
Air Blank	0.000	00:47	
Subject Sample	# 1 0.175	00:50	Volume Not Met
Breath Volume	0.617	00:50	
Air Blank	0.000	00:50	
Air Blank	0.000	00:52	
Subject Sample	# 2	00:55	Sequence Aborted
Breath Volume	0.000	00:55	
Air Blank	0.000	00:56	

Data Download Date/Time: 07/22/2019 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAROTTA, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: CLAGUE, ERIC
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 06/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 07/06/2019 01:02:27

Diagnostics Check	OK	00:59	
Air Blank	0.000	00:59	
Control Test	0.078	01:00	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:01	
Control Test	0.078	01:01	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 07/22/2019 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,ZACHARY,B
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: CLAGUE,ERIC
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 06/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: TELATYCKI

DATE/TIME OF TEST: 07/11/2019 23:28:42

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:26	
Control Test	0.079	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 1	23:27	Subject Test Refused
Breath Volume	0.000	23:27	
Air Blank	0.000	23:27	
Control Test	0.079	23:28	
Air Blank	0.000	23:28	
Diagnostics Check	OK	23:28	

Data Download Date/Time: 07/22/2019 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CARTER,TIMOTHY,R

BREATH TEST OPERATOR AGENCY: Gulfport PD

ARRESTING OFFICER: CARTER,TIMOTHY

ARRESTING OFFICER AGENCY: Gulfport PD

LAST AGENCY INSPECTION: 06/17/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17817080A2

DRY GAS EXPIRY DATE: 08/05/2019

VIOLATION CODE: DUI

ARREST TIME: 02:56

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: KHITLIK

DATE/TIME OF TEST: 07/14/2019 03:39:35

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.080	03:35	
Air Blank	0.000	03:36	
Subject Sample	# 1	03:38	Subject Test Refused
Breath Volume	0.000	03:38	
Air Blank	0.000	03:38	
Control Test	0.079	03:39	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 07/22/2019 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: CARTER,TIMOTHY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 06/17/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/15/2019 20:49:05

Diagnostics Check		OK	20:42
Air Blank		0.000	20:42
Control Test		0.079	20:43
Air Blank		0.000	20:43
Subject Sample	# 1	0.241	20:44
Breath Volume		1.671	20:44
Air Blank		0.000	20:44
Air Blank		0.000	20:46
Subject Sample	# 2	0.234	20:47
Breath Volume		1.808	20:47
Air Blank		0.000	20:48
Control Test		0.078	20:48
Air Blank		0.000	20:49
Diagnostics Check		OK	20:49

Data Download Date/Time: 07/22/2019 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAROTTA, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: CLAGUE, ERIC
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: MEURER

DATE/TIME OF TEST: 07/30/2019 19:38:09

Diagnostics Check	OK	19:31	
Air Blank	0.000	19:32	
Control Test	0.079	19:32	
Air Blank	0.000	19:32	
Subject Sample # 1	0.301	19:33	
Breath Volume	1.468	19:33	
Air Blank	0.000	19:34	
Air Blank	0.000	19:36	
Subject Sample # 2	0.301	19:36	
Breath Volume	1.507	19:36	
Air Blank	0.000	19:37	
Control Test	0.074	19:37	Control Outside Tolerance
Air Blank	0.000	19:38	

Data Download Date/Time: 09/03/2019 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,OFC. A.,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HAZIME,OFC. A.
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 07/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: WAGENHOFFER

DATE/TIME OF TEST: 07/19/2019 01:55:57

Diagnostics Check		OK	01:49
Air Blank		0.000	01:50
Control Test		0.077	01:50
Air Blank		0.000	01:51
Subject Sample	# 1	0.084	01:51
Breath Volume		2.203	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.094	01:54
Breath Volume		1.769	01:54
Air Blank		0.000	01:55
Control Test		0.078	01:55
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 08/27/2019 22:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,MARC,T
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: TOLDO,MICHAEL
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 07/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:03

SUBJECT'S LAST NAME: DACCHILLE

DATE/TIME OF TEST: 07/19/2019 16:30:00

Diagnostics Check	OK	16:23
Air Blank	0.000	16:24
Control Test	0.078	16:24
Air Blank	0.000	16:25
Subject Sample # 1	0.000	16:25
Breath Volume	2.921	16:25
Air Blank	0.000	16:26
Air Blank	0.000	16:28
Subject Sample # 2	0.000	16:28
Breath Volume	2.839	16:28
Air Blank	0.000	16:29
Control Test	0.080	16:29
Air Blank	0.000	16:29
Diagnostics Check	OK	16:30

Data Download Date/Time: 08/27/2019 22:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGEL,JASON,S
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: SABO,TYLER
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 07/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: EARHART

DATE/TIME OF TEST: 07/19/2019 21:22:58

Diagnostics Check	OK	21:16
Air Blank	0.000	21:16
Control Test	0.079	21:17
Air Blank	0.000	21:17
Subject Sample # 1	0.203	21:18
Breath Volume	1.296	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:20
Subject Sample # 2	0.197	21:21
Breath Volume	1.535	21:21
Air Blank	0.000	21:22
Control Test	0.078	21:22
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 08/27/2019 22:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGEL,JASON,S
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: BERGE,TIMOTHY
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 07/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 07/22/2019 02:58:25

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:55	
Control Test	0.078	02:55	
Air Blank	0.000	02:56	
Subject Sample	# 1	02:56	RFI Detect
Breath Volume	0.351	02:56	
Air Blank	0.000	02:57	
Control Test	0.079	02:57	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

Data Download Date/Time: 08/27/2019 22:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGEL,JASON,S
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: BERGE,TIMOTHY
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 07/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 07/22/2019 03:06:23

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.077	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.165	03:01
Breath Volume	2.074	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:03
Subject Sample # 2	0.159	03:04
Breath Volume	2.789	03:04
Air Blank	0.000	03:05
Control Test	0.076	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 08/27/2019 22:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,OFC. A.,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HAZIME,OFC. A.
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 07/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: BLAKLEY

DATE/TIME OF TEST: 07/22/2019 20:41:14

Diagnostics Check	OK	20:35
Air Blank	0.000	20:35
Control Test	0.075	20:35
Air Blank	0.000	20:36
Subject Sample # 1	0.245	20:36
Breath Volume	1.742	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.237	20:39
Breath Volume	1.347	20:39
Air Blank	0.000	20:40
Control Test	0.077	20:40
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 08/27/2019 22:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,OFC. A.,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HAZIME,OFC. A.
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 07/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 07/23/2019 22:26:20

Diagnostics Check	OK	22:17	
Air Blank	0.000	22:18	
Control Test	0.078	22:18	
Air Blank	0.000	22:19	
Subject Sample	# 1	22:22	No Sample Provided
Breath Volume	0.000	22:22	
Air Blank	0.000	22:22	
Air Blank	0.000	22:24	
Subject Sample	# 2	22:24	Subject Test Refused
Breath Volume	0.000	22:24	
Air Blank	0.000	22:25	
Control Test	0.079	22:25	
Air Blank	0.000	22:26	
Diagnostics Check	OK	22:26	

Data Download Date/Time: 08/27/2019 22:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,OFC. A.,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HAZIME,OFC. A.
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 07/07/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 07/27/2019 03:06:27

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.077	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.168	03:02
Breath Volume	1.574	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.169	03:04
Breath Volume	1.441	03:04
Air Blank	0.000	03:05
Control Test	0.078	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 08/27/2019 22:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,JACOB,R
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: MILLER,JACOB
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: PANTZOS

DATE/TIME OF TEST: 07/02/2019 02:09:50

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.080	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.187	02:05
Breath Volume		2.078	02:05
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.185	02:08
Breath Volume		1.492	02:08
Air Blank		0.000	02:08
Control Test		0.079	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 07/22/2019 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPATZ,JOHN,M
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: WARWICK,ANDREW
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: CROUNSE

DATE/TIME OF TEST: 07/18/2019 23:42:01

Diagnostics Check	OK	23:39	
Air Blank	0.000	23:39	
Control Test	0.079	23:39	
Air Blank	0.000	23:40	
Subject Sample	# 1	23:40	Subject Test Refused
Breath Volume	0.000	23:40	
Air Blank	0.000	23:41	
Control Test	0.080	23:41	
Air Blank	0.000	23:41	
Diagnostics Check	OK	23:42	

Data Download Date/Time: 07/22/2019 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONDOS,ROBERT,E
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: RONDOS,ROBERT
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: CHUDY

DATE/TIME OF TEST: 07/25/2019 21:43:27

Diagnostics Check	OK	21:37
Air Blank	0.000	21:37
Control Test	0.080	21:38
Air Blank	0.000	21:38
Subject Sample # 1	0.000	21:39
Breath Volume	2.480	21:39
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.000	21:42
Breath Volume	2.179	21:42
Air Blank	0.000	21:42
Control Test	0.081	21:42
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 08/13/2019 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: HEGSTROM

DATE/TIME OF TEST: 07/02/2019 05:09:36

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.082	05:03
Air Blank	0.000	05:04
Subject Sample # 1	0.039	05:05
Breath Volume	3.429	05:05
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.041	05:08
Breath Volume	4.187	05:08
Air Blank	0.000	05:08
Control Test	0.082	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 07/10/2019 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BROWN,ERIK
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 07/02/2019 19:20:58

Diagnostics Check	OK	19:14
Air Blank	0.000	19:15
Control Test	0.082	19:15
Air Blank	0.000	19:15
Subject Sample # 1	0.000	19:16
Breath Volume	1.429	19:16
Air Blank	0.000	19:16
Air Blank	0.000	19:18
Subject Sample # 2	0.000	19:19
Breath Volume	1.390	19:19
Air Blank	0.000	19:20
Control Test	0.082	19:20
Air Blank	0.000	19:20
Diagnostics Check	OK	19:20

Data Download Date/Time: 07/10/2019 21:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: MOOD

DATE/TIME OF TEST: 07/03/2019 04:11:51

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.000	04:07
Breath Volume	1.257	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.000	04:10
Breath Volume	1.535	04:10
Air Blank	0.000	04:10
Control Test	0.080	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 07/10/2019 21:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CARLSON,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: SALAME

DATE/TIME OF TEST: 07/04/2019 01:42:22

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:39	
Control Test	0.082	01:39	
Air Blank	0.000	01:40	
Subject Sample	# 1	01:41	Subject Test Refused
Breath Volume	0.000	01:41	
Air Blank	0.000	01:41	
Control Test	0.080	01:41	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 07/10/2019 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREMUR,KIP,E
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: MCDONALD,KIMBERLY
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: TUTTLE

DATE/TIME OF TEST: 07/04/2019 18:41:59

Diagnostics Check	OK	18:35
Air Blank	0.000	18:36
Control Test	0.082	18:36
Air Blank	0.000	18:36
Subject Sample # 1	0.000	18:37
Breath Volume	3.523	18:37
Air Blank	0.000	18:37
Air Blank	0.000	18:39
Subject Sample # 2	0.000	18:40
Breath Volume	2.550	18:40
Air Blank	0.000	18:41
Control Test	0.081	18:41
Air Blank	0.000	18:41
Diagnostics Check	OK	18:41

Data Download Date/Time: 07/10/2019 21:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: VU

DATE/TIME OF TEST: 07/04/2019 23:21:01

Diagnostics Check	OK	23:14
Air Blank	0.000	23:14
Control Test	0.082	23:15
Air Blank	0.000	23:15
Subject Sample # 1	0.177	23:16
Breath Volume	1.367	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:18
Subject Sample # 2	0.193	23:19
Breath Volume	1.656	23:19
Air Blank	0.000	23:20
Control Test	0.081	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:21

Data Download Date/Time: 07/10/2019 21:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TIMMY,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WILLIAMS,TIMMY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: DURAN

DATE/TIME OF TEST: 07/05/2019 00:00:38

Diagnostics Check	OK	23:54
Air Blank	0.000	23:55
Control Test	0.081	23:55
Air Blank	0.000	23:55
Subject Sample # 1	0.175	23:56
Breath Volume	1.265	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.167	23:59
Breath Volume	2.187	23:59
Air Blank	0.000	23:59
Control Test	0.080	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 07/10/2019 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ALVIS,REECE
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: BUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: LOJEWSKI

DATE/TIME OF TEST: 07/05/2019 00:56:47

Diagnostics Check		OK	00:50
Air Blank		0.000	00:51
Control Test		0.082	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.116	00:52
Breath Volume		2.492	00:52
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.119	00:55
Breath Volume		2.617	00:55
Air Blank		0.000	00:55
Control Test		0.081	00:56
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 07/10/2019 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: MISKOVICH

DATE/TIME OF TEST: 07/05/2019 03:20:11

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.082	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.122	03:15
Breath Volume	1.859	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.122	03:18
Breath Volume	1.691	03:18
Air Blank	0.000	03:19
Control Test	0.081	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 07/10/2019 21:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MARTINEAU,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: CROWLEY

DATE/TIME OF TEST: 07/05/2019 23:59:24

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.082	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.000	23:54
Breath Volume	1.816	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.000	23:58
Breath Volume	1.550	23:58
Air Blank	0.000	23:58
Control Test	0.082	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 07/10/2019 21:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: YODER

DATE/TIME OF TEST: 07/06/2019 00:59:34

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.082	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.087	00:54
Breath Volume		2.242	00:54
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.086	00:58
Breath Volume		2.593	00:58
Air Blank		0.000	00:58
Control Test		0.081	00:58
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 07/10/2019 21:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: REIDY,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/06/2019 04:32:24

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.082	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.066	04:28
Breath Volume	1.843	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.065	04:30
Breath Volume	1.753	04:30
Air Blank	0.000	04:31
Control Test	0.081	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 07/10/2019 21:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: GALLEGO

DATE/TIME OF TEST: 07/06/2019 05:22:38

Diagnostics Check	OK	05:11	
Air Blank	0.000	05:11	
Control Test	0.081	05:12	
Air Blank	0.000	05:12	
Subject Sample	# 1	05:15	No Sample Provided
Breath Volume	0.000	05:15	
Air Blank	0.000	05:16	
Air Blank	0.000	05:18	
Subject Sample	# 2	05:21	No Sample Provided
Breath Volume	0.000	05:21	
Air Blank	0.000	05:21	
Control Test	0.081	05:22	
Air Blank	0.000	05:22	
Diagnostics Check	OK	05:22	

Data Download Date/Time: 07/10/2019 21:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: STYLES,L
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 07/06/2019 23:41:48

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.082	23:35
Air Blank	0.000	23:35
Subject Sample # 1	0.317	23:36
Breath Volume	1.386	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.303	23:40
Breath Volume	1.554	23:40
Air Blank	0.000	23:40
Control Test	0.079	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 07/10/2019 21:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FREDA,KYLE
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 14:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:04

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 07/07/2019 16:31:17

Diagnostics Check	OK	16:24
Air Blank	0.000	16:25
Control Test	0.082	16:25
Air Blank	0.000	16:26
Subject Sample # 1	0.146	16:26
Breath Volume	1.570	16:26
Air Blank	0.000	16:27
Air Blank	0.000	16:29
Subject Sample # 2	0.140	16:29
Breath Volume	1.140	16:29
Air Blank	0.000	16:30
Control Test	0.081	16:30
Air Blank	0.000	16:31
Diagnostics Check	OK	16:31

Data Download Date/Time: 07/10/2019 21:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: EASTTY,MARK
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: FALSO

DATE/TIME OF TEST: 07/07/2019 20:46:00

Diagnostics Check	OK	20:38
Air Blank	0.000	20:39
Control Test	0.082	20:39
Air Blank	0.000	20:40
Subject Sample # 1	0.245	20:41
Breath Volume	1.539	20:41
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.246	20:44
Breath Volume	1.414	20:44
Air Blank	0.000	20:45
Control Test	0.081	20:45
Air Blank	0.000	20:45
Diagnostics Check	OK	20:46

Data Download Date/Time: 07/10/2019 21:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 07/07/2019 23:33:50

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.082	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.000	23:29
Breath Volume	1.914	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.000	23:32
Breath Volume	2.621	23:32
Air Blank	0.000	23:32
Control Test	0.082	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 07/10/2019 21:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BYRD,W
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/08/2019 03:56:45

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:53	
Control Test	0.082	03:53	
Air Blank	0.000	03:54	
Subject Sample	# 1	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:55	
Control Test	0.082	03:56	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 07/10/2019 21:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: PEASLEE

DATE/TIME OF TEST: 07/09/2019 00:50:23

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:47	
Control Test	0.083	00:47	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:49	Subject Test Refused
Breath Volume	0.000	00:49	
Air Blank	0.000	00:49	
Control Test	0.082	00:49	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

Data Download Date/Time: 07/10/2019 21:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: AVALOS

DATE/TIME OF TEST: 07/10/2019 03:06:59

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.082	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.128	03:02
Breath Volume	1.367	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.139	03:05
Breath Volume	1.820	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 07/10/2019 21:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JAGODA,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 07/12/2019 02:34:22

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:23	
Control Test	0.083	02:23	
Air Blank	0.000	02:24	
Subject Sample	# 1	02:27	No Sample Provided
Breath Volume	0.000	02:27	
Air Blank	0.000	02:27	
Air Blank	0.000	02:29	
Subject Sample	# 2	02:32	No Sample Provided
Breath Volume	0.000	02:32	
Air Blank	0.000	02:33	
Control Test	0.082	02:33	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 08/03/2019 22:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JAGODA,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 07/12/2019 02:52:30

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.083	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.204	02:48
Breath Volume	2.832	02:48
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.210	02:50
Breath Volume	1.656	02:50
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 08/03/2019 22:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MICHELL,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: LUGO-ESTRELLA

DATE/TIME OF TEST: 07/13/2019 04:25:45

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.082	04:20
Air Blank	0.000	04:20
Subject Sample # 1	0.000	04:21
Breath Volume	3.050	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.000	04:24
Breath Volume	3.148	04:24
Air Blank	0.000	04:24
Control Test	0.081	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 08/03/2019 22:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALERMITI,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: ROQUE RODRIGUEZ

DATE/TIME OF TEST: 07/13/2019 04:50:55

Diagnostics Check	OK	04:44
Air Blank	0.000	04:44
Control Test	0.082	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.032	04:46
Breath Volume	1.160	04:46
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.030	04:49
Breath Volume	2.179	04:49
Air Blank	0.000	04:50
Control Test	0.081	04:50
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 08/03/2019 22:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTHINGTON,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WORTHINGTON,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: TOMLINSON

DATE/TIME OF TEST: 07/14/2019 00:10:02

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:07	
Control Test	0.081	00:07	
Air Blank	0.000	00:08	
Subject Sample	# 1	00:08	RFI Detect
Breath Volume	0.000	00:08	
Air Blank	0.000	00:09	
Control Test	0.081	00:09	
Air Blank	0.000	00:09	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 08/03/2019 22:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTHINGTON,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WORTHINGTON,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: TOMLINSON

DATE/TIME OF TEST: 07/14/2019 00:17:00

Diagnostics Check	OK	00:14	
Air Blank	0.000	00:14	
Control Test	0.082	00:14	
Air Blank	0.000	00:15	
Subject Sample	# 1	00:15	Subject Test Refused
Breath Volume	0.000	00:15	
Air Blank	0.000	00:16	
Control Test	0.081	00:16	
Air Blank	0.000	00:16	
Diagnostics Check	OK	00:17	

Data Download Date/Time: 08/03/2019 22:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 07/14/2019 04:36:58

Diagnostics Check	OK	04:34	
Air Blank	0.000	04:34	
Control Test	0.083	04:34	
Air Blank	0.000	04:35	
Subject Sample	# 1	04:35	Subject Test Refused
Breath Volume	0.000	04:35	
Air Blank	0.000	04:36	
Control Test	0.081	04:36	
Air Blank	0.000	04:36	
Diagnostics Check	OK	04:36	

Data Download Date/Time: 08/03/2019 22:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVIS
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PUPKE,LUKE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:03

SUBJECT'S LAST NAME: BLOUIN

DATE/TIME OF TEST: 07/14/2019 06:34:44

Diagnostics Check	OK	06:28
Air Blank	0.000	06:28
Control Test	0.082	06:28
Air Blank	0.000	06:29
Subject Sample # 1	0.036	06:30
Breath Volume	1.281	06:30
Air Blank	0.000	06:30
Air Blank	0.000	06:32
Subject Sample # 2	0.035	06:33
Breath Volume	1.355	06:33
Air Blank	0.000	06:33
Control Test	0.082	06:34
Air Blank	0.000	06:34
Diagnostics Check	OK	06:34

Data Download Date/Time: 08/03/2019 22:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:17

SUBJECT'S LAST NAME: REPPERT

DATE/TIME OF TEST: 07/14/2019 07:44:30

Diagnostics Check	OK	07:38
Air Blank	0.000	07:38
Control Test	0.082	07:39
Air Blank	0.000	07:39
Subject Sample # 1	0.000	07:40
Breath Volume	1.988	07:40
Air Blank	0.000	07:40
Air Blank	0.000	07:42
Subject Sample # 2	0.000	07:43
Breath Volume	1.750	07:43
Air Blank	0.000	07:43
Control Test	0.082	07:43
Air Blank	0.000	07:44
Diagnostics Check	OK	07:44

Data Download Date/Time: 08/03/2019 22:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS,M,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COOK,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CRUZ-PENA

DATE/TIME OF TEST: 07/15/2019 03:53:05

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.082	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.158	03:48
Breath Volume		2.515	03:48
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.141	03:51
Breath Volume		1.109	03:51
Air Blank		0.000	03:52
Control Test		0.081	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 08/03/2019 22:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MARTINI,I
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: YINGLING

DATE/TIME OF TEST: 07/16/2019 03:24:56

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.082	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.053	03:20
Breath Volume	1.976	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.054	03:23
Breath Volume	1.824	03:23
Air Blank	0.000	03:24
Control Test	0.081	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 08/03/2019 22:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIOTT,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MERI

DATE/TIME OF TEST: 07/18/2019 02:49:03

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:46	
Control Test	0.083	02:46	
Air Blank	0.000	02:47	
Subject Sample	# 1	02:47	Subject Test Refused
Breath Volume	0.000	02:47	
Air Blank	0.000	02:48	
Control Test	0.082	02:48	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:49	

Data Download Date/Time: 08/03/2019 22:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CHMELIK

DATE/TIME OF TEST: 07/18/2019 03:03:23

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:01	
Control Test	0.081	03:01	
Air Blank	0.000	03:01	
Subject Sample	# 1	03:02	Subject Test Refused
Breath Volume	0.000	03:02	
Air Blank	0.000	03:02	
Control Test	0.081	03:02	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CRANE,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ENGH

DATE/TIME OF TEST: 07/19/2019 01:09:12

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.082	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.146	01:04
Breath Volume	1.593	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:06
Subject Sample # 2	0.142	01:07
Breath Volume	1.539	01:07
Air Blank	0.000	01:08
Control Test	0.081	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/19/2019 05:52:21

Diagnostics Check	OK	05:45
Air Blank	0.000	05:46
Control Test	0.082	05:46
Air Blank	0.000	05:46
Subject Sample # 1	0.232	05:47
Breath Volume	2.195	05:47
Air Blank	0.000	05:48
Air Blank	0.000	05:50
Subject Sample # 2	0.235	05:50
Breath Volume	2.359	05:50
Air Blank	0.000	05:51
Control Test	0.081	05:51
Air Blank	0.000	05:52
Diagnostics Check	OK	05:52

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANGANARO,JESSICA,H
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MANGANARO,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:44

SUBJECT'S LAST NAME: LAURO

DATE/TIME OF TEST: 07/19/2019 15:11:21

Diagnostics Check	OK	15:05
Air Blank	0.000	15:05
Control Test	0.082	15:05
Air Blank	0.000	15:06
Subject Sample # 1	0.000	15:07
Breath Volume	1.710	15:07
Air Blank	0.000	15:07
Air Blank	0.000	15:09
Subject Sample # 2	0.000	15:09
Breath Volume	2.667	15:09
Air Blank	0.000	15:10
Control Test	0.081	15:10
Air Blank	0.000	15:11
Diagnostics Check	OK	15:11

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: CLEMONS

DATE/TIME OF TEST: 07/20/2019 03:07:29

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.082	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.000	03:03
Breath Volume	1.210	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.000	03:06
Breath Volume	2.171	03:06
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/21/2019 01:21:08

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:18	
Control Test	0.082	01:18	
Air Blank	0.000	01:18	
Subject Sample	# 1	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:20	
Control Test	0.081	01:20	
Air Blank	0.000	01:21	
Diagnostics Check	OK	01:21	

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI, JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: LEBLANC JR

DATE/TIME OF TEST: 07/21/2019 02:01:31

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.081	01:54
Air Blank	0.000	01:54
Subject Sample # 1	0.310	01:55
Breath Volume	1.746	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.299	01:59
Breath Volume	2.085	01:59
Air Blank	0.000	02:00
Control Test	0.081	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: BARNEY

DATE/TIME OF TEST: 07/21/2019 21:03:24

Diagnostics Check	OK	21:00	
Air Blank	0.000	21:01	
Control Test	0.082	21:01	
Air Blank	0.000	21:01	
Subject Sample	# 1	21:02	Subject Test Refused
Breath Volume	0.000	21:02	
Air Blank	0.000	21:02	
Control Test	0.082	21:02	
Air Blank	0.000	21:03	
Diagnostics Check	OK	21:03	

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS, TOMESHIA, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ARBANAS, SHANE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/24/2019 22:23:23

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.081	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.075	22:18
Breath Volume	1.453	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.066	22:21
Breath Volume	1.664	22:21
Air Blank	0.000	22:22
Control Test	0.080	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: CREAMER

DATE/TIME OF TEST: 07/25/2019 04:39:54

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.083	04:34
Air Blank	0.000	04:34
Subject Sample # 1	0.196	04:35
Breath Volume	1.671	04:35
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.187	04:38
Breath Volume	1.582	04:38
Air Blank	0.000	04:38
Control Test	0.082	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: MCGOVERN

DATE/TIME OF TEST: 07/26/2019 01:51:30

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:49	
Control Test	0.082	01:49	
Air Blank	0.000	01:49	
Subject Sample	# 1	01:50	Subject Test Refused
Breath Volume	0.000	01:50	
Air Blank	0.000	01:50	
Control Test	0.082	01:50	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FOSTER,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: DEDA

DATE/TIME OF TEST: 07/26/2019 02:22:15

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.083	02:15
Air Blank		0.000	02:16
Subject Sample	# 1	0.061	02:17
Breath Volume		1.710	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.061	02:20
Breath Volume		1.199	02:20
Air Blank		0.000	02:21
Control Test		0.082	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MARTINI,I
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: PROCTOR

DATE/TIME OF TEST: 07/26/2019 02:55:02

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.082	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.131	02:50
Breath Volume	1.496	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.127	02:53
Breath Volume	1.679	02:53
Air Blank	0.000	02:54
Control Test	0.082	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:55

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALERMITI,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/26/2019 04:29:45

Diagnostics Check	OK	04:25	
Air Blank	0.000	04:25	
Control Test	0.082	04:25	
Air Blank	0.000	04:26	
Subject Sample	# 1	04:28	Subject Test Refused
Breath Volume	0.000	04:28	
Air Blank	0.000	04:28	
Control Test	0.082	04:29	
Air Blank	0.000	04:29	
Diagnostics Check	OK	04:29	

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOD,COREY,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VAN WORP,JUDSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:41

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/26/2019 08:03:59

Diagnostics Check	OK	08:01	
Air Blank	0.000	08:01	
Control Test	0.083	08:01	
Air Blank	0.000	08:02	
Subject Sample	# 1	08:02	Subject Test Refused
Breath Volume	0.000	08:02	
Air Blank	0.000	08:03	
Control Test	0.082	08:03	
Air Blank	0.000	08:03	
Diagnostics Check	OK	08:03	

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTHINGTON,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WORTHINGTON,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 07/27/2019 03:30:43

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.083	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.307	03:26
Breath Volume	1.796	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:28
Subject Sample # 2	0.298	03:29
Breath Volume	3.750	03:29
Air Blank	0.000	03:29
Control Test	0.081	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTHINGTON,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WORTHINGTON,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: ACHU

DATE/TIME OF TEST: 07/27/2019 23:25:16

Diagnostics Check	OK	23:18
Air Blank	0.000	23:18
Control Test	0.082	23:18
Air Blank	0.000	23:19
Subject Sample # 1	0.116	23:20
Breath Volume	2.378	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.110	23:23
Breath Volume	1.574	23:23
Air Blank	0.000	23:24
Control Test	0.081	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,JOHN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: PAVLAT-MORRIS

DATE/TIME OF TEST: 07/28/2019 23:05:03

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.083	22:58
Air Blank	0.000	22:59
Subject Sample # 1	0.174	23:00
Breath Volume	2.812	23:00
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.175	23:03
Breath Volume	2.589	23:03
Air Blank	0.000	23:04
Control Test	0.082	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:05

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTWICK,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:29

SUBJECT'S LAST NAME: MAZZA

DATE/TIME OF TEST: 07/30/2019 06:53:21

Diagnostics Check	OK	06:49	
Air Blank	0.000	06:50	
Control Test	0.083	06:50	
Air Blank	0.000	06:51	
Subject Sample	# 1	06:51	Subject Test Refused
Breath Volume	0.000	06:51	
Air Blank	0.000	06:52	
Control Test	0.082	06:52	
Air Blank	0.000	06:53	
Diagnostics Check	OK	06:53	

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VAN WORP,JUDSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: CLEARY

DATE/TIME OF TEST: 07/31/2019 02:56:08

Diagnostics Check	OK	02:55	
Air Blank		02:55	RFI Detect
Air Blank		02:56	RFI Detect

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VAN WORP,JUDSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: CLEARY

DATE/TIME OF TEST: 07/31/2019 03:11:19

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:58	
Control Test	0.083	02:59	
Air Blank	0.000	02:59	
Subject Sample	# 1 0.205	03:01	
Breath Volume	1.843	03:01	
Air Blank	0.000	03:01	
Air Blank	0.000	03:03	
Subject Sample	# 2	03:06	No Sample Provided
Breath Volume	0.000	03:06	
Air Blank	0.000	03:07	
Air Blank	0.000	03:09	
Subject Sample	# 3 0.201	03:09	
Breath Volume	2.042	03:09	
Air Blank	0.000	03:10	
Control Test	0.082	03:10	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: HLAVA

DATE/TIME OF TEST: 07/31/2019 04:31:29

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.082	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.139	04:27
Breath Volume	1.937	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.134	04:30
Breath Volume	1.636	04:30
Air Blank	0.000	04:30
Control Test	0.082	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWETNICH,JARED,A
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: SWETNICH,JARED
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 07/10/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 07/31/2019 23:39:08

Diagnostics Check	OK	23:36	
Air Blank	0.000	23:36	
Control Test	0.081	23:37	
Air Blank	0.000	23:37	
Subject Sample	# 1	23:37	Subject Test Refused
Breath Volume	0.000	23:37	
Air Blank	0.000	23:38	
Control Test	0.081	23:38	
Air Blank	0.000	23:39	
Diagnostics Check	OK	23:39	

Data Download Date/Time: 08/03/2019 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: HERNANDEZ MORALES

DATE/TIME OF TEST: 07/01/2019 01:24:40

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.078	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.092	01:20
Breath Volume	2.664	01:20
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.091	01:23
Breath Volume	2.332	01:23
Air Blank	0.000	01:23
Control Test	0.079	01:24
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 07/03/2019 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: CANAY-LAYNEZ

DATE/TIME OF TEST: 07/02/2019 23:42:23

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.078	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.151	23:37
Breath Volume	1.972	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.132	23:40
Breath Volume	2.312	23:40
Air Blank	0.000	23:41
Control Test	0.078	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 07/03/2019 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/13/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 07/03/2019 03:19:14

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.199	03:14
Breath Volume	1.781	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.190	03:17
Breath Volume	1.312	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 07/03/2019 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: ECHEVERRI SOSSA

DATE/TIME OF TEST: 07/04/2019 03:48:52

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.078	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.059	03:44
Breath Volume	1.949	03:44
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.057	03:47
Breath Volume	1.574	03:47
Air Blank	0.000	03:47
Control Test	0.078	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUC,NORBEN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: ROZENBAUM

DATE/TIME OF TEST: 07/04/2019 04:55:18

Diagnostics Check	OK	04:48
Air Blank	0.000	04:48
Control Test	0.078	04:48
Air Blank	0.000	04:49
Subject Sample # 1	0.238	04:50
Breath Volume	1.675	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:52
Subject Sample # 2	0.224	04:53
Breath Volume	1.136	04:53
Air Blank	0.000	04:54
Control Test	0.077	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: HAMZE

DATE/TIME OF TEST: 07/04/2019 22:01:34

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.078	21:55
Air Blank	0.000	21:56
Subject Sample # 1	0.165	21:57
Breath Volume	2.843	21:57
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.166	22:00
Breath Volume	3.429	22:00
Air Blank	0.000	22:00
Control Test	0.079	22:00
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TRABUE,THOMAS
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 07/05/2019 02:11:20

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.078	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.139	02:06
Breath Volume	1.226	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.128	02:09
Breath Volume	1.257	02:09
Air Blank	0.000	02:10
Control Test	0.079	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ORTUNO,CARLOS
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: RAMIREZ FLORES

DATE/TIME OF TEST: 07/05/2019 21:10:05

Diagnostics Check	OK	21:02
Air Blank	0.000	21:03
Control Test	0.080	21:03
Air Blank	0.000	21:04
Subject Sample # 1	0.192	21:05
Breath Volume	2.382	21:05
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.192	21:08
Breath Volume	1.695	21:08
Air Blank	0.000	21:09
Control Test	0.078	21:09
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLIHER,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 07/08/2019 01:14:54

Diagnostics Check		OK	01:06
Air Blank		0.000	01:07
Control Test		0.079	01:07
Air Blank		0.000	01:08
Subject Sample	# 1	0.195	01:09
Breath Volume		2.203	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.187	01:13
Breath Volume		1.945	01:13
Air Blank		0.000	01:14
Control Test		0.078	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WALLACE,JONATHAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: JEUDI

DATE/TIME OF TEST: 07/08/2019 02:55:14

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.078	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.051	02:50
Breath Volume	2.691	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.051	02:53
Breath Volume	2.621	02:53
Air Blank	0.000	02:54
Control Test	0.078	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BERMAN,MARC
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: AMARSINGH

DATE/TIME OF TEST: 07/09/2019 03:04:39

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.078	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.243	03:00
Breath Volume	1.257	03:00
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.237	03:03
Breath Volume	1.851	03:03
Air Blank	0.000	03:03
Control Test	0.078	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: STEARNS

DATE/TIME OF TEST: 07/10/2019 00:56:55

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.078	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.309	00:52
Breath Volume	2.843	00:52
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.315	00:55
Breath Volume	2.679	00:55
Air Blank	0.000	00:56
Control Test	0.078	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 07/10/2019 04:41:18

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.078	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.119	04:36
Breath Volume	1.527	04:36
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.105	04:39
Breath Volume	1.277	04:39
Air Blank	0.000	04:40
Control Test	0.079	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: HANIFF

DATE/TIME OF TEST: 07/12/2019 05:35:56

Diagnostics Check	OK	05:29
Air Blank	0.000	05:29
Control Test	0.078	05:30
Air Blank	0.000	05:30
Subject Sample # 1	0.178	05:31
Breath Volume	1.750	05:31
Air Blank	0.000	05:32
Air Blank	0.000	05:33
Subject Sample # 2	0.180	05:34
Breath Volume	1.578	05:34
Air Blank	0.000	05:35
Control Test	0.078	05:35
Air Blank	0.000	05:35
Diagnostics Check	OK	05:35

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SGT. BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 07/12/2019 23:35:10

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.078	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.193	23:30
Breath Volume	2.621	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:32
Subject Sample # 2	0.196	23:33
Breath Volume	2.507	23:33
Air Blank	0.000	23:34
Control Test	0.076	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: DESMARAIS

DATE/TIME OF TEST: 07/15/2019 01:00:31

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.078	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.195	00:56
Breath Volume	1.324	00:56
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.199	00:59
Breath Volume	1.433	00:59
Air Blank	0.000	00:59
Control Test	0.078	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,JASON
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 07/15/2019 03:42:33

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.078	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.172	03:38
Breath Volume	2.242	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.158	03:41
Breath Volume	1.890	03:41
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BERMAN,MARC
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: GERECs

DATE/TIME OF TEST: 07/15/2019 23:11:33

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.078	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.151	23:06
Breath Volume	1.656	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.146	23:10
Breath Volume	1.511	23:10
Air Blank	0.000	23:10
Control Test	0.079	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/16/2019 01:27:03

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.078	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.202	01:22
Breath Volume	2.335	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.202	01:25
Breath Volume	2.472	01:25
Air Blank	0.000	01:26
Control Test	0.078	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:27

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/16/2019 05:11:44

Diagnostics Check	OK	05:05
Air Blank	0.000	05:05
Control Test	0.079	05:05
Air Blank	0.000	05:06
Subject Sample # 1	0.079	05:07
Breath Volume	2.328	05:07
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.092	05:10
Breath Volume	3.187	05:10
Air Blank	0.000	05:10
Control Test	0.079	05:11
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: RAMRATTAN

DATE/TIME OF TEST: 07/18/2019 05:49:32

Diagnostics Check	OK	05:42
Air Blank	0.000	05:42
Control Test	0.079	05:43
Air Blank	0.000	05:43
Subject Sample # 1	0.168	05:44
Breath Volume	1.925	05:44
Air Blank	0.000	05:44
Air Blank	0.000	05:46
Subject Sample # 2	0.170	05:48
Breath Volume	2.660	05:48
Air Blank	0.000	05:48
Control Test	0.079	05:48
Air Blank	0.000	05:49
Diagnostics Check	OK	05:49

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: OLEA,ELIEZER
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: SAINT-VICTOR

DATE/TIME OF TEST: 07/19/2019 09:36:17

Diagnostics Check	OK	09:28
Air Blank	0.000	09:28
Control Test	0.079	09:29
Air Blank	0.000	09:29
Subject Sample # 1	0.073	09:31
Breath Volume	1.734	09:31
Air Blank	0.000	09:32
Air Blank	0.000	09:34
Subject Sample # 2	0.068	09:34
Breath Volume	1.281	09:34
Air Blank	0.000	09:35
Control Test	0.079	09:35
Air Blank	0.000	09:36
Diagnostics Check	OK	09:36

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MCCLELLAND,ANDREW
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: SZEILER

DATE/TIME OF TEST: 07/20/2019 23:42:39

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.204	23:35
Breath Volume	1.390	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:37
Subject Sample # 2	0.204	23:41
Breath Volume	1.968	23:41
Air Blank	0.000	23:41
Control Test	0.077	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: GODSEY

DATE/TIME OF TEST: 07/21/2019 23:32:02

Diagnostics Check		OK	23:19	
Air Blank		0.000	23:20	
Control Test		0.079	23:20	
Air Blank		0.000	23:20	
Subject Sample	# 1	0.224	23:21	No .020 agreement
Breath Volume		1.343	23:21	
Air Blank		0.000	23:22	
Air Blank		0.000	23:24	
Subject Sample	# 2	0.185	23:27	Volume Not Met
Breath Volume		0.890	23:27	
Air Blank		0.000	23:28	
Air Blank		0.000	23:30	
Subject Sample	# 3	0.191	23:30	No .020 agreement
Breath Volume		1.812	23:30	
Air Blank		0.000	23:31	
Control Test		0.077	23:31	
Air Blank		0.000	23:31	
Diagnostics Check		OK	23:32	

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GODSEY

DATE/TIME OF TEST: 07/22/2019 01:28:00

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.077	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.187	01:23
Breath Volume		2.062	01:23
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.178	01:26
Breath Volume		1.859	01:26
Air Blank		0.000	01:27
Control Test		0.078	01:27
Air Blank		0.000	01:27
Diagnostics Check		OK	01:28

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MCCLELAND,ANDREW
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: BASTIEN

DATE/TIME OF TEST: 07/23/2019 07:58:49

Diagnostics Check	OK	07:50
Air Blank	0.000	07:50
Control Test	0.079	07:50
Air Blank	0.000	07:51
Subject Sample # 1	0.159	07:52
Breath Volume	1.351	07:52
Air Blank	0.000	07:53
Air Blank	0.000	07:54
Subject Sample # 2	0.162	07:57
Breath Volume	2.261	07:57
Air Blank	0.000	07:57
Control Test	0.079	07:58
Air Blank	0.000	07:58
Diagnostics Check	OK	07:58

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BURNETTE,JAVON
ARRESTING OFFICER AGENCY: Lauderhill PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 07/24/2019 04:23:30

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.078	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.146	04:18
Breath Volume	3.019	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.147	04:22
Breath Volume	3.261	04:22
Air Blank	0.000	04:22
Control Test	0.078	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BERMAN,MARC
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 07/24/2019 23:13:55

Diagnostics Check	OK	23:07
Air Blank	0.000	23:08
Control Test	0.078	23:08
Air Blank	0.000	23:08
Subject Sample # 1	0.161	23:09
Breath Volume	2.125	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:11
Subject Sample # 2	0.159	23:12
Breath Volume	2.164	23:12
Air Blank	0.000	23:13
Control Test	0.078	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MAS,MILES
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: PINTO PAZ

DATE/TIME OF TEST: 07/26/2019 09:55:49

Diagnostics Check		OK	09:48
Air Blank		0.000	09:49
Control Test		0.079	09:49
Air Blank		0.000	09:50
Subject Sample	# 1	0.165	09:51
Breath Volume		1.996	09:51
Air Blank		0.000	09:51
Air Blank		0.000	09:53
Subject Sample	# 2	0.157	09:54
Breath Volume		1.546	09:54
Air Blank		0.000	09:54
Control Test		0.078	09:55
Air Blank		0.000	09:55
Diagnostics Check		OK	09:55

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,DANNY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: FATTER

DATE/TIME OF TEST: 07/26/2019 11:21:50

Diagnostics Check	OK	11:14
Air Blank	0.000	11:15
Control Test	0.079	11:15
Air Blank	0.000	11:15
Subject Sample # 1	0.000	11:16
Breath Volume	1.796	11:16
Air Blank	0.000	11:17
Air Blank	0.000	11:19
Subject Sample # 2	0.000	11:20
Breath Volume	1.554	11:20
Air Blank	0.000	11:20
Control Test	0.079	11:21
Air Blank	0.000	11:21
Diagnostics Check	OK	11:21

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: IRELAN,ANDREW
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: CHEVERE

DATE/TIME OF TEST: 07/27/2019 04:10:26

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.079	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.191	04:05
Breath Volume	3.570	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.192	04:08
Breath Volume	3.476	04:08
Air Blank	0.000	04:09
Control Test	0.078	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: INGLETON

DATE/TIME OF TEST: 07/27/2019 05:04:51

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.078	04:59
Air Blank	0.000	04:59
Subject Sample # 1	0.167	05:00
Breath Volume	1.410	05:00
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.175	05:03
Breath Volume	2.582	05:03
Air Blank	0.000	05:03
Control Test	0.078	05:04
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/03/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/30/2019 20:19:26

Diagnostics Check	OK	20:13
Air Blank	0.000	20:13
Control Test	0.078	20:14
Air Blank	0.000	20:14
Subject Sample # 1	0.152	20:15
Breath Volume	3.406	20:15
Air Blank	0.000	20:15
Air Blank	0.000	20:17
Subject Sample # 2	0.153	20:17
Breath Volume	3.273	20:17
Air Blank	0.000	20:18
Control Test	0.077	20:18
Air Blank	0.000	20:19
Diagnostics Check	OK	20:19

Data Download Date/Time: 07/31/2019 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/16/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: AMLANI

DATE/TIME OF TEST: 07/21/2019 02:24:05

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.079	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.138	02:19
Breath Volume		2.261	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:21
Subject Sample	# 2	0.143	02:22
Breath Volume		2.046	02:22
Air Blank		0.000	02:23
Control Test		0.079	02:23
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 08/23/2019 20:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/06/2019 03:36:58

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.079	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.138	03:32
Breath Volume	1.531	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.147	03:35
Breath Volume	1.683	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 07/25/2019 11:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,CAMERON,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SENTERS,B
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 07/21/2019 02:35:13

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.174	02:30
Breath Volume	2.285	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.176	02:33
Breath Volume	1.785	02:33
Air Blank	0.000	02:34
Control Test	0.079	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 07/25/2019 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,C,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: FOREMAN

DATE/TIME OF TEST: 07/21/2019 17:03:15

Diagnostics Check	OK	16:54
Air Blank	0.000	16:55
Control Test	0.080	16:55
Air Blank	0.000	16:56
Subject Sample # 1	0.141	16:58
Breath Volume	1.656	16:58
Air Blank	0.000	16:58
Air Blank	0.000	17:00
Subject Sample # 2	0.142	17:01
Breath Volume	1.363	17:01
Air Blank	0.000	17:02
Control Test	0.079	17:02
Air Blank	0.000	17:03
Diagnostics Check	OK	17:03

Data Download Date/Time: 07/25/2019 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASLINE, T'NEISHA, S
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HART, B
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: GALLION

DATE/TIME OF TEST: 07/25/2019 22:38:28

Diagnostics Check		OK	22:32
Air Blank		0.000	22:32
Control Test		0.079	22:32
Air Blank		0.000	22:33
Subject Sample	# 1	0.262	22:34
Breath Volume		1.179	22:34
Air Blank		0.000	22:34
Air Blank		0.000	22:36
Subject Sample	# 2	0.257	22:37
Breath Volume		1.179	22:37
Air Blank		0.000	22:37
Control Test		0.079	22:37
Air Blank		0.000	22:38
Diagnostics Check		OK	22:38

Data Download Date/Time: 08/26/2019 01:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARCHER,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: VALDES,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: SHOUP

DATE/TIME OF TEST: 07/26/2019 04:02:40

Diagnostics Check		OK	03:55
Air Blank		0.000	03:56
Control Test		0.079	03:56
Air Blank		0.000	03:56
Subject Sample	# 1	0.188	03:57
Breath Volume		2.250	03:57
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.184	04:01
Breath Volume		2.113	04:01
Air Blank		0.000	04:01
Control Test		0.080	04:02
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 08/26/2019 01:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JENNINGS,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: SABOURIN

DATE/TIME OF TEST: 07/31/2019 10:01:04

Diagnostics Check	OK	09:52	
Air Blank	0.000	09:52	
Control Test	0.079	09:53	
Air Blank	0.000	09:53	
Subject Sample	# 1 0.000	09:56	Volume Not Met
Breath Volume	0.496	09:56	
Air Blank	0.000	09:57	
Air Blank	0.000	09:59	
Subject Sample	# 2	09:59	RFI Detect
Breath Volume	0.000	09:59	
Air Blank	0.000	10:00	
Control Test	0.079	10:00	
Air Blank	0.000	10:00	
Diagnostics Check	OK	10:01	

Data Download Date/Time: 08/26/2019 01:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JENNINGS,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: SABOURIN

DATE/TIME OF TEST: 07/31/2019 10:11:47

Diagnostics Check	OK	10:04	
Air Blank	0.000	10:04	
Control Test	0.080	10:04	
Air Blank	0.000	10:05	
Subject Sample # 1	0.000	10:08	
Breath Volume	1.132	10:08	
Air Blank	0.000	10:08	
Air Blank	0.000	10:10	
Subject Sample # 2		10:10	RFI Detect
Breath Volume	0.000	10:10	
Air Blank	0.000	10:11	
Control Test	0.078	10:11	
Air Blank		10:11	RFI Detect
Air Blank		10:11	RFI Detect

Data Download Date/Time: 08/26/2019 01:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JENNINGS,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/25/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: SABOURIN

DATE/TIME OF TEST: 07/31/2019 10:41:07

Diagnostics Check	OK	10:34
Air Blank	0.000	10:34
Control Test	0.079	10:35
Air Blank	0.000	10:35
Subject Sample # 1	0.000	10:36
Breath Volume	1.179	10:36
Air Blank	0.000	10:37
Air Blank	0.000	10:39
Subject Sample # 2	0.000	10:39
Breath Volume	1.308	10:39
Air Blank	0.000	10:40
Control Test	0.080	10:40
Air Blank	0.000	10:41
Diagnostics Check	OK	10:41

Data Download Date/Time: 08/26/2019 01:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REAY,ALEX,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REAY,ALEX
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/18/2019

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: TROXELL

DATE/TIME OF TEST: 07/07/2019 03:01:46

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:50	
Control Test		0.079	02:51	
Air Blank		0.000	02:51	
Subject Sample	# 1	0.130	02:54	Volume Not Met
Breath Volume		0.171	02:54	
Air Blank		0.000	02:55	
Air Blank		0.000	02:57	
Subject Sample	# 2		03:00	No Sample Provided
Breath Volume		0.000	03:00	
Air Blank		0.000	03:00	
Control Test		0.077	03:01	
Air Blank		0.000	03:01	
Diagnostics Check		OK	03:01	

Data Download Date/Time: 07/15/2019 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: FRANSCA

DATE/TIME OF TEST: 07/03/2019 23:40:22

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.081	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.148	23:35
Breath Volume	1.261	23:35
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.155	23:38
Breath Volume	1.164	23:38
Air Blank	0.000	23:39
Control Test	0.081	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 08/19/2019 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: MENENDEZ

DATE/TIME OF TEST: 07/04/2019 02:00:12

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.081	01:55
Breath Volume	2.687	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.084	01:58
Breath Volume	2.863	01:58
Air Blank	0.000	01:59
Control Test	0.080	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 08/19/2019 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: DALMAU

DATE/TIME OF TEST: 07/04/2019 04:49:20

Diagnostics Check	OK	04:41
Air Blank	0.000	04:42
Control Test	0.081	04:42
Air Blank	0.000	04:43
Subject Sample # 1	0.113	04:44
Breath Volume	3.964	04:44
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.113	04:47
Breath Volume	2.683	04:47
Air Blank	0.000	04:48
Control Test	0.080	04:48
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 08/19/2019 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATZ,JUSTIN,T
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: KATZ,JUSTIN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: TELLECHEA

DATE/TIME OF TEST: 07/04/2019 17:24:45

Diagnostics Check	OK	17:21	
Air Blank	0.000	17:21	
Control Test	0.080	17:21	
Air Blank	0.000	17:22	
Subject Sample	# 1	17:23	Subject Test Refused
Breath Volume	0.000	17:23	
Air Blank	0.000	17:23	
Control Test	0.080	17:24	
Air Blank	0.000	17:24	
Diagnostics Check	OK	17:24	

Data Download Date/Time: 08/19/2019 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ZUMALE-VASQUEZ

DATE/TIME OF TEST: 07/05/2019 03:06:38

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.081	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.121	03:02
Breath Volume	3.007	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.119	03:05
Breath Volume	2.253	03:05
Air Blank	0.000	03:05
Control Test	0.080	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 08/19/2019 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: SCHEDLER

DATE/TIME OF TEST: 07/06/2019 22:06:03

Diagnostics Check	OK	21:59
Air Blank	0.000	21:59
Control Test	0.080	21:59
Air Blank	0.000	22:00
Subject Sample # 1	0.161	22:01
Breath Volume	2.519	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.166	22:04
Breath Volume	2.082	22:04
Air Blank	0.000	22:05
Control Test	0.080	22:05
Air Blank	0.000	22:05
Diagnostics Check	OK	22:06

Data Download Date/Time: 08/19/2019 16:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17618080A3

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 00:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: MADDEN

DATE/TIME OF TEST: 07/07/2019 00:53:24

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:50	
Control Test	0.080	00:51	
Air Blank	0.000	00:51	
Subject Sample	# 1	00:52	Subject Test Refused
Breath Volume	0.000	00:52	
Air Blank	0.000	00:52	
Control Test	0.081	00:52	
Air Blank	0.000	00:53	
Diagnostics Check	OK	00:53	

Data Download Date/Time: 08/19/2019 16:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17618080A3

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 00:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: KLEIN

DATE/TIME OF TEST: 07/07/2019 01:39:11

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.080	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.116	01:34
Breath Volume	2.929	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:36
Subject Sample # 2	0.105	01:37
Breath Volume	1.890	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 08/19/2019 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,ROBERT,C
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: COOK,ROBERT
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:01

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/10/2019 10:29:19

Diagnostics Check	OK	10:23
Air Blank	0.000	10:23
Control Test	0.080	10:24
Air Blank	0.000	10:24
Subject Sample # 1	0.185	10:24
Breath Volume	2.054	10:24
Air Blank	0.000	10:25
Air Blank	0.000	10:27
Subject Sample # 2	0.184	10:27
Breath Volume	1.765	10:27
Air Blank	0.000	10:28
Control Test	0.080	10:28
Air Blank	0.000	10:29
Diagnostics Check	OK	10:29

Data Download Date/Time: 08/19/2019 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:07

SUBJECT'S LAST NAME: YUNGERT

DATE/TIME OF TEST: 07/10/2019 18:35:25

Diagnostics Check	OK	18:32	
Air Blank	0.000	18:32	
Control Test	0.080	18:33	
Air Blank	0.000	18:33	
Subject Sample	# 1	18:34	Subject Test Refused
Breath Volume	0.000	18:34	
Air Blank	0.000	18:34	
Control Test	0.080	18:34	
Air Blank	0.000	18:35	
Diagnostics Check	OK	18:35	

Data Download Date/Time: 08/19/2019 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: SADLER

DATE/TIME OF TEST: 07/11/2019 04:43:18

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:36
Subject Sample # 1	0.212	04:38
Breath Volume	1.394	04:38
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.219	04:41
Breath Volume	1.902	04:41
Air Blank	0.000	04:42
Control Test	0.080	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 08/19/2019 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: MAYNARD

DATE/TIME OF TEST: 07/12/2019 02:17:38

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.080	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.129	02:12
Breath Volume	2.402	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.129	02:16
Breath Volume	2.085	02:16
Air Blank	0.000	02:16
Control Test	0.080	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 08/19/2019 16:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17618080A3

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 23:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 07/12/2019 23:53:10

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:50	
Control Test	0.080	23:50	
Air Blank	0.000	23:51	
Subject Sample	# 1	23:51	Subject Test Refused
Breath Volume	0.000	23:51	
Air Blank	0.000	23:52	
Control Test	0.080	23:52	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:53	

Data Download Date/Time: 08/19/2019 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER, MATTHEW, D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: KYNION, JASMINE
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 07/13/2019 01:42:07

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.268	01:37
Breath Volume	1.863	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.280	01:40
Breath Volume	2.222	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 08/19/2019 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTIEN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 07/14/2019 00:10:46

Diagnostics Check	OK	00:03
Air Blank	0.000	00:04
Control Test	0.080	00:04
Air Blank	0.000	00:05
Subject Sample # 1	0.169	00:06
Breath Volume	1.523	00:06
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.152	00:09
Breath Volume	2.160	00:09
Air Blank	0.000	00:09
Control Test	0.079	00:10
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 08/19/2019 16:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17618080A3

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 03:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: WILKES

DATE/TIME OF TEST: 07/14/2019 04:21:16

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:18	
Control Test	0.080	04:18	
Air Blank	0.000	04:18	
Subject Sample	# 1	04:19	Subject Test Refused
Breath Volume	0.000	04:19	
Air Blank	0.000	04:20	
Control Test	0.080	04:20	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:21	

Data Download Date/Time: 08/19/2019 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMONDS,RICHARD,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: TINDALL,RICHARD
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: FINAN

DATE/TIME OF TEST: 07/16/2019 02:50:42

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:48	
Control Test	0.080	02:48	
Air Blank	0.000	02:49	
Subject Sample	# 1	02:49	Subject Test Refused
Breath Volume	0.000	02:49	
Air Blank	0.000	02:49	
Control Test	0.080	02:50	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

Data Download Date/Time: 08/19/2019 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: MERCADO FIGUEROA

DATE/TIME OF TEST: 07/18/2019 00:50:54

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.081	00:46
Breath Volume	3.203	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:48
Subject Sample # 2	0.080	00:49
Breath Volume	3.074	00:49
Air Blank	0.000	00:49
Control Test	0.080	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 08/19/2019 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GOLOMBEK

DATE/TIME OF TEST: 07/19/2019 01:58:27

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.081	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.143	01:54
Breath Volume	3.511	01:54
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.144	01:56
Breath Volume	2.910	01:56
Air Blank	0.000	01:57
Control Test	0.080	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 08/19/2019 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: WILFONG

DATE/TIME OF TEST: 07/19/2019 05:39:33

Diagnostics Check	OK	05:31
Air Blank	0.000	05:32
Control Test	0.080	05:32
Air Blank	0.000	05:33
Subject Sample # 1	0.081	05:35
Breath Volume	5.300	05:35
Air Blank	0.000	05:35
Air Blank	0.000	05:37
Subject Sample # 2	0.081	05:38
Breath Volume	2.921	05:38
Air Blank	0.000	05:38
Control Test	0.080	05:38
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

Data Download Date/Time: 08/19/2019 16:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER, MATTHEW, D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: KYNION, JASMINE

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17618080A3

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 02:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: SUSI

DATE/TIME OF TEST: 07/20/2019 03:03:03

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.136	02:58
Breath Volume	3.332	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.133	03:01
Breath Volume	3.488	03:01
Air Blank	0.000	03:02
Control Test	0.080	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:03

Data Download Date/Time: 08/19/2019 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: HERNANDEZ GARCIA

DATE/TIME OF TEST: 07/20/2019 22:09:09

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.081	22:01
Air Blank	0.000	22:02
Subject Sample # 1	0.151	22:03
Breath Volume	1.785	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.145	22:07
Breath Volume	1.945	22:07
Air Blank	0.000	22:08
Control Test	0.079	22:08
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 08/19/2019 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17618080A3

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 00:49

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: MAYHEW

DATE/TIME OF TEST: 07/21/2019 01:41:23

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:38	
Control Test	0.081	01:39	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:40	Subject Test Refused
Breath Volume	0.000	01:40	
Air Blank	0.000	01:40	
Control Test	0.080	01:40	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

Data Download Date/Time: 08/19/2019 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: EHRHART, MATTHEW, B

BREATH TEST OPERATOR AGENCY: Indian Shores PD

ARRESTING OFFICER: HICKS, JASON

ARRESTING OFFICER AGENCY: Indian Shores PD

LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17618080A3

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 19:12

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: PENDERS

DATE/TIME OF TEST: 07/22/2019 20:27:29

Diagnostics Check	OK	20:20
Air Blank	0.000	20:20
Control Test	0.080	20:21
Air Blank	0.000	20:21
Subject Sample # 1	0.326	20:22
Breath Volume	1.789	20:22
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.328	20:25
Breath Volume	1.691	20:25
Air Blank	0.000	20:26
Control Test	0.079	20:26
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 08/19/2019 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATZ,JUSTIN,T
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: KATZ,JUSTIN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:46

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 07/26/2019 17:12:43

Diagnostics Check	OK	17:06
Air Blank	0.000	17:06
Control Test	0.080	17:07
Air Blank	0.000	17:07
Subject Sample # 1	0.274	17:08
Breath Volume	2.539	17:08
Air Blank	0.000	17:08
Air Blank	0.000	17:10
Subject Sample # 2	0.278	17:11
Breath Volume	2.253	17:11
Air Blank	0.000	17:11
Control Test	0.080	17:12
Air Blank	0.000	17:12
Diagnostics Check	OK	17:12

Data Download Date/Time: 08/19/2019 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: HASTINGS,KENNETH

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17618080A3

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 20:22

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: HUTCHINS

DATE/TIME OF TEST: 07/26/2019 21:13:25

Diagnostics Check	OK	21:06
Air Blank	0.000	21:07
Control Test	0.080	21:07
Air Blank	0.000	21:08
Subject Sample # 1	0.105	21:08
Breath Volume	1.578	21:08
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample # 2	0.105	21:11
Breath Volume	1.429	21:11
Air Blank	0.000	21:12
Control Test	0.081	21:12
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

Data Download Date/Time: 08/19/2019 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 07/01/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: CAREY

DATE/TIME OF TEST: 07/27/2019 19:35:27

Diagnostics Check	OK	19:32	
Air Blank	0.000	19:33	
Control Test	0.081	19:33	
Air Blank	0.000	19:33	
Subject Sample	# 1	19:34	Subject Test Refused
Breath Volume	0.000	19:34	
Air Blank	0.000	19:34	
Control Test	0.081	19:34	
Air Blank	0.000	19:35	
Diagnostics Check	OK	19:35	

Data Download Date/Time: 08/19/2019 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: WILTON

DATE/TIME OF TEST: 07/01/2019 00:41:08

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.081	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.115	00:35
Breath Volume	1.714	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.119	00:39
Breath Volume	2.644	00:39
Air Blank	0.000	00:40
Control Test	0.081	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 07/12/2019 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY, WESLEY, R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KELLY, WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 07/01/2019 19:07:12

Diagnostics Check	OK	19:00
Air Blank	0.000	19:01
Control Test	0.082	19:01
Air Blank	0.000	19:02
Subject Sample # 1	0.034	19:02
Breath Volume	1.644	19:02
Air Blank	0.000	19:03
Air Blank	0.000	19:05
Subject Sample # 2	0.031	19:05
Breath Volume	2.320	19:05
Air Blank	0.000	19:06
Control Test	0.081	19:06
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

Data Download Date/Time: 07/12/2019 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAROW,M,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: ASHLEY,R
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:43

SUBJECT'S LAST NAME: CHISHOLM

DATE/TIME OF TEST: 07/02/2019 10:51:00

Diagnostics Check	OK	10:44
Air Blank	0.000	10:45
Control Test	0.081	10:45
Air Blank	0.000	10:46
Subject Sample # 1	0.000	10:46
Breath Volume	2.609	10:46
Air Blank	0.000	10:47
Air Blank	0.000	10:49
Subject Sample # 2	0.000	10:49
Breath Volume	3.078	10:49
Air Blank	0.000	10:50
Control Test	0.081	10:50
Air Blank	0.000	10:50
Diagnostics Check	OK	10:51

Data Download Date/Time: 07/12/2019 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONNELL,JACOB,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CLOUTIER,AUSTIN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: O'LEARY

DATE/TIME OF TEST: 07/03/2019 02:25:52

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.082	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.000	02:21
Breath Volume	2.957	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.000	02:24
Breath Volume	3.562	02:24
Air Blank	0.000	02:24
Control Test	0.080	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 07/12/2019 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KELLY,WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:19

SUBJECT'S LAST NAME: FERRARO

DATE/TIME OF TEST: 07/03/2019 11:45:24

Diagnostics Check	OK	11:39
Air Blank	0.000	11:39
Control Test	0.082	11:40
Air Blank	0.000	11:40
Subject Sample # 1	0.000	11:41
Breath Volume	1.734	11:41
Air Blank	0.000	11:41
Air Blank	0.000	11:43
Subject Sample # 2	0.000	11:44
Breath Volume	2.003	11:44
Air Blank	0.000	11:44
Control Test	0.080	11:44
Air Blank	0.000	11:45
Diagnostics Check	OK	11:45

Data Download Date/Time: 07/12/2019 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,MARC,T
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LANE,MARC
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: PRENTICE

DATE/TIME OF TEST: 07/05/2019 00:08:54

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:04	
Control Test	0.081	00:04	
Air Blank	0.000	00:05	
Subject Sample	# 1	00:07	Subject Test Refused
Breath Volume	0.000	00:07	
Air Blank	0.000	00:08	
Control Test	0.080	00:08	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:08	

Data Download Date/Time: 07/12/2019 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: KISTLER

DATE/TIME OF TEST: 07/06/2019 00:10:43

Diagnostics Check		OK	00:04
Air Blank		0.000	00:04
Control Test		0.081	00:05
Air Blank		0.000	00:05
Subject Sample	# 1	0.186	00:05
Breath Volume		1.773	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.172	00:09
Breath Volume		1.128	00:09
Air Blank		0.000	00:09
Control Test		0.080	00:10
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 07/12/2019 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULASKI,DAMON,J
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: ROTHCHILD,FOREST
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: BUI
ARREST TIME: 17:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: LEVEROCK

DATE/TIME OF TEST: 07/07/2019 20:01:52

Diagnostics Check	OK	19:55
Air Blank	0.000	19:55
Control Test	0.081	19:56
Air Blank	0.000	19:56
Subject Sample # 1	0.042	19:57
Breath Volume	2.507	19:57
Air Blank	0.000	19:57
Air Blank	0.000	19:59
Subject Sample # 2	0.037	20:00
Breath Volume	1.664	20:00
Air Blank	0.000	20:00
Control Test	0.082	20:01
Air Blank	0.000	20:01
Diagnostics Check	OK	20:01

Data Download Date/Time: 07/12/2019 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWARZ,RILEY,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SCHWARZ,RILEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 07/09/2019 01:08:23

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:05	
Control Test	0.080	01:06	
Air Blank	0.000	01:06	
Subject Sample	# 1	01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:07	
Control Test	0.081	01:07	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 07/12/2019 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULASKI,DAMON,J
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: BALFOUR,KEVIN
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: BUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: CHRISTOFFERSON

DATE/TIME OF TEST: 07/10/2019 23:55:58

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:53	
Control Test	0.081	23:53	
Air Blank	0.000	23:54	
Subject Sample	# 1	23:54	Subject Test Refused
Breath Volume	0.000	23:54	
Air Blank	0.000	23:55	
Control Test	0.082	23:55	
Air Blank	0.000	23:55	
Diagnostics Check	OK	23:55	

Data Download Date/Time: 07/12/2019 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH-LONG,JEFFERSON,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SMITH-LONG,JEFFERSON
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: FRIESEN

DATE/TIME OF TEST: 07/11/2019 00:43:46

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.082	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.000	00:39
Breath Volume	3.824	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.000	00:42
Breath Volume	1.914	00:42
Air Blank	0.000	00:42
Control Test	0.084	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 07/12/2019 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: FUSSEL

DATE/TIME OF TEST: 07/12/2019 02:07:57

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.146	02:03
Breath Volume	1.917	02:03
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.158	02:06
Breath Volume	2.097	02:06
Air Blank	0.000	02:07
Control Test	0.081	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 07/12/2019 19:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NETTLES,TRISTAN,O
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SEQUEIRA,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:46

SUBJECT'S LAST NAME: TICHANUK

DATE/TIME OF TEST: 07/12/2019 12:21:27

Diagnostics Check	OK	12:07	
Air Blank	0.000	12:07	
Control Test	0.081	12:08	
Air Blank	0.000	12:08	
Subject Sample	# 1 0.007	12:09	
Breath Volume	1.449	12:09	
Air Blank	0.000	12:10	
Air Blank	0.000	12:12	
Subject Sample	# 2 0.007	12:15	Volume Not Met
Breath Volume	0.187	12:15	
Air Blank	0.000	12:15	
Air Blank	0.000	12:17	
Subject Sample	# 3 0.000	12:20	
Breath Volume	1.683	12:20	
Air Blank	0.000	12:20	
Control Test	0.081	12:20	
Air Blank	0.000	12:21	
Diagnostics Check	OK	12:21	

Data Download Date/Time: 07/12/2019 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK, TYLER, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIRK, TYLER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 07/13/2019 01:33:36

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:31	
Control Test	0.081	01:31	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:32	Subject Test Refused
Breath Volume	0.000	01:32	
Air Blank	0.000	01:32	
Control Test	0.082	01:33	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

Data Download Date/Time: 08/20/2019 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: DEFRANCESCO

DATE/TIME OF TEST: 07/14/2019 03:07:48

Diagnostics Check		OK	02:56	
Air Blank		0.000	02:56	
Control Test		0.080	02:57	
Air Blank		0.000	02:57	
Subject Sample	# 1	0.000	03:00	Volume Not Met
Breath Volume		1.027	03:00	
Air Blank		0.000	03:01	
Air Blank		0.000	03:03	
Subject Sample	# 2	0.000	03:06	Volume Not Met
Breath Volume		0.316	03:06	
Air Blank		0.000	03:06	
Control Test		0.081	03:07	
Air Blank		0.000	03:07	
Diagnostics Check		OK	03:07	

Data Download Date/Time: 08/20/2019 16:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: SPARKS

DATE/TIME OF TEST: 07/14/2019 22:45:06

Diagnostics Check		OK	22:32	
Air Blank		0.000	22:32	
Control Test		0.081	22:32	
Air Blank		0.000	22:33	
Subject Sample	# 1	0.224	22:36	Volume Not Met
Breath Volume		0.503	22:36	
Air Blank		0.000	22:37	
Air Blank		0.000	22:38	
Subject Sample	# 2	0.233	22:40	
Breath Volume		1.476	22:40	
Air Blank		0.000	22:41	
Air Blank		0.000	22:43	
Subject Sample	# 3	0.224	22:43	
Breath Volume		1.226	22:43	
Air Blank		0.000	22:44	
Control Test		0.078	22:44	
Air Blank		0.000	22:45	
Diagnostics Check		OK	22:45	

Data Download Date/Time: 08/20/2019 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEQUEIRA,NATHAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SEQUEIRA,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 09:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 07/17/2019 10:45:36

Diagnostics Check		OK	10:31	
Air Blank		0.000	10:32	
Control Test		0.081	10:32	
Air Blank		0.000	10:33	
Subject Sample	# 1	0.075	10:36	Volume Not Met
Breath Volume		0.722	10:36	
Air Blank		0.000	10:36	
Air Blank		0.000	10:38	
Subject Sample	# 2	0.080	10:40	
Breath Volume		1.242	10:40	
Air Blank		0.000	10:40	
Air Blank		0.000	10:42	
Subject Sample	# 3	0.078	10:44	
Breath Volume		1.460	10:44	
Air Blank		0.000	10:44	
Control Test		0.080	10:45	
Air Blank		0.000	10:45	
Diagnostics Check		OK	10:45	

Data Download Date/Time: 08/20/2019 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AIKEN,DURRIEL,R
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: AIKEN,DURRIEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: KORTEPETER

DATE/TIME OF TEST: 07/17/2019 20:08:19

Diagnostics Check		OK	20:02
Air Blank		0.000	20:02
Control Test		0.081	20:03
Air Blank		0.000	20:03
Subject Sample	# 1	0.000	20:04
Breath Volume		2.597	20:04
Air Blank		0.000	20:04
Air Blank		0.000	20:06
Subject Sample	# 2	0.000	20:06
Breath Volume		2.550	20:06
Air Blank		0.000	20:07
Control Test		0.081	20:07
Air Blank		0.000	20:08
Diagnostics Check		OK	20:08

Data Download Date/Time: 08/20/2019 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINTERS,LANE,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SHOUP,MICHAEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: LARIVIERE

DATE/TIME OF TEST: 07/18/2019 01:14:22

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.081	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.000	01:09
Breath Volume	1.980	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.000	01:13
Breath Volume	1.996	01:13
Air Blank	0.000	01:13
Control Test	0.081	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 08/20/2019 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 07/18/2019 03:25:21

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.080	03:17
Air Blank		0.000	03:17
Subject Sample	# 1	0.200	03:20
Breath Volume		2.027	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:22
Subject Sample	# 2	0.198	03:23
Breath Volume		2.234	03:23
Air Blank		0.000	03:24
Control Test		0.080	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 08/20/2019 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWARZ,RILEY,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SCHWARZ,RILEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: COONTZ

DATE/TIME OF TEST: 07/18/2019 03:59:37

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:51	
Control Test	0.080	03:51	
Air Blank	0.000	03:52	
Subject Sample	# 1	03:55	No Sample Provided
Breath Volume	0.000	03:55	
Air Blank	0.000	03:56	
Air Blank	0.000	03:58	
Subject Sample	# 2	03:58	Subject Test Refused
Breath Volume	0.000	03:58	
Air Blank	0.000	03:58	
Control Test	0.080	03:59	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

Data Download Date/Time: 08/20/2019 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWARZ,RILEY,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SCHWARZ,RILEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 07/19/2019 04:03:00

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:55	
Control Test	0.080	03:55	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:58	No Sample Provided
Breath Volume	0.000	03:58	
Air Blank	0.000	03:59	
Air Blank	0.000	04:01	
Subject Sample	# 2	04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:02	
Control Test	0.081	04:02	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 08/20/2019 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON, MATTHEW, F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 07/20/2019 04:03:52

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.081	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.156	03:59
Breath Volume	2.656	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:01
Subject Sample # 2	0.156	04:02
Breath Volume	1.187	04:02
Air Blank	0.000	04:02
Control Test	0.079	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 08/20/2019 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON, MATTHEW, F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: MCKENNA

DATE/TIME OF TEST: 07/21/2019 07:14:01

Diagnostics Check	OK	07:05	
Air Blank	0.000	07:06	
Control Test	0.081	07:06	
Air Blank	0.000	07:06	
Subject Sample	# 1	07:09	No Sample Provided
Breath Volume	0.000	07:09	
Air Blank	0.000	07:10	
Air Blank	0.000	07:12	
Subject Sample	# 2	07:12	Subject Test Refused
Breath Volume	0.000	07:12	
Air Blank	0.000	07:13	
Control Test	0.081	07:13	
Air Blank	0.000	07:13	
Diagnostics Check	OK	07:14	

Data Download Date/Time: 08/20/2019 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON, MATTHEW, F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: ALONSO

DATE/TIME OF TEST: 07/21/2019 20:27:37

Diagnostics Check	OK	20:20
Air Blank	0.000	20:20
Control Test	0.081	20:21
Air Blank	0.000	20:21
Subject Sample # 1	0.141	20:23
Breath Volume	1.343	20:23
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.161	20:26
Breath Volume	1.898	20:26
Air Blank	0.000	20:26
Control Test	0.081	20:27
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 08/20/2019 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON, MATTHEW, F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 07/22/2019 01:30:33

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.210	01:25
Breath Volume	3.460	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.214	01:28
Breath Volume	3.156	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 08/20/2019 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON, MATTHEW, F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SCHMIDT

DATE/TIME OF TEST: 07/25/2019 01:41:07

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:32	
Control Test	0.082	01:32	
Air Blank	0.000	01:33	
Subject Sample	# 1	01:36	No Sample Provided
Breath Volume	0.000	01:36	
Air Blank	0.000	01:36	
Air Blank	0.000	01:38	
Subject Sample	# 2	01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:40	
Control Test	0.081	01:40	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

Data Download Date/Time: 08/20/2019 16:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSTOS,ALEJANDRO,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BUSTOS,ALEJANDRO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:16

SUBJECT'S LAST NAME: USHER

DATE/TIME OF TEST: 07/25/2019 18:45:51

Diagnostics Check	OK	18:39
Air Blank	0.000	18:40
Control Test	0.082	18:40
Air Blank	0.000	18:40
Subject Sample # 1	0.000	18:41
Breath Volume	2.398	18:41
Air Blank	0.000	18:41
Air Blank	0.000	18:43
Subject Sample # 2	0.000	18:44
Breath Volume	1.671	18:44
Air Blank	0.000	18:44
Control Test	0.081	18:45
Air Blank	0.000	18:45
Diagnostics Check	OK	18:45

Data Download Date/Time: 08/20/2019 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON, MATTHEW, F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HANNON, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/25/2019 23:37:20

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.081	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.000	23:33
Breath Volume	2.125	23:33
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.000	23:35
Breath Volume	1.816	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 08/20/2019 16:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK, TYLER, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COMSTOCK, MICHAEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: W236501550640

DATE/TIME OF TEST: 07/27/2019 00:41:50

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.047	00:37
Breath Volume	3.480	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:39
Subject Sample # 2	0.046	00:40
Breath Volume	3.414	00:40
Air Blank	0.000	00:40
Control Test	0.081	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 08/20/2019 16:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK, TYLER, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIRK, TYLER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: SENESAC

DATE/TIME OF TEST: 07/28/2019 17:52:18

Diagnostics Check	OK	17:46
Air Blank	0.000	17:46
Control Test	0.080	17:46
Air Blank	0.000	17:47
Subject Sample # 1	0.282	17:47
Breath Volume	1.507	17:47
Air Blank	0.000	17:48
Air Blank	0.000	17:50
Subject Sample # 2	0.290	17:50
Breath Volume	2.394	17:50
Air Blank	0.000	17:51
Control Test	0.081	17:51
Air Blank	0.000	17:52
Diagnostics Check	OK	17:52

Data Download Date/Time: 08/20/2019 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: BATLEY

DATE/TIME OF TEST: 07/31/2019 22:31:38

Diagnostics Check	OK	22:25
Air Blank	0.000	22:25
Control Test	0.080	22:25
Air Blank	0.000	22:26
Subject Sample # 1	0.283	22:26
Breath Volume	3.949	22:26
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.284	22:30
Breath Volume	4.128	22:30
Air Blank	0.000	22:30
Control Test	0.082	22:31
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 08/20/2019 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARQUHARSON,KEVIN,D
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 06/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: RIVERA LUGO

DATE/TIME OF TEST: 07/05/2019 01:48:55

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.078	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.158	01:42
Breath Volume	1.546	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.150	01:47
Breath Volume	1.457	01:47
Air Blank	0.000	01:47
Control Test	0.077	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 07/25/2019 19:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROSKE, OFC J,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: MAGANA, OFC E
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 06/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: CARSWELL

DATE/TIME OF TEST: 07/16/2019 23:44:19

Diagnostics Check	OK	23:38
Air Blank	0.000	23:39
Control Test	0.078	23:39
Air Blank	0.000	23:39
Subject Sample # 1	0.000	23:40
Breath Volume	1.574	23:40
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.000	23:42
Breath Volume	1.753	23:42
Air Blank	0.000	23:43
Control Test	0.077	23:43
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 07/25/2019 19:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 07/03/2019 00:20:47

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.081	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.178	00:15
Breath Volume	1.546	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.176	00:19
Breath Volume	1.480	00:19
Air Blank	0.000	00:19
Control Test	0.081	00:20
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 07/18/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISSETT,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: SALINAS

DATE/TIME OF TEST: 07/03/2019 03:31:33

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.082	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.098	03:26
Breath Volume	2.382	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.095	03:30
Breath Volume	2.566	03:30
Air Blank	0.000	03:30
Control Test	0.081	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 07/18/2019 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 07/04/2019 03:30:08

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.082	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.216	03:25
Breath Volume	1.523	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:27
Subject Sample # 2	0.228	03:28
Breath Volume	1.742	03:28
Air Blank	0.000	03:29
Control Test	0.081	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 07/18/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ, M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: SANG

DATE/TIME OF TEST: 07/05/2019 19:54:08

Diagnostics Check	OK	19:47
Air Blank	0.000	19:48
Control Test	0.082	19:48
Air Blank	0.000	19:48
Subject Sample # 1	0.000	19:49
Breath Volume	1.796	19:49
Air Blank	0.000	19:50
Air Blank	0.000	19:52
Subject Sample # 2	0.000	19:52
Breath Volume	1.957	19:52
Air Blank	0.000	19:53
Control Test	0.081	19:53
Air Blank	0.000	19:54
Diagnostics Check	OK	19:54

Data Download Date/Time: 07/18/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: KAHLER

DATE/TIME OF TEST: 07/05/2019 23:56:55

Diagnostics Check	OK	23:48
Air Blank	0.000	23:48
Control Test	0.082	23:49
Air Blank	0.000	23:49
Subject Sample # 1	0.165	23:51
Breath Volume	1.671	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:53
Subject Sample # 2	0.161	23:55
Breath Volume	1.750	23:55
Air Blank	0.000	23:56
Control Test	0.081	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 07/18/2019 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ, M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: KELLY JR

DATE/TIME OF TEST: 07/06/2019 01:30:48

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:28	
Control Test	0.080	01:28	
Air Blank	0.000	01:29	
Subject Sample	# 1	01:29	Subject Test Refused
Breath Volume	0.000	01:29	
Air Blank	0.000	01:29	
Control Test	0.082	01:30	
Air Blank	0.000	01:30	
Diagnostics Check	OK	01:30	

Data Download Date/Time: 07/18/2019 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: BRETT

DATE/TIME OF TEST: 07/12/2019 01:03:41

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:01	
Control Test	0.081	01:01	
Air Blank	0.000	01:02	
Subject Sample	# 1	01:02	Subject Test Refused
Breath Volume	0.000	01:02	
Air Blank	0.000	01:02	
Control Test	0.081	01:03	
Air Blank	0.000	01:03	
Diagnostics Check	OK	01:03	

Data Download Date/Time: 07/18/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,S,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: FRANCO

DATE/TIME OF TEST: 07/12/2019 02:12:59

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.081	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.191	02:08
Breath Volume	1.917	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.185	02:11
Breath Volume	2.671	02:11
Air Blank	0.000	02:12
Control Test	0.082	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 07/18/2019 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEAN,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: FERNANDEZ RIVERA

DATE/TIME OF TEST: 07/12/2019 04:06:03

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:03	
Control Test	0.081	04:03	
Air Blank	0.000	04:04	
Subject Sample	# 1	04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:05	
Control Test	0.081	04:05	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 07/18/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN, D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 07/12/2019 23:54:16

Diagnostics Check	OK	23:46
Air Blank	0.000	23:47
Control Test	0.081	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.139	23:49
Breath Volume	1.378	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.144	23:52
Breath Volume	2.308	23:52
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 07/18/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: RIVAS HERNANDEZ

DATE/TIME OF TEST: 07/13/2019 02:59:52

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.081	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.140	02:55
Breath Volume		1.230	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:57
Subject Sample	# 2	0.132	02:58
Breath Volume		2.074	02:58
Air Blank		0.000	02:58
Control Test		0.081	02:59
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 07/18/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COLE, D
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: KEILI

DATE/TIME OF TEST: 07/13/2019 04:45:45

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:43	
Control Test	0.081	04:43	
Air Blank	0.000	04:44	
Subject Sample	# 1	04:44	Subject Test Refused
Breath Volume	0.000	04:44	
Air Blank	0.000	04:44	
Control Test	0.081	04:45	
Air Blank	0.000	04:45	
Diagnostics Check	OK	04:45	

Data Download Date/Time: 07/18/2019 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTELL,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: BAUGH

DATE/TIME OF TEST: 07/13/2019 18:51:31

Diagnostics Check	OK	18:48	
Air Blank	0.000	18:49	
Control Test	0.081	18:49	
Air Blank	0.000	18:49	
Subject Sample	# 1	18:50	Subject Test Refused
Breath Volume	0.000	18:50	
Air Blank	0.000	18:50	
Control Test	0.081	18:50	
Air Blank	0.000	18:51	
Diagnostics Check	OK	18:51	

Data Download Date/Time: 07/18/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 07/13/2019 22:17:14

Diagnostics Check	OK	22:14	
Air Blank	0.000	22:14	
Control Test	0.082	22:15	
Air Blank	0.000	22:15	
Subject Sample	# 1	22:15	Subject Test Refused
Breath Volume	0.000	22:15	
Air Blank	0.000	22:16	
Control Test	0.081	22:16	
Air Blank	0.000	22:17	
Diagnostics Check	OK	22:17	

Data Download Date/Time: 07/18/2019 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MCINTYRE

DATE/TIME OF TEST: 07/14/2019 03:02:11

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.013	02:57
Breath Volume	1.355	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.013	03:00
Breath Volume	1.558	03:00
Air Blank	0.000	03:01
Control Test	0.081	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 07/18/2019 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: PAYAN

DATE/TIME OF TEST: 07/14/2019 05:36:18

Diagnostics Check	OK	05:29
Air Blank	0.000	05:30
Control Test	0.081	05:30
Air Blank	0.000	05:31
Subject Sample # 1	0.096	05:31
Breath Volume	1.742	05:31
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.090	05:34
Breath Volume	1.324	05:34
Air Blank	0.000	05:35
Control Test	0.081	05:35
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 07/18/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ, M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 07/15/2019 02:29:59

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.082	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.055	02:25
Breath Volume	1.621	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.060	02:28
Breath Volume	1.777	02:28
Air Blank	0.000	02:29
Control Test	0.081	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 07/18/2019 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: ESPINO

DATE/TIME OF TEST: 07/16/2019 01:25:44

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.081	01:18
Air Blank	0.000	01:18
Subject Sample # 1	0.177	01:19
Breath Volume	1.691	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:21
Subject Sample # 2	0.166	01:24
Breath Volume	1.218	01:24
Air Blank	0.000	01:24
Control Test	0.081	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 07/18/2019 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: DENSON

DATE/TIME OF TEST: 07/21/2019 02:02:38

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.081	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.132	01:58
Breath Volume	2.464	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.127	02:01
Breath Volume	2.269	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 07/30/2019 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HALBRITTER,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: VARNES

DATE/TIME OF TEST: 07/21/2019 04:02:24

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.081	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.224	03:57
Breath Volume	3.902	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.229	04:00
Breath Volume	4.101	04:00
Air Blank	0.000	04:01
Control Test	0.082	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 07/30/2019 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWES,T
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: NUNNALLY

DATE/TIME OF TEST: 07/21/2019 12:05:45

Diagnostics Check	OK	12:02	
Air Blank	0.000	12:02	
Control Test	0.081	12:02	
Air Blank	0.000	12:03	
Subject Sample	# 1	12:04	Subject Test Refused
Breath Volume	0.000	12:04	
Air Blank	0.000	12:04	
Control Test	0.082	12:05	
Air Blank	0.000	12:05	
Diagnostics Check	OK	12:05	

Data Download Date/Time: 07/30/2019 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WOLFORD,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: STEINER

DATE/TIME OF TEST: 07/21/2019 16:51:57

Diagnostics Check	OK	16:49	
Air Blank	0.000	16:49	
Control Test	0.081	16:49	
Air Blank	0.000	16:50	
Subject Sample	# 1	16:50	Subject Test Refused
Breath Volume	0.000	16:50	
Air Blank	0.000	16:51	
Control Test	0.081	16:51	
Air Blank	0.000	16:51	
Diagnostics Check	OK	16:51	

Data Download Date/Time: 07/30/2019 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 07/21/2019 23:26:48

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.081	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.158	23:22
Breath Volume	2.011	23:22
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.171	23:25
Breath Volume	2.757	23:25
Air Blank	0.000	23:25
Control Test	0.079	23:26
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 07/30/2019 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MCFARLAND,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: HAILEYESUS

DATE/TIME OF TEST: 07/22/2019 05:24:06

Diagnostics Check	OK	05:21	
Air Blank	0.000	05:21	
Control Test	0.081	05:22	
Air Blank	0.000	05:22	
Subject Sample	# 1	05:22	Subject Test Refused
Breath Volume	0.000	05:22	
Air Blank	0.000	05:23	
Control Test	0.081	05:23	
Air Blank	0.000	05:24	
Diagnostics Check	OK	05:24	

Data Download Date/Time: 07/30/2019 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CALVY

DATE/TIME OF TEST: 07/23/2019 03:44:04

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:41	
Control Test	0.081	03:42	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:43	
Control Test	0.081	03:43	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 07/30/2019 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RUSSELL,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: BOWLING

DATE/TIME OF TEST: 07/23/2019 16:43:04

Diagnostics Check	OK	16:36
Air Blank	0.000	16:37
Control Test	0.082	16:37
Air Blank	0.000	16:38
Subject Sample # 1	0.176	16:38
Breath Volume	2.007	16:38
Air Blank	0.000	16:39
Air Blank	0.000	16:40
Subject Sample # 2	0.171	16:41
Breath Volume	2.402	16:41
Air Blank	0.000	16:42
Control Test	0.081	16:42
Air Blank	0.000	16:42
Diagnostics Check	OK	16:43

Data Download Date/Time: 07/30/2019 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HURLOCK,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 07/25/2019 00:56:54

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.081	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.000	00:51
Breath Volume	1.781	00:51
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.000	00:55
Breath Volume	1.632	00:55
Air Blank	0.000	00:55
Control Test	0.081	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 07/30/2019 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 07/25/2019 04:03:32

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.081	03:55	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:59	No Sample Provided
Breath Volume	0.000	03:59	
Air Blank	0.000	03:59	
Air Blank	0.000	04:02	
Subject Sample	# 2	04:02	Subject Test Refused
Breath Volume	0.000	04:02	
Air Blank	0.000	04:02	
Control Test	0.081	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 07/30/2019 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/18/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: MINK

DATE/TIME OF TEST: 07/30/2019 08:54:58

Diagnostics Check	OK	08:52	
Air Blank	0.000	08:52	
Control Test	0.081	08:52	
Air Blank	0.000	08:53	
Subject Sample	# 1	08:53	Subject Test Refused
Breath Volume	0.000	08:53	
Air Blank	0.000	08:54	
Control Test	0.081	08:54	
Air Blank	0.000	08:54	
Diagnostics Check	OK	08:54	

Data Download Date/Time: 07/30/2019 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN, D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: PROEFROCK

DATE/TIME OF TEST: 07/31/2019 02:06:31

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.081	02:00
Air Blank	0.000	02:01
Subject Sample # 1	0.191	02:02
Breath Volume	1.148	02:02
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.201	02:04
Breath Volume	1.480	02:04
Air Blank	0.000	02:05
Control Test	0.082	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 08/06/2019 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,TABARIE,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,TABARIE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 07/20/2019 03:33:18

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:32	
Control Test	0.036	03:32	Control Outside Tolerance
Air Blank	0.000	03:33	

Data Download Date/Time: 07/24/2019 19:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,TABARIE,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,TABARIE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 07/20/2019 03:43:37

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.076	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.144	03:39
Breath Volume	2.148	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.133	03:42
Breath Volume	1.339	03:42
Air Blank	0.000	03:42
Control Test	0.076	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 07/24/2019 19:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,TABARIE,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,TABARIE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: LIVINGSTON

DATE/TIME OF TEST: 07/25/2019 00:44:35

Diagnostics Check	OK	00:44	
Air Blank		00:44	RFI Detect
Air Blank	0.000	00:44	

Data Download Date/Time: 08/29/2019 18:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,TABARIE,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,TABARIE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: LIVINGSTON

DATE/TIME OF TEST: 07/25/2019 00:53:29

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.077	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.147	00:48
Breath Volume	2.351	00:48
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.147	00:51
Breath Volume	2.027	00:51
Air Blank	0.000	00:52
Control Test	0.077	00:52
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 08/29/2019 18:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALCON,IVAN,N
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: FALCON,IVAN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: ALCALA

DATE/TIME OF TEST: 07/20/2019 19:06:17

Diagnostics Check	OK	18:58
Air Blank	0.000	18:59
Control Test	0.082	18:59
Air Blank	0.000	19:00
Subject Sample # 1	0.209	19:01
Breath Volume	2.570	19:01
Air Blank	0.000	19:02
Air Blank	0.000	19:04
Subject Sample # 2	0.219	19:04
Breath Volume	1.675	19:04
Air Blank	0.000	19:05
Control Test	0.080	19:05
Air Blank	0.000	19:06
Diagnostics Check	OK	19:06

Data Download Date/Time: 07/31/2019 21:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTOPASSI,BARON,H
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: LEWIS,PAM
ARRESTING OFFICER AGENCY: Apalachicola PD
LAST AGENCY INSPECTION: 06/06/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 09:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:36

SUBJECT'S LAST NAME: NOBLE

DATE/TIME OF TEST: 07/04/2019 10:08:05

Diagnostics Check	OK	09:56	
Air Blank	0.000	09:57	
Control Test	0.078	09:57	
Air Blank	0.000	09:57	
Subject Sample	# 1	10:01	No Sample Provided
Breath Volume	0.000	10:01	
Air Blank	0.000	10:01	
Air Blank	0.000	10:03	
Subject Sample	# 2	10:06	No Sample Provided
Breath Volume	0.000	10:06	
Air Blank	0.000	10:07	
Control Test	0.074	10:07	Control Outside Tolerance
Air Blank	0.000	10:08	

Data Download Date/Time: 07/15/2019 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOZADA RIVERA, ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP TROOP K
ARRESTING OFFICER: LOZADA RIVERA, ANTHONY
ARRESTING OFFICER AGENCY: FHP TROOP K
LAST AGENCY INSPECTION: 06/28/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 07/07/2019 16:13:43

Diagnostics Check	OK	16:07
Air Blank	0.000	16:07
Control Test	0.080	16:08
Air Blank	0.000	16:08
Subject Sample # 1	0.326	16:09
Breath Volume	3.054	16:09
Air Blank	0.000	16:09
Air Blank	0.000	16:11
Subject Sample # 2	0.320	16:12
Breath Volume	3.144	16:12
Air Blank	0.000	16:12
Control Test	0.080	16:13
Air Blank	0.000	16:13
Diagnostics Check	OK	16:13

Data Download Date/Time: 07/31/2019 22:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,ADAM,H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: POTTER,ADAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: HAMPTON

DATE/TIME OF TEST: 07/01/2019 02:40:52

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:36	
Control Test	0.081	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:39	Subject Test Refused
Breath Volume	0.000	02:39	
Air Blank	0.000	02:39	
Control Test	0.081	02:40	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:40	

Data Download Date/Time: 07/25/2019 07:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 07/02/2019 20:11:37

Diagnostics Check	OK	20:08	
Air Blank	0.000	20:09	
Control Test	0.079	20:09	
Air Blank	0.000	20:09	
Subject Sample	# 1	20:10	Subject Test Refused
Breath Volume	0.000	20:10	
Air Blank	0.000	20:10	
Control Test	0.079	20:11	
Air Blank	0.000	20:11	
Diagnostics Check	OK	20:11	

Data Download Date/Time: 07/25/2019 07:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINET,RICKY,H
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BINET,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: PACE

DATE/TIME OF TEST: 07/03/2019 03:19:41

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.081	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.148	03:15
Breath Volume	1.921	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.159	03:18
Breath Volume	2.031	03:18
Air Blank	0.000	03:18
Control Test	0.079	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 07/25/2019 07:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTS,JAMES,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: TURNAGE,MELISSA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: ALLARD

DATE/TIME OF TEST: 07/03/2019 17:12:41

Diagnostics Check	OK	17:01	
Air Blank	0.000	17:01	
Control Test	0.078	17:01	
Air Blank	0.000	17:02	
Subject Sample	# 1	17:05	No Sample Provided
Breath Volume	0.000	17:05	
Air Blank	0.000	17:06	
Air Blank	0.000	17:08	
Subject Sample	# 2 0.207	17:11	Volume Not Met
Breath Volume	0.621	17:11	
Air Blank	0.000	17:11	
Control Test	0.081	17:12	
Air Blank	0.000	17:12	
Diagnostics Check	OK	17:12	

Data Download Date/Time: 07/25/2019 07:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, ADAM, H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER, ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: FULTON

DATE/TIME OF TEST: 07/04/2019 00:41:00

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.079	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.215	00:35
Breath Volume		1.253	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.216	00:39
Breath Volume		1.261	00:39
Air Blank		0.000	00:40
Control Test		0.077	00:40
Air Blank		0.000	00:40
Diagnostics Check		OK	00:41

Data Download Date/Time: 07/25/2019 07:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,ROBERT,C
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MENDEZ,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: KAREUS

DATE/TIME OF TEST: 07/04/2019 11:45:16

Diagnostics Check	OK	11:42	
Air Blank	0.000	11:42	
Control Test	0.080	11:43	
Air Blank	0.000	11:43	
Subject Sample	# 1	11:43	Subject Test Refused
Breath Volume	0.000	11:43	
Air Blank	0.000	11:44	
Control Test	0.079	11:44	
Air Blank	0.000	11:45	
Diagnostics Check	OK	11:45	

Data Download Date/Time: 07/25/2019 07:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: JOINER

DATE/TIME OF TEST: 07/05/2019 01:32:37

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.148	01:28
Breath Volume	1.847	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.149	01:31
Breath Volume	1.546	01:31
Air Blank	0.000	01:31
Control Test	0.080	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 07/25/2019 07:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/06/2019 01:28:59

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.162	01:24
Breath Volume	1.199	01:24
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.159	01:27
Breath Volume	1.480	01:27
Air Blank	0.000	01:28
Control Test	0.078	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 07/25/2019 07:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: SCHWARTZ

DATE/TIME OF TEST: 07/06/2019 04:17:38

Diagnostics Check	OK	04:14	
Air Blank	0.000	04:15	
Control Test	0.080	04:15	
Air Blank	0.000	04:16	
Subject Sample	# 1	04:16	Subject Test Refused
Breath Volume	0.000	04:16	
Air Blank	0.000	04:16	
Control Test	0.081	04:17	
Air Blank	0.000	04:17	
Diagnostics Check	OK	04:17	

Data Download Date/Time: 07/25/2019 07:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 07/06/2019 20:35:50

Diagnostics Check	OK	20:29
Air Blank	0.000	20:29
Control Test	0.081	20:30
Air Blank	0.000	20:30
Subject Sample # 1	0.256	20:31
Breath Volume	2.375	20:31
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample # 2	0.266	20:34
Breath Volume	1.566	20:34
Air Blank	0.000	20:34
Control Test	0.079	20:35
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

Data Download Date/Time: 07/25/2019 07:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: URIBE

DATE/TIME OF TEST: 07/07/2019 18:28:42

Diagnostics Check	OK	18:22
Air Blank	0.000	18:22
Control Test	0.081	18:23
Air Blank	0.000	18:23
Subject Sample # 1	0.152	18:24
Breath Volume	1.691	18:24
Air Blank	0.000	18:24
Air Blank	0.000	18:26
Subject Sample # 2	0.152	18:27
Breath Volume	1.574	18:27
Air Blank	0.000	18:27
Control Test	0.080	18:28
Air Blank	0.000	18:28
Diagnostics Check	OK	18:28

Data Download Date/Time: 07/25/2019 07:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: GUADALUPE

DATE/TIME OF TEST: 07/08/2019 01:03:10

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.080	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.242	00:58
Breath Volume	1.246	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.232	01:01
Breath Volume	2.312	01:01
Air Blank	0.000	01:02
Control Test	0.078	01:02
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 07/25/2019 07:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTON,XAVIER,V
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/09/2019 20:25:43

Diagnostics Check	OK	20:19
Air Blank	0.000	20:20
Control Test	0.080	20:20
Air Blank	0.000	20:21
Subject Sample # 1	0.241	20:21
Breath Volume	2.070	20:21
Air Blank	0.000	20:22
Air Blank	0.000	20:24
Subject Sample # 2	0.242	20:24
Breath Volume	1.660	20:24
Air Blank	0.000	20:24
Control Test	0.079	20:25
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 07/25/2019 07:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKE,CLEVELAND,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SCHEIDEGG,COREY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: SINGLETON

DATE/TIME OF TEST: 07/10/2019 01:44:43

Diagnostics Check		OK	01:37
Air Blank		0.000	01:38
Control Test		0.079	01:38
Air Blank		0.000	01:39
Subject Sample	# 1	0.282	01:40
Breath Volume		1.417	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:42
Subject Sample	# 2	0.283	01:43
Breath Volume		1.222	01:43
Air Blank		0.000	01:43
Control Test		0.078	01:44
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 07/25/2019 07:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDAH,PATTY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: YATES,SAMUEL
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:09

SUBJECT'S LAST NAME: CHAMPAGNE

DATE/TIME OF TEST: 07/13/2019 16:41:15

Diagnostics Check	OK	16:32
Air Blank	0.000	16:33
Control Test	0.081	16:33
Air Blank	0.000	16:34
Subject Sample # 1	0.000	16:35
Breath Volume	1.101	16:35
Air Blank	0.000	16:35
Air Blank	0.000	16:37
Subject Sample # 2	0.000	16:39
Breath Volume	1.277	16:39
Air Blank	0.000	16:40
Control Test	0.082	16:40
Air Blank	0.000	16:41
Diagnostics Check	OK	16:41

Data Download Date/Time: 07/25/2019 07:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTON,XAVIER,V
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: SERMONS

DATE/TIME OF TEST: 07/14/2019 03:02:14

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.079	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.184	02:57
Breath Volume	1.835	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.186	03:00
Breath Volume	1.375	03:00
Air Blank	0.000	03:01
Control Test	0.078	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 07/25/2019 07:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITCHIE,SHELBY,W
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RITCHIE,SHELBY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: FORSMAN

DATE/TIME OF TEST: 07/14/2019 23:10:17

Diagnostics Check	OK	23:06	
Air Blank	0.000	23:06	
Control Test	0.081	23:07	
Air Blank	0.000	23:07	
Subject Sample	# 1	23:08	Subject Test Refused
Breath Volume	0.000	23:08	
Air Blank	0.000	23:09	
Control Test	0.080	23:09	
Air Blank	0.000	23:10	
Diagnostics Check	OK	23:10	

Data Download Date/Time: 07/25/2019 07:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,JORDAN,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: RAMOS FLORES

DATE/TIME OF TEST: 07/16/2019 06:32:59

Diagnostics Check	OK	06:23
Air Blank	0.000	06:24
Control Test	0.079	06:24
Air Blank	0.000	06:24
Subject Sample # 1	0.000	06:26
Breath Volume	2.171	06:26
Air Blank	0.000	06:27
Air Blank	0.000	06:29
Subject Sample # 2	0.000	06:31
Breath Volume	2.042	06:31
Air Blank	0.000	06:32
Control Test	0.080	06:32
Air Blank	0.000	06:32
Diagnostics Check	OK	06:32

Data Download Date/Time: 07/25/2019 07:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: TRUJILLO

DATE/TIME OF TEST: 07/20/2019 00:26:54

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.079	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.147	00:22
Breath Volume	2.882	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.148	00:25
Breath Volume	2.636	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 07/25/2019 07:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: BUCCERI

DATE/TIME OF TEST: 07/20/2019 01:12:38

Diagnostics Check		OK	01:06
Air Blank		0.000	01:06
Control Test		0.080	01:07
Air Blank		0.000	01:07
Subject Sample	# 1	0.228	01:07
Breath Volume		2.250	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.219	01:11
Breath Volume		1.777	01:11
Air Blank		0.000	01:11
Control Test		0.080	01:12
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 07/25/2019 07:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: FAUSTINO

DATE/TIME OF TEST: 07/20/2019 01:49:21

Diagnostics Check		OK	01:40
Air Blank		0.000	01:40
Control Test		0.081	01:40
Air Blank		0.000	01:41
Subject Sample	# 1	0.203	01:42
Breath Volume		1.695	01:42
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.182	01:44
Breath Volume		2.453	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 3	0.187	01:47
Breath Volume		1.859	01:47
Air Blank		0.000	01:48
Control Test		0.080	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 07/25/2019 07:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONCION,CHRISTIAN,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MONCION,CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: BERNIER

DATE/TIME OF TEST: 07/20/2019 18:34:48

Diagnostics Check	OK	18:28
Air Blank	0.000	18:28
Control Test	0.079	18:29
Air Blank	0.000	18:29
Subject Sample # 1	0.139	18:30
Breath Volume	1.445	18:30
Air Blank	0.000	18:31
Air Blank	0.000	18:32
Subject Sample # 2	0.137	18:33
Breath Volume	1.753	18:33
Air Blank	0.000	18:33
Control Test	0.080	18:34
Air Blank	0.000	18:34
Diagnostics Check	OK	18:34

Data Download Date/Time: 07/25/2019 07:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES, STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES, STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: HAYDOCK

DATE/TIME OF TEST: 07/20/2019 22:41:11

Diagnostics Check	OK	22:27	
Air Blank	0.000	22:27	
Control Test	0.080	22:27	
Air Blank	0.000	22:28	
Subject Sample	# 1	22:31	No Sample Provided
Breath Volume	0.000	22:31	
Air Blank	0.000	22:32	
Air Blank	0.000	22:34	
Subject Sample	# 2	0.320	22:36
Breath Volume	1.597	22:36	
Air Blank	0.000	22:37	
Air Blank	0.000	22:38	
Subject Sample	# 3	0.338	22:39
Breath Volume	1.632	22:39	
Air Blank	0.000	22:40	
Control Test	0.081	22:40	
Air Blank	0.000	22:41	
Diagnostics Check	OK	22:41	

Data Download Date/Time: 07/25/2019 07:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: HUSE

DATE/TIME OF TEST: 07/21/2019 01:10:52

Diagnostics Check	OK	00:59	
Air Blank	0.000	00:59	
Control Test	0.081	01:00	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:03	No Sample Provided
Breath Volume	0.000	01:03	
Air Blank	0.000	01:04	
Air Blank	0.000	01:06	
Subject Sample	# 2	01:09	No Sample Provided
Breath Volume	0.000	01:09	
Air Blank	0.000	01:09	
Control Test	0.080	01:10	
Air Blank	0.000	01:10	
Diagnostics Check	OK	01:10	

Data Download Date/Time: 07/25/2019 07:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: BAMACA

DATE/TIME OF TEST: 07/21/2019 03:57:40

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.079	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.221	03:53
Breath Volume		1.761	03:53
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.216	03:56
Breath Volume		2.421	03:56
Air Blank		0.000	03:56
Control Test		0.080	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 07/25/2019 07:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BING, TRAE, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CARTER, JACOB
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: BUI
ARREST TIME: 16:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: REGISTER

DATE/TIME OF TEST: 07/21/2019 18:30:48

Diagnostics Check	OK	18:29	
Air Blank	0.000	18:29	
Control Test	0.074	18:30	Control Outside Tolerance
Air Blank	0.000	18:30	

Data Download Date/Time: 07/25/2019 07:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BING, TRAE, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CARTER, JACOB
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: BUI
ARREST TIME: 16:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: REGISTER

DATE/TIME OF TEST: 07/21/2019 18:40:48

Diagnostics Check	OK	18:33
Air Blank	0.000	18:33
Control Test	0.079	18:34
Air Blank	0.000	18:34
Subject Sample # 1	0.107	18:35
Breath Volume	1.285	18:35
Air Blank	0.000	18:36
Air Blank	0.000	18:38
Subject Sample # 2	0.123	18:39
Breath Volume	3.218	18:39
Air Blank	0.000	18:39
Control Test	0.078	18:40
Air Blank	0.000	18:40
Diagnostics Check	OK	18:40

Data Download Date/Time: 07/25/2019 07:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DAVIS,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: WHALEN

DATE/TIME OF TEST: 07/22/2019 00:26:38

Diagnostics Check		OK	00:19
Air Blank		0.000	00:20
Control Test		0.081	00:20
Air Blank		0.000	00:21
Subject Sample	# 1	0.204	00:21
Breath Volume		1.421	00:21
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.198	00:25
Breath Volume		2.074	00:25
Air Blank		0.000	00:25
Control Test		0.080	00:26
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 07/25/2019 07:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINTOSH,HALLIE,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: VINES

DATE/TIME OF TEST: 07/22/2019 21:31:11

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:26	
Control Test	0.079	21:26	
Air Blank	0.000	21:27	
Subject Sample	# 1	21:29	Subject Test Refused
Breath Volume	0.000	21:29	
Air Blank	0.000	21:30	
Control Test	0.081	21:30	
Air Blank	0.000	21:31	
Diagnostics Check	OK	21:31	

Data Download Date/Time: 07/25/2019 07:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINTOSH,HALLIE,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/26/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: GRIDER

DATE/TIME OF TEST: 07/24/2019 04:44:52

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.079	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.154	04:39
Breath Volume	1.488	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.143	04:43
Breath Volume	2.003	04:43
Air Blank	0.000	04:43
Control Test	0.079	04:44
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 07/25/2019 07:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:49

SUBJECT'S LAST NAME: WALLINGFORD

DATE/TIME OF TEST: 07/25/2019 16:17:47

Diagnostics Check		OK	16:09
Air Blank		0.000	16:10
Control Test		0.080	16:10
Air Blank		0.000	16:10
Subject Sample	# 1	0.206	16:13
Breath Volume		2.281	16:13
Air Blank		0.000	16:13
Air Blank		0.000	16:15
Subject Sample	# 2	0.198	16:16
Breath Volume		1.417	16:16
Air Blank		0.000	16:16
Control Test		0.078	16:17
Air Blank		0.000	16:17
Diagnostics Check		OK	16:17

Data Download Date/Time: 08/27/2019 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 07/25/2019 17:25:18

Diagnostics Check	OK	17:21	
Air Blank	0.000	17:21	
Control Test	0.082	17:21	
Air Blank	0.000	17:22	
Subject Sample	# 1	17:23	Interferent Detect
Breath Volume	1.800	17:23	
Air Blank	0.000	17:24	
Control Test	0.084	17:24	
Air Blank	0.000	17:25	
Diagnostics Check	OK	17:25	

Data Download Date/Time: 08/27/2019 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 07/25/2019 17:32:33

Diagnostics Check	OK	17:26
Air Blank	0.000	17:26
Control Test	0.079	17:27
Air Blank	0.000	17:27
Subject Sample # 1	0.000	17:28
Breath Volume	1.386	17:28
Air Blank	0.000	17:28
Air Blank	0.000	17:30
Subject Sample # 2	0.000	17:31
Breath Volume	1.125	17:31
Air Blank	0.000	17:31
Control Test	0.080	17:31
Air Blank	0.000	17:32
Diagnostics Check	OK	17:32

Data Download Date/Time: 08/27/2019 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, ADAM, H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER, ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: ESPINAL

DATE/TIME OF TEST: 07/27/2019 01:03:16

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.115	00:58
Breath Volume	2.121	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.117	01:01
Breath Volume	2.351	01:01
Air Blank	0.000	01:02
Control Test	0.081	01:02
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 08/27/2019 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,ADAM,H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: POTTER,ADAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/24/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A6
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: LECLAIR

DATE/TIME OF TEST: 07/28/2019 19:34:06

Diagnostics Check	OK	19:22	
Air Blank	0.000	19:23	
Control Test	0.080	19:23	
Air Blank	0.000	19:24	
Subject Sample	# 1	19:27	No Sample Provided
Breath Volume	0.000	19:27	
Air Blank	0.000	19:27	
Air Blank	0.000	19:29	
Subject Sample	# 2	19:32	No Sample Provided
Breath Volume	0.000	19:32	
Air Blank	0.000	19:33	
Control Test	0.078	19:33	
Air Blank	0.000	19:34	
Diagnostics Check	OK	19:34	

Data Download Date/Time: 08/27/2019 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JASON,G
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BROWN,JASON
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/11/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13413080A1
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:53

SUBJECT'S LAST NAME: MERCER

DATE/TIME OF TEST: 07/06/2019 11:17:01

Diagnostics Check	OK	11:14	
Air Blank	0.000	11:14	
Control Test	0.080	11:14	
Air Blank	0.000	11:15	
Subject Sample	# 1	11:15	Subject Test Refused
Breath Volume	0.000	11:15	
Air Blank	0.000	11:16	
Control Test	0.079	11:16	
Air Blank	0.000	11:16	
Diagnostics Check	OK	11:17	

Data Download Date/Time: 07/22/2019 19:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JARED,R
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: DAVIS,JARED
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 07/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/25/2019 03:58:58

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.130	03:54
Breath Volume	1.265	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:56
Subject Sample # 2	0.133	03:57
Breath Volume	2.195	03:57
Air Blank	0.000	03:58
Control Test	0.076	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 08/20/2019 18:54